

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
OCT 30 2018
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS COUNTY SO
Time of Inspection: 22:33

Date of Inspection: 10/24/2018

Serial Number: 80-001003
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG815602 Exp: 06/05/2020
0.000	0.044 / 0.047	0.077	0.192	0.076
0.000	0.045 / 0.048	0.078	0.193	0.076
0.000	0.044 / 0.048	0.077	0.194	0.076

Number of Simulators Used: 5

Remarks:

A F / M A: Ambient Fail. 05: Control Outside Tolerance UNK PROB REPEAT OK.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

M. Deane _____ M DEANE
Signature and Printed Name

10/24/2018
Date

SP
11/6/18
SP