

Florida Department of Law Enforcement Alcohol Testing Program

12/26/18
JE

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAINES CITY PD
Time of Inspection: 13:35

Date of Inspection: 12/25/2018

Serial Number: 80-001007
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 684131 Exp: 01/23/2020
0.000	0.049	0.080	0.199	0.081
0.000	0.051	0.081	0.200	0.081
0.000	0.050	0.081	0.200	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 *RAY M. WEBER* ^{AC96}

RAY M WEBER
Signature and Printed Name

12/25/2018
Date

Florida Department of Law Enforcement BK Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAINES CITY PD
Time of Inspection: 07:08

Date of Inspection: 11/26/2018

Serial Number: 80-001007
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK	Yes	No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 684131 Exp: 01/23/2020
0.000	0.051	0.082	0.200	0.081
0.000	0.052	0.081	0.200	0.081
0.000	0.053	0.082	0.199	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 RAY M WEBER
Signature and Printed Name

11/26/2018
Date

80-001007 *MS*

RECEIVED

OCT 30 2018

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAINES CITY PD
Time of Inspection: 08:22

Date of Inspection: 10/30/2018

Serial Number: 80-001007
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 684131 Exp: 01/23/2020
0.000	0.051	0.081	0.199	0.082
0.000	0.051	0.081	0.200	0.081
0.000	0.052	0.080	0.200	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 cpl. WEBER #896 RAY M WEBER
Signature and Printed Name

10/30/2018
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

SEP 11 2018

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAINES CITY PD

Time of Inspection: 13:44

Date of Inspection: 09/11/2018

Serial Number: 80-001007

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 684131 Exp: 01/23/2020
0.000	0.050	0.080	0.201	0.082
0.000	0.050	0.080	0.201	0.083
0.000	0.051	0.080	0.201	0.082

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



RAY M. WEBER #896

RAY M WEBER

Signature and Printed Name

09/11/2018
Date

80-001007 ~~800~~

RECEIVED

AUG 22 2018

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAINES CITY PD

Time of Inspection: 16:32

Date of Inspection: 08/22/2018

Serial Number: 80-001007

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:684131 Exp: 01/23/2020
0.000	0.050	0.080	0.199	0.082
0.000	0.050	0.081	0.200	0.082
0.000	0.051	0.081	0.200	0.083

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 RAY M WEBER
Signature and Printed Name

08/22/2018
Date

80-001007 ~~008~~

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JUL 08 2018

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAINES CITY PD

Time of Inspection: 19:10

Date of Inspection: 07/08/2018

Serial Number: 80-001007

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:684131 Exp: 01/23/2020
0.000	0.050	0.080	0.200	0.081
0.000	0.051	0.080	0.200	0.081
0.000	0.050	0.081	0.200	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

CPL. WEBER # 896

Signature and Printed Name

RAY M WEBER

07/08/2018

Date

80-001007 **RECEIVED**

JUN 10 2018

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAINES CITY PD
Time of Inspection: 22:12
Date of Inspection: 06/10/2018
Serial Number: 80-001007
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:684131 Exp: 01/23/2020
0.000	0.053	0.081	0.202	0.082
0.000	0.053	0.081	0.202	0.082
0.000	0.054	0.081	0.202	0.082

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 **RAY M WEBER**
Signature and Printed Name
06/10/2018
Date

80-001007 *me*

RECEIVED

MAY 13 2018

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAINES CITY PD

Time of Inspection: 00:51

Date of Inspection: 05/13/2018

Serial Number: 80-001007

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1684131 Exp: 01/23/2020
0.000	0.050	0.081	0.201	0.083
0.000	0.051	0.080	0.201	0.083
0.000	0.051	0.080	0.201	0.083

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


 RAY M WEBER #896

Signature and Printed Name

05/13/2018
Date

RECEIVED

APR 15 2018

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAINES CITY PD
Time of Inspection: 17:35

Date of Inspection: 04/15/2018

Serial Number: 80-001007
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:684131 Exp: 01/23/2020
0.000	0.051	0.080	0.200	0.082
0.000	0.052	0.080	0.200	0.082
0.000	0.051	0.081	0.201	0.082

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

RAY M WEBER

Signature and Printed Name

04/15/2018
Date

80-001007 RNS

RECEIVED
MAR 27 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAINES CITY PD
Time of Inspection: 00:49

Date of Inspection: 03/27/2018

Serial Number: 80-001007
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:684131 Exp: 01/23/2020
0.000	0.051	0.081	0.201	0.082
0.000	0.052	0.081	0.202	0.083
0.000	0.052	0.082	0.201	0.083

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 CPL WEBER #896 RAY M WEBER
Signature and Printed Name

03/27/2018
Date

80-001007 RMB

RECEIVED

FEB 28 2018

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAINES CITY PD
Time of Inspection: 06:45

Date of Inspection: 02/28/2018

Serial Number: 80-001007
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:684131 Exp: 01/23/2020
0.000	0.052	0.081	0.201	0.081
0.000	0.051	0.081	0.201	0.082
0.000	0.052	0.080	0.201	0.082

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 RAY M WEBER
Signature and Printed Name

02/28/2018
Date

RECEIVED

JAN 12 2018

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAINES CITY PD

Time of Inspection: 11:31

Date of Inspection: 01/10/2018

Serial Number: 80-001007

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:684131 Exp: 01/23/2020
0.000	0.051	0.081	0.200	0.082
0.000	0.051	0.081	0.200	0.081
0.000	0.051	0.081	0.201	0.082

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

RAY M WEBER

Signature and Printed Name

01/10/2018
Date