

Florida Department of Law Enforcement Alcohol Testing Program

1008
48

RECEIVED

JAN 30 2018

AGENCY INSPECTION REPORT - INTOXILYZER 8000

FDLE

Agency: WEST MELBOURNE P.D.
Time of Inspection: 11:14

Date of Inspection: 01/25/2018

Serial Number: 81001008
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 04216080A3 Exp: 03/05/2018
0.000	0.049	0.078	0.194	0.079
0.000	0.049	0.078	0.195	0.078
0.000	0.049	0.078	0.195	0.079

Number of Simulators Used: 4

Remarks:

It was found that the mouth Alcohol solution and the Acetone stock solution that was used during this inspection are both expired. The mouth Alcohol solution expired 03-27-17 and the Acetone stock solution expired 01-13-18. This did not effect the inspection outcome. AMENDED 01-29-18

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

SHANE Z POPA
Signature and Printed Name

01/25/2018
Date