

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COCONUT CREEK PD
Time of Inspection: 10:37

Date of Inspection: 05/30/2018

Serial Number: 80-001044
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:686331 Exp: 01/31/2020
0.000	0.049	0.076	0.000 / 0.193	0.078
0.000	0.049	0.077	0.000 / 0.195	0.079
0.000	0.050	0.078	0.000 / 0.196	0.078

Number of Simulators Used: 4

Remarks:

20: Control Outside Tolerance. SEAL BETWEEN HEAD UNIT & CONTAINER FAILED AND CAUSED AIR LEAK. CHANGED HEAD UNIT & PASSED.


RECEIVED

JUN 12 2018

**FDLE
Alcohol Testing Program**

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 #17979
KYLE DUFFEY
Signature and Printed Name

05/30/2018
Date