80-001048 OOAS

RECEIVED
OCT 29 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE ALFRED PD
Time of Inspection: 17:43 Date of Inspection: 10/19/2018
Serial Number: 80-001048
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:31517080A1 Exp: 01/02/2020
0.000	0.048	0.077	0.200	0.081
0.000	0.048	0.078	0.200	0.081
0.000	0.049	0.078	0.201	0.081

Number of Simulators Used: 4

Remarks:
TEMP REG TEST FAILEDNon-compliance:TEMP REG FAILED DURING DIAGNOSTICS.

The above instrument complies ( ) does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

OPA #25 GAPD ANTHONY W GETTLE
Signature and Printed Name

10/19/2018
Date

FDLE/ATP Form 40 – March 2004

80-001048 DOS

RECEIVED
AUG 30 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE ALFRED PD
Time of Inspection: 17:37 Date of Inspection: 08/29/2018
Serial Number: 80-001048
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:31517080A1 Exp: 01/05/2020
/ 0.000	0.047	0.076	0.201	0.080
/ 0.000	0.049	0.077	0.202	0.080
/ 0.000	0.049	0.078	0.203	0.081

Number of Simulators Used: 4

Remarks:
00: RFI DetectRADIO KEYED UP.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

OFC. #25 LAPD
ANTHONY W GETTLE
Signature and Printed Name

08/29/2018
Date

FDLE/ATP Form 40 – March 2004

80-001048 2018

RECEIVED
JUL 19 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE ALFRED PD
Time of Inspection: 10:01 Date of Inspection: 07/19/2018
Serial Number: 80-001048
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:31517080A1 Exp: 01/05/2020
0.000	0.047	0.078	0.202	0.082
0.000	0.049	0.078	0.202	0.081
0.000	0.049	0.079	0.202	0.081

Number of Simulators Used: 4

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Off. #25 LAPD ANTHONY W GETTLE
Signature and Printed Name

07/19/2018
Date

FDLE/ATP Form 40 – March 2004

80-001048 DDS

RECEIVED
JUN 07 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE ALFRED PD
Time of Inspection: 16:15 Date of Inspection: 06/07/2018
Serial Number: 80-001048
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:31517080A1 Exp: 01/05/2020
0.000	0.048	0.079	0.202	0.081
0.000	0.048	0.079	0.202	0.081
0.000	0.049	0.079	0.202	0.081

Number of Simulators Used: 4

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

OFC #25 LAPD ANTHONY W GETTLE
Signature and Printed Name

06/07/2018
Date

FDLE/ATP Form 40 – March 2004

80-001048

RECEIVED
MAY 20 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE ALFRED PD
Time of Inspection: 23:54 Date of Inspection: 05/19/2018
Serial Number: 80-001048
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:31517080A1 Exp: 01/05/2020
0.000	0.048	0.077	0.202	0.082
0.000	0.049	0.078	0.202	0.081
0.000	0.049	0.078	0.202	0.081

Number of Simulators Used: 4

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

#25 LAPD

ANTHONY W GETTLE
Signature and Printed Name

05/19/2018
Date

FDLE/ATP Form 40 – March 2004

80-001048

RECEIVED

APR 22 2018

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE ALFRED PD
Time of Inspection: 11:25 Date of Inspection: 04/22/2018 Serial Number: 80-001048
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:31517080A1 Exp: 01/05/2020
0.000	0.049	0.078	0.202	0.082
0.000	0.050	0.079	0.202	0.081
0.000	0.050	0.080	0.202	0.082

Number of Simulators Used: 4

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

OFC. #25 LAPD ANTHONY W GETTLE
Signature and Printed Name

04/22/2018
Date

FDLE/ATP Form 40 – March 2004

80-001048 OVS

RECEIVED
MAR 29 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE ALFRED PD
Time of Inspection: 18:06
Date of Inspection: 03/29/2018
Serial Number: 80-001048
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:31517080A1 Exp: 01/05/2020
0.000	0.049	0.079	0.201	0.081
0.000	0.050	0.079	0.201	0.081
0.000	0.050	0.079	0.201	0.081

Number of Simulators Used: 4

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

OFC. [signature] #25 CAPD

ANTHONY W GETTLE
Signature and Printed Name

03/29/2018
Date

FDLE/ATP Form 40 – March 2004

80-001048

RECEIVED
FEB 19 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE ALFRED PD
Time of Inspection: 12:39 Date of Inspection: 02/19/2018
Serial Number: 80-001048
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:31517080A1 Exp: 01/05/2020
0.000	0.049	0.079	0.202	0.082
0.000	0.050	0.078	0.201	0.081
0.000	0.050	0.079	0.201	0.081

Number of Simulators Used: 4

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

OFC. #25 CAPP ANTHONY W GETTLE
Signature and Printed Name

02/19/2018
Date

FDLE/ATP Form 40 – March 2004

80-001048 EDMS

RECEIVED
JAN 31 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE ALFRED PD
Time of Inspection: 11:18 Date of Inspection: 01/31/2018
Serial Number: 80-001048
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:31517080A1 Exp: 01/05/2020
0.000	0.050	0.080	0.199	0.083
0.000	0.050	0.081	0.199	0.082
0.000	0.050	0.081	0.200	0.083

Number of Simulators Used: 3

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Off. #25 WPD

ANTHONY W GETTLE
Signature and Printed Name

01/31/2018
Date

FDLE/ATP Form 40 – March 2004