

80-001051 MB

# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

OCT 16 2018

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ST PETERSBURG PD  
Time of Inspection: 08:26

Date of Inspection: 10/15/2018

Serial Number: 80-001051  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 22817080A2 * Exp: 10/05/2019
0.000	0.052	0.083	0.203	0.078
0.000	0.052	0.082	0.202	0.079
0.000	0.051	0.082	0.202	0.079

Number of Simulators Used: 5

Remarks:  
IN COMPLIANCE

\* Entered lot number incorrectly. Lot Number is 22817080A4

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MICHAEL D WEISKOPF

Signature and Printed Name

10/15/2018  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

80-001051 008

**RECEIVED**

JUN 07 2018

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ST PETERSBURG PD  
Time of Inspection: 08:50

Date of Inspection: 06/07/2018

Serial Number: 80-001051  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 22817080A2 ✖ Exp: 10/05/2019
0.000	0.053	0.084	0.205	0.078
0.000	0.052	0.083	0.205	0.079
0.000	0.052	0.083	0.204	0.079

Number of Simulators Used: 5

Remarks:  
IN COMPLIANCE

✖ Mistyped lot number for Dry Gas. Should be 22817080A4

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Michael D Weiskopf MICHAEL D WEISKOPF  
Signature and Printed Name

06/07/2018  
Date