1054
SP

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 07:49
Date of Inspection: 11/27/2018
Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22817080A5 Exp: 10/05/2019
0.000	0.049	0.077	0.198	0.080
0.000	0.049	0.077	0.198	0.078
0.000	0.048	0.077	0.197	0.079

Number of Simulators Used: 5

Remarks:

A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY Shon W. Abeln 2184 D.S.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184
SHON W ABELN
Signature and Printed Name

11/27/2018
Date

FDLE/ATP Form 40 – March 2004

1054
SP

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 12:38
Date of Inspection: 10/20/2018
Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22817080A5 Exp: 10/05/2019
0.000	0.047	0.079	0.197	0.079
0.000	0.047	0.078	0.197	0.079
0.000	0.048	0.079	0.195	0.079

Number of Simulators Used: 5

Remarks:

A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY Shon W. Abeln 2184 D.S.

RECEIVED
OCT 25 2018
FDLE
Alcohol Testing Program

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184 SHON W ABELN
Signature and Printed Name

10/20/2018
Date

FDLE/ATP Form 40 – March 2004

1054
8P

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 22:44
Date of Inspection: 09/26/2018
Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22817080A5 Exp: 10/05/2019
0.000	0.046	0.074 / 0.077	0.194	0.079
0.000	0.046	0.075 / 0.076	0.195	0.078
0.000	0.046	0.076 / 0.077	0.195	0.079

Number of Simulators Used: 5

Remarks:
08: Control Outside Tolerance.
Simulator JBR Lid not tight enough. tightened then repeated.

RECEIVED
OCT 01 2018
FDLE
Alcohol Testing Program

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GREGORY S GLASCO
Signature and Printed Name

09/26/2018
Date

1054
SP

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 22:33
Date of Inspection: 08/26/2018
Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22817080A5 Exp: 10/05/2019
0.000	0.046	0.075	0.192	0.079
0.000	0.045	0.076	0.194	0.078
0.000	0.046	0.076	0.194	0.077

Number of Simulators Used: 5

Remarks:

A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY_____D.S.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GREGORY S GLASCO
Signature and Printed Name

08/26/2018
Date

1054 SP

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 08:29 Date of Inspection: 06/29/2018

Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22817080A5 Exp: 10/05/2019
0.000	0.049	0.078	0.197	0.078
0.000	0.050	0.078	0.196	0.078
0.000	0.049	0.078	0.197	0.077

Number of Simulators Used: 5

Remarks:

RECEIVED
JUL 05 2018
FDLE
Alcohol Testing Program

A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY Shon W. Abeln 2184 D.S.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184 SHON W ABELN
Signature and Printed Name

06/29/2018
Date

FDLE/ATP Form 40 – March 2004

1054
SP

RECEIVED
JUN 14 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 15:15
Date of Inspection: 05/27/2018
Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22817080A5 Exp: 10/25/2019
0.000	0.048	0.076	0.187 / 0.195	0.077
0.000	0.049	0.076	0.192 / 0.196	0.077
0.000	0.049	0.076	0.194 / 0.197	0.076

Number of Simulators Used: 5

Remarks:
20: Control Outside Tolerance.
Simulator was off and wasn't up to heat yet. Shon W. Abeln
Amended 6/14/2018 Sgt. Abeln 2184

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184
SHON W ABELN
Signature and Printed Name

05/27/2018
Date

FDLE/ATP Form 40 – March 2004

FDLE
Florida Department of Law Enforcement

AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: HIGHLANDS County SO Instrument Serial Number: 80-00 1054

AGENCY INSPECTION DATA REVIEW

Agency Inspector: SHON ABELN | Date of Inspection: 5/27/18 | Time of Inspection: 15:15

Agency Inspection Discrepancy: ☒ Incomplete ☐ Untimely/Not Received ☐ Erroneous Information ☐ Procedural ☐ Other ____

☐ Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.

☐ Lot Number ☐Expiration Date for ____g/ 210L ☐Alcohol Reference Solution ☐Dry Gas Standard is ☐Incorrect ☐Expired.

☒ FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report – Intoxilyzer 8000. The ☐REASON for repeating the following test(s); OR the ☒Possible Cause and Corrective Action Taken on the following test(s) was not recorded:

☐ Alcohol Free Subject Test ☐ Mouth Alcohol Test ☐ Alcohol Free Test ☐ Interferent Detect Test
☐ 0.05 g/210L Test ☐ 0.08 g/210L Test ☒ 0.20 g/210L Test ☐ 0.08 g/210L Dry Gas Standard Test

☐ FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector."

- ☐ The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC.
- ☐ The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use.
- ☐ The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.

☐ The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.

☐ Other: ____

OTHER ELECTRONIC DATA REVIEW

☐ Login Records Date: ____	Comments: ____
☐ Cylinder Change Records Date: ____	
☐ Control Test Records Date: ____	
☐ Diagnostic Check Records Date: ____	

CORRECTIVE ACTION

☒ Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by 6/22 (Date).
☐ Provide a written explanation regarding the referenced item(s) to the Department Inspector by ____ (Date).
☐ Upload the Agency Inspection(s).
☐ Remove the instrument from evidentiary use until otherwise directed by the Department.
☐ No action required
☐ Other: ____

Shayla Platt
Signature of Alcohol Testing Program Staff Member

6/13/2018
Date

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

Agency HIGHLANDS COUNTY SO
Inspector ABELN,SHON,W

Date 05/27/2018
Time 15:15:49

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A
Expiration 03/07/2019

0.08 Lot # 201709A
Expiration 09/12/2019

0.20 Lot # 201702B
Expiration 02/23/2019

0.08 Gas Lot # 22817080A5
Expiration 10/25/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.187
0.20 g/210L Test # (Repeat) 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.076
0.20 g/210L Test # 2	0.192
0.20 g/210L Test # (Repeat) 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.076
0.20 g/210L Test # 3	0.194
0.20 g/210L Test # (Repeat) 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.076
Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance.

Data Download Date/Time 05/27/2018 15:17

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement
Alcohol Testing Program

1054
SP

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 15:15
Date of Inspection: 05/27/2018
Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22817080A5 Exp: 10/25/2019
0.000	0.048	0.076	0.187 / 0.195	0.077
0.000	0.049	0.076	0.192 / 0.196	0.077
0.000	0.049	0.076	0.194 / 0.197	0.076

Number of Simulators Used: 5

Remarks:
20: Control Outside Tolerance.

Simulator was off and wasn't up to heat yet. Shon W. Abeln

A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY Shon W. Abeln 2184 D.S.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184 SHON W ABELN
Signature and Printed Name

05/27/2018
Date

FDLE/ATP Form 40 – March 2004

1054
58

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 07:30
Date of Inspection: 04/25/2018
Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32316080A3 Exp: 01/05/2019
0.000	0.049	0.076	0.195	0.080
0.000	0.049	0.077	0.196	0.081
0.000	0.048	0.077	0.196	0.080

Number of Simulators Used: 5

Remarks:

A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY Shon W. Abeln 2184 D.S.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184
SHON W ABELN
Signature and Printed Name

04/25/2018
Date

FDLE/ATP Form 40 – March 2004

1054
58

RECEIVED
FEB 01 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 03:53
Date of Inspection: 01/27/2018
Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32316080A3 Exp: 01/05/2019
0.000	0.046	0.076	0.193	0.079
0.000	0.046	0.076	0.194	0.079
0.000	0.047	0.076	0.195	0.079

Number of Simulators Used: 5

Remarks:

A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY_____D.S.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GREGORY S GLASCO
Signature and Printed Name

01/27/2018
Date

FDLE/ATP Form 40 – March 2004

1054
SP

RECEIVED
FEB 01 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Serial Number: 80-001054
Time of Inspection: 03:12
Date of Inspection: 01/27/2018
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: 5

Remarks:
A F / M A: Range Exceeded. — I had too much alcohol solution in my mouth, then H.H in compliance by mistake. Immediately re inspected with no Issues.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GREGORY S GLASCO
Signature and Printed Name

01/27/2018
Date

FDLE/ATP Form 40 – March 2004