

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULFPORT PD
Time of Inspection: 17:31

Date of Inspection: 09/30/2018

Serial Number: 80-001061
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/01/2019 <i>M</i> <i>12</i>	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17817080A2 Exp: 08/05/2019
0.000	0.048	0.076	/ 0.195	0.080
0.000	0.048	0.077	/ 0.195	0.080
0.000	0.047	0.077	/ 0.195	0.080

Number of Simulators Used: 3

Remarks:
20: RFI Detect.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Michael S. Marotta
Signature and Printed Name

MICHAEL S MAROTTA

09/30/2018
Date

8D-001061

RECEIVED

AUG 02 2018

FDLE Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULFPORT PD

Serial Number: 80-001061

Time of Inspection: 15:16

Date of Inspection: 07/31/2018

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: <i>M 201703A</i> Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 17817080A2 Exp: 08/05/2019
0.000	0.000 / 0.050	0.074 / 0.077	0.188 / 0.193	0.081
0.000	0.000 / 0.050	0.075 / 0.077	0.191 / 0.195	0.080
0.000	0.000 / 0.050	0.076 / 0.077	0.193 / 0.194	0.081

Number of Simulators Used: 3

Remarks:

05: Control Outside Tolerance. LOOSE HOSE

M
~~e. LOOSE HOSE~~

.08 Allowed sim to heat longer. *M*
.20 Allowed sim to heat longer. *M*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MICHAEL S MAROTTA

Signature and Printed Name