

Florida Department of Law Enforcement
Alcohol Testing Program

80-001064

RECEIVED
OCT 29 2018

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: NEW PORT RICHEY PD

Serial Number: 80-001064

Time of Inspection: 22:51

Date of Inspection: 10/21/2018

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: Exp:
0.000	0.046			
0.000	0.047			
0.000	0.047			

Number of Simulators Used: 4

Remarks: 10-29-18 JSE
Y. UNABLE TO BACK OUT OF MENU AFTER REVIEWING RESULTS. POSSIBLE KEY BOARD CONNECTOR WAS LOOSE. ADJUSTED THE CONNECTOR AND RESTARTED TEST WITH NO ISSUES.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 Jason S Engel NP407
JASON S ENGEL
Signature and Printed Name

10/21/2018
Date