

80-001076 *MB*

RECEIVED

APR 12 2018

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ZEPHYRHILLS PD

Time of Inspection: 15:29

Date of Inspection: 04/04/2018

Serial Number: 80-001076

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.04 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: 4

Remarks:

4/4/18

*TOO MUCH MOUTH ALCOHOL REPEATED TEST BUT DID NOT REPEAT ALCOHOL
4/4/18 FREE TEST WAS ABORTED.*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector's License and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

04/04/2018
Date