

80-001078

RECEIVED

FEB 13 2018

FDLE  
Alcohol Testing Program

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ST PETERSBURG PD  
Time of Inspection: 10:41

Date of Inspection: 02/05/2018

Serial Number: 80-001078  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 26326080A2 * Exp: 11/05/2018
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.079	0.199	0.080
0.000	0.050	0.079	0.199	0.080

Number of Simulators Used: 5

Remarks:  
IN COMPLIANCE

Entered wrong digit  
Correct lot number is  
#26316080A2

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MICHAEL D WEISKOPF  
Signature and Printed Name

02/05/2018  
Date