

# Florida Department of Law Enforcement Alcohol Testing Program

12/20/18  
JO

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO  
Time of Inspection: 18:11

Date of Inspection: 12/19/2018

Serial Number: 80-001117  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#: 201802B<br>Exp: 02/06/2020 | 0.08g/210L Test (g/210L)<br>Lot#: 201709A<br>Exp: 09/12/2019 | 0.20g/210L Test (g/210L)<br>Lot#: 201702B<br>Exp: 02/23/2019 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#: 02618080A3<br>Exp: 03/05/2020 |
|----------------------------|--|--|--|--|
| 0.000                      | 0.018 / 0.047  | 0.077  | 0.201  | 0.082  |
| 0.000                      | 0.019 / 0.048  | 0.078  | 0.203  | 0.082  |
| 0.000                      | 0.022 / 0.048  | 0.078  | 0.204  | 0.082  |


Number of Simulators Used: 5

**Remarks:**

05: Control Outside Tolerance SIM WAS TURNED OFF. SIM WAS TURNED BACK ON AND REPEATED

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



BARRY NIXON

Signature and Printed Name

12/19/2018  
Date

# Florida Department of Law Enforcement BK Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO  
Time of Inspection: 18:57

Date of Inspection: 11/26/2018

Serial Number: 80-001117  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201802B<br>Exp: 02/06/2020 | 0.08g/210L Test (g/210L)<br>Lot#:201709A<br>Exp: 09/12/2019 | 0.20g/210L Test (g/210L)<br>Lot#:201702B<br>Exp: 02/23/2019 | 0.08 g/210L Dry Gas Std Test (g/210L) 02018080A<br>Lot#:02018080A3<br>Exp: 03/05/2020 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.047   | 0.075   | 0.194   | 0.000 / 0.082   |
| 0.000                      | 0.047   | 0.075   | 0.195   | 0.082 / 0.082   |
| 0.000                      | 0.047   | 0.076   | 0.196   | 0.082 / 0.082   |

*3 39/2  
BJ  
11/26/18*

Number of Simulators Used: 5

**Remarks:**

08: Control Outside Tolerance FORGOT TO RECONNECT HOSE TO Dry Gas. Connected hose then retested 11/26/18 Bo 392

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

\_\_\_\_\_  
Signature and Printed Name BARRY NIXON

11/26/2018  
Date