

6/14/18

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP D ORLANDO  
Time of Inspection: 02:03

Date of Inspection: 05/25/2018

Serial Number: 80-001123  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 22817980AB Exp: 10/05/2019
0.000 / 0.000	0.049	0.078	0.197	0.079
RFI / 0.000	0.049	0.078	0.198	0.079
/ 0.000	0.048	0.078	0.198	0.078

Number of Simulators Used: 1

Remarks:

RFI Detect.

RFI Detect. on head in with radio on

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JACOB A. KUEHN  
Signature and Printed Name

05/25/2018  
Date