

Florida Department of Law Enforcement  
Alcohol Testing Program

**RECEIVED**

NOV 09 2018

FDLE  
Alcohol Testing Program

**AGENCY INSPECTION REPORT - INTOXILYZER 8000**

Agency: VOLUSIA COUNTY SO

Serial Number: 80-001154

Time of Inspection: 11:37

Date of Inspection: 11/09/2018

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: Exp:
0.000	0.048	0.076	0.242 / 0.187	
0.000	0.049	0.078	0.225 / 0.194	
0.000	0.049	0.078	0.216 / 0.197	

Number of Simulators Used: 5**Remarks:**

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: . Aerator disconnected on both 34L simulator during inspection. Reconnected Aerator. Changed Alcohol reference solution and completed new inspection. Informed FDLE Department Inspector D. Bell of the incident.  
The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Jimmie N. Stone*

JIMMIE N STONE

Signature and Printed Name

11/09/2018

Date