

1166
SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BREVARD COUNTY SO
Time of Inspection: 16:45

Date of Inspection: 12/27/2018

Serial Number: 80-001166
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:19817080A3 Exp: 09/05/2019
0.000	0.050	0.081	0.200	0.086 / 0.086
0.000	0.050	0.081	0.201	0.086 / 0.086
0.000	0.050	0.080	0.201	0.086 / 0.086

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:CONTROL OUTSIDE TOLERANCE 0 N DRY GAS STANDAR.

I SPOKE WITH T.J. GRAHAM, FOLE/ATP, WHO STATED TO TRY ANOTHER DRY GAS TANK AND RUN SIX STABILITY CHECKS. HE STATED IF THE PROBLEM PERSISTED W/ NEW TANK, THEN IT IS AN INSTRUMENT ISSUE. IF PROBLEM DIDNT PERSIST, THEN RUN ANOTHER INSPECTION. I RAN TEN STABILITY CHECKS AND STILL HAD THE ISSUE OF CONTROL OUTSIDE TOLERANCE (0.086), EVEN WITH NEW TANK.

THIS INSTRUMENT IS TAKEN OUT OF SERVICE AND IS BEING SENT TO FOLE FOR SERVICE AND ANNUAL INSPECTION.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Thomas A Garrity

THOMAS A GARRITY

Signature and Printed Name

12/27/2018

Date

1166
SP

BREVARD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001166
12/27/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	17:56
Control Test	0.087	17:56
Air Blank	0.000	17:56
Control Test	0.086	17:57
Air Blank	0.000	17:57
Control Test	0.085	17:58
Air Blank	0.000	17:58
Control Test	0.085	17:58
Air Blank	0.000	17:59
Control Test	0.085	17:59
Air Blank	0.000	18:00
Control Test	0.085	18:00
Air Blank	0.000	18:01
Control Test Stats		
Average	0.0855	
Std Dev	0.0008	
Rel Std Dev(%)	0.9785	

BREVARD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001166
12/27/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	18:16
Control Test	0.086	18:17
Air Blank	0.000	18:18
Control Test	0.085	18:18
Air Blank	0.000	18:18
Control Test	0.084	18:19
Air Blank	0.000	18:19
Control Test	0.085	18:19
Air Blank	0.000	18:20
Control Test Stats		
Average	0.0850	
Std Dev	0.0008	
Rel Std Dev(%)	0.9606	

Paula J. [Signature] 446
Operator's Signature

Paula J. [Signature] 446
Operator's Signature

AMENDED
7/14/18

11/26
SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BREVARD COUNTY SO
Time of Inspection: 13:06

Date of Inspection: 06/13/2018

Serial Number: 80-001166
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) <i>201611B</i> Lot#: 201703A <i>TKS</i> Exp: 03/07/2019 <i>11/15/2018</i>	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 19817080A3 Exp: 09/05/2019
0.000	0.050	0.080	0.200	0.085
0.000	0.051	0.080	0.200	0.084
0.000	0.051	0.081	0.200	0.085

Number of Simulators Used: 5

Remarks:

Asst By: DEP Reg #121 (WAGNER)

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Thomas A. Garrity
THOMAS A GARRITY
Signature and Printed Name

06/13/2018
Date