

1189 SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MARTIN COUNTY SO
Time of Inspection: 10:52

Date of Inspection: 04/27/2018

Serial Number: 80-001189
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:06717080A7 Exp: 05/05/2019
0.000	0.050	0.079	0.201	0.081
0.000	0.050	0.080	0.201	0.081
0.000	0.051	0.079	0.201	0.081

Number of Simulators Used: 5

Remarks:

RECEIVED
MAY 02 2018
FDLE
Alcohol Testing Program

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

P/S C. [Signature] 5/14/1602
CHRISTOPHER L THOMAS
Signature and Printed Name

04/27/2018
Date