

# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JUN 07 2018

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: AUBURNDALE PD  
Time of Inspection: 22:18

Date of Inspection: 05/18/2018

Serial Number: 80-001231  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 03417080A3 Exp: 03/05/2019
0.000	0.049	0.070 / 0.075	0.200	0.080
0.000	0.049	0.072 / 0.077	0.200	0.080
0.000	0.050	0.075 / 0.077	0.200	0.080

Number of Simulators Used: 4

Remarks:

08: Control Outside Tolerance.

Agitator malfunction, attempted before 20 min time frame allowing to heat up properly.

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

THOMAS J LOKIETEK

Signature and Printed Name

05/18/2018  
Date