1261 SP

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.
Time of Inspection: 10:02 Date of Inspection: 10/29/2018

Serial Number: 80-001261
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:572920 Exp: 09/13/2019
0.000	0.051	0.080	0.202	0.080
0.000	0.050	0.082	0.203	0.080
0.000	0.051	0.080	0.203	0.081

Number of Simulators Used: 4

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

SGT. #185 PATRICK W KENNEDY

Signature and Printed Name

10/29/2018

Date

FDLE/ATP Form 40 – March 2004

1261 SP

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.
Time of Inspection: 04:42 Date of Inspection: 10/23/2018

Serial Number: 80-001261
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:572920 Exp: 09/13/2019
0.000	0.049	0.072 / 0.078	0.128 / 0.203	0.080
0.000	0.049	0.075 / 0.079	0.139 / 0.203	0.079
0.000	0.049	0.078 / 0.078	0.139 / 0.205	0.079

Number of Simulators Used: 4

Remarks:
08: Control Outside Tolerance SEAL NOT TIGHT. HEATING ISSUE (20 TEST)
20: HEATING ISSUE, CHECKED SEAL, ~~WAITING~~ WAITED ADDITIONAL 20 MINUTES REPEATED TEST. 10/28/18 #180

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

#180 IVETTE GOMEZ
Signature and Printed Name

10/23/2018
Date

1261 SP

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.
Time of Inspection: 10:04 Date of Inspection: 09/30/2018

Serial Number: 80-001261
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:572920 Exp: 09/13/2019
0.000	0.049	0.080	0.198	0.080
0.000	0.050	0.080	0.202	0.080
0.000	0.050	0.081	0.199	0.081

Number of Simulators Used: 4

Remarks:

RECEIVED
OCT 04 2018
FDLE
Alcohol Testing Program

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

#180 IVETTE GOMEZ
Signature and Printed Name

09/30/2018
Date

FDLE/ATP Form 40 – March 2004

1261 SP

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.
Time of Inspection: 09:12
Date of Inspection: 08/27/2018
Serial Number: 80-001261
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:572920 Exp: 09/13/2019
0.000	0.052	0.083	0.205	0.080
0.000	0.052	0.082	0.206	0.080
0.000	0.052	0.082	0.206	0.080

Number of Simulators Used: 4

Remarks:

RECEIVED
AUG 28 2018
FDLE
Alcohol Testing Program

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt Ivette Gomez #180 IVETTE GOMEZ
Signature and Printed Name

08/27/2018
Date

FDLE/ATP Form 40 – March 2004

1261
SP

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.
Time of Inspection: 07:41 Date of Inspection: 08/27/2018

Serial Number: 80-001261
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.046 / 0.043			
0.000	0.047 / 0.045			
0.000	0.044 / 0.052			

Number of Simulators Used: 4

Remarks:
SIMULATOR ISSUE WITH AIR LEAK. SIMULATOR ISSUE, WILL TAKE OFF LINE TO CORRECT PROBLEM.

ISSUE WITH LEAK WAS CORRECTED, WET CONTROL TEST CONDUCTED TO ASSURE RELIABILITY. SIMULATOR BACK ONLINE. TEST WILL BE REPEATED.

SEE ATTACHED PAPERWORK
WET CONTROL TEST

Sgt Ivette Gomez #180

RECEIVED
AUG 28 2018
FDLE
Alcohol Testing Program

The above instrument complies ( ) does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt Ivette Gomez #180 IVETTE GOMEZ
Signature and Printed Name

08/27/2018
Date

SP

ROCKLEDGE P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001261
08/27/2018
Software: 8100.27

| WET CONTROL TEST |

Test	g/210L	Time
Air Blank	0.000	07:53
Control Test	0.052	07:53
Air Blank	0.000	07:54

#180
Operator's Signature

RECEIVED
AUG 28 2018
FDLE
Alcohol Testing Program

SP

ROCKLEDGE P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001261
08/27/2018
Software: 8100.27

| WET CONTROL TEST |

Test	g/210L	Time
Air Blank	0.000	07:56
Control Test	0.052	07:56
Air Blank	0.000	07:57

#180
Operator's Signature

RECEIVED
AUG 28 2018
FDLE
Alcohol Testing Program

8P

ROCKLEDGE P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001261
08/27/2018
Software: 8100.27

| WET CONTROL TEST |

Test	g/210L	Time
Air Blank	0.000	07:58
Control Test	0.053	07:59
Air Blank	0.000	07:59

#180
Operator's Signature

RECEIVED
AUG 28 2018
FDLE
Alcohol Testing Program

1261 SP

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.
Time of Inspection: 09:12
Date of Inspection: 08/27/2018
Serial Number: 80-001261
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:572920 Exp: 09/13/2019
0.000	0.052	0.083	0.205	0.080
0.000	0.052	0.082	0.206	0.080
0.000	0.052	0.082	0.206	0.080

Number of Simulators Used: 4

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt Anthony Rosa #180 IVETTE GOMEZ
Signature and Printed Name

08/27/2018
Date

FDLE/ATP Form 40 – March 2004

1261 8P

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.
Time of Inspection: 08:14 Date of Inspection: 07/31/2018

Serial Number: 80-001261
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:572920 Exp: 09/13/2019
0.000	0.049	0.080	0.196	0.080
0.000	0.050	0.080	0.198	0.080
0.000	0.050	0.081	0.198	0.080

Number of Simulators Used: 4

Remarks:
Time-Date changed.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

180 IVETTE GOMEZ
Signature and Printed Name

07/31/2018
Date

FDLE/ATP Form 40 – March 2004

1261 SP

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.
Time of Inspection: 17:23 Date of Inspection: 06/30/2018

Serial Number: 80-001261
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:572920 Exp: 09/13/2019
0.000	0.051	0.080	0.200	0.080
0.000	0.051	0.079	0.202	0.080
0.000	0.051	0.081	0.202	0.080

Number of Simulators Used: 4

Remarks:

RECEIVED
JUL 19 2018
FDLE
Alcohol Testing Program

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

SGT. #85 PATRICK W KENNEDY
Signature and Printed Name

06/30/2018
Date

FDLE/ATP Form 40 – March 2004

Florida Department of Law Enforcement
Alcohol Testing Program

1201 SR

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.
Serial Number: 80-001261
Time of Inspection: 10:25
Date of Inspection: 05/20/2018
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:572920 Exp: 09/13/2019
0.000	0.050	0.079	0.203	0.079
0.000	0.052	0.080	0.203	0.080
0.000	0.050	0.080	0.203	0.080

Number of Simulators Used: 4

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

SGT #185 PATRICK W KENNEDY
Signature and Printed Name

05/20/2018
Date

FDLE/ATP Form 40 – March 2004

1261 58

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.
Time of Inspection: 12:47 Date of Inspection: 04/30/2018

Serial Number: 80-001261
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:572920 Exp: 09/13/2019
0.000	0.050	0.082	0.205	0.081
0.000	0.051	0.082	0.207	0.080
0.000	0.051	0.082	0.207	0.080

Number of Simulators Used: 4

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

SGT. #785 PATRICK W KENNEDY
Signature and Printed Name

04/30/2018
Date

FDLE/ATP Form 40 – March 2004

1261 SP

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.
Time of Inspection: 04:41 Date of Inspection: 03/12/2018

Serial Number: 80-001261
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:572920 Exp: 09/13/2019
0.000	0.051	0.082	0.206	0.080
0.000	0.051	0.083	0.207	0.079
0.000	0.051	0.083	0.207	0.080

Number of Simulators Used: 4

Remarks:
NONE

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

SGT. 785 PATRICK W KENNEDY
Signature and Printed Name

03/12/2018
Date

FDLE/ATP Form 40 – March 2004

1261 SP

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.
Time of Inspection: 04:41 Date of Inspection: 03/12/2018

Serial Number: 80-001261
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:572920 Exp: 09/13/2019
0.000	0.051	0.082	0.206	0.080
0.000	0.051	0.083	0.207	0.079
0.000	0.051	0.083	0.207	0.080

Number of Simulators Used: 4

Remarks:
NONE

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

SGT. 785 PATRICK W KENNEDY
Signature and Printed Name

03/12/2018
Date

FDLE/ATP Form 40 – March 2004