

1263 SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: INDIAN HARBOUR PD
Time of Inspection: 03:34

Date of Inspection: 08/17/2018

Serial Number: 80-001263
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: 4

Remarks:

A F / M A: Range Exceeded Ambient Fail Range Exceeded Range Ex. Non-compliance: RANGE EXCEEDED MOUTH ALC OHOL.

Too much mouth Alcohol / Second
RECEIVED / Inspection completed
 AUG 21 2018 and in
 FDLE compliance
 Alcohol Testing Program on
 8/17/2018

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MATTHEW S WIMMER
Signature and Printed Name

08/17/2018
Date