

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI-DADE PD
Time of Inspection: 12:42

Date of Inspection: 10/05/2018

Serial Number: 80-001276
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: N 201709A (PE) Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 34917080A4 Exp: 02/05/2020
0.000	0.050	0.078	0.201	0.081
0.000	0.050	0.078	0.201	0.081
0.000	0.050	0.079	0.200	0.081

Number of Simulators Used: 5

Remarks:

AMENDED AI

.08 SOLUTION LOT 201709A EXP 9-12-2019

(PE)
11-8-2018

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Pablo Espinoza

PABLO ESPINOZA

Signature and Printed Name

10/05/2018
Date