

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 20:10

Date of Inspection: 03/27/2018

Serial Number: 80-001286
Software: 8100.27

2

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02618080A2 Exp: 03/05/2020
0.000	0.046	0.076	0.200	0.078
0.000	0.047	0.077	0.202	0.078
0.000	0.047	0.077	0.202	0.079

Number of Simulators Used: 5

Remarks:

Rec'd
4/24/18

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Corey S Wilson

COREY S WILSON

Signature and Printed Name

03/27/2018
Date