

Florida Department of Law Enforcement Alcohol Testing Program

BK
90-001287

AGENCY INSPECTION REPORT - INTOXILYZER 800

Agency: BAKER COUNTY SO
Time of Inspection: 08:50

Date of Inspection: 12/22/2018

Serial Number: 80-001287
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.1g/210L Dry Gas Std Test (g/210L) Lot#:32316080A3 Exp: 01/05/2019
0.000	0.048	0.078	0.197	0.083
0.000	0.048	0.078	0.197	0.083
0.000	0.049	0.079	0.197	0.083

Number of Simulators Used: 5

Remarks:


RECEIVED

DEC 27 2018

BAKER COUNTY SHERIFFS OFFICE

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



COREY S WILSON

Signature and Printed Name

12/22/2018
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 07:33

Date of Inspection: 10/31/2018

Serial Number: 80-001287
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L)	0.08g/210L Test (g/210L)	0.20g/210L Test (g/210L)	0.08 g/210L Dry Gas Std Test (g/210L)
	Lot#: 201802B Exp: 02/06/2020	Lot#: 201709A Exp: 09/12/2019	Lot#: 201802G Exp: 02/22/2020	Lot#: 32316086A3 Exp: 01/05/2019
0.000	0.048	0.077	0.196	0.083
0.000	0.048	0.077	0.197	0.083
0.000	0.049	0.078	0.198	0.082

Number of Simulations Run: 5

Remarks:

RECEIVED

NOV 1 2018

BAKER COUNTY
SHERIFFS OFFICE

The above information complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Corey S Wilson

COREY S WILSON

Signature and Printed Name

10/31/2018
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 23:37

Date of Inspection: 11/15/2018

Serial Number: 02-001287
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 30A3 Exp: 01/05/2019
0.000	0.048	0.077	0.197	0.083
0.000	0.048	0.078	0.197	0.083
0.000	0.048	0.078	0.198	0.082

Number of Simulators Used: 5

Remarks:


RECEIVED

NOV 16 2018

BAKER COUNTY SHERIFFS OFFICE

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



COREY S WILSON

Signature and Printed Name

11/15/2018
Date

Florida Department of Law Enforcement Alcohol Testing Program

FDLE

OCT 19 2018

Alcohol Testing
Program**AGENCY INSPECTION REPORT - INTOXILYZER 8000**Agency: BAKER COUNTY SO
Time of Inspection: 20:18

Date of Inspection: 09/25/2018

Serial Number: 80-001287
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.08g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 32316080A3 Exp: 01/05/2019
0.000	0.047	0.077	0.196	0.082
0.000	0.048	0.078	0.197	0.082
0.000	0.048	0.078	0.197	0.082


Number of Simulators Used: 5

Remarks:

RECEIVED
OCT 04 2018
BAKER COUNTY
SHERIFF'S OFFICE

The above instrument ☒ does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name

COREY S WILSON

09/25/2018
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: **BAKER COUNTY SO**
Time of Inspection: **08:30**

Date of Inspection: **08/19/2018**

Serial Number: **80-001287**
Software: **8100.27**

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 32316080A3 Exp: 01/05/2019
0.000	0.048	0.077	0.196	0.081
0.000	0.048	0.077	0.197	0.081
0.000	0.049	0.078	0.197	0.081

Number of Simulators Used: **5**

Remarks:

RECEIVED

AUG 27 2018

**BAKER COUNTY
SHERIFF'S OFFICE**

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I am a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Corey S. Wilson

Signature and Printed Name

COREY S WILSON

08/19/2018
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

RECEIVED

JUL 31 2018

FDLE
Alcohol Testing Program

Agency: BAKER COUNTY SO
Time of Inspection: 10:18

Date of Inspection: 07/22/2018

Serial Number: 80-001287
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 32316080A3 Exp: 01/05/2019
0.000	0.048	0.076	0.196	0.081
0.000	0.048	0.076	0.197	0.081
0.000	0.048	0.077	0.197	0.080

Number of Simulators Used: 5

Remarks:
Time-Date changed

RECEIVED

JUL 23 2018

BAKER COUNTY
SHERIFF'S OFFICE

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

COREY S WILSON

07/22/2018
Date

Florida Department of Law Enforcement Alcohol Testing Program

7/2/18

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 20:17

Date of Inspection: 06/27/2018

Serial Number: 80-001287
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 32316080A3 Exp: 01/05/2019
0.000	0.049	0.077	0.197	0.081
0.000	0.049	0.078	0.197	0.081
0.000	0.049	0.078	0.197	0.081

Number of Simulators Used: 5

Remarks:

RECEIVED

JUL 02 2018

BAKER COUNTY
SHERIFF'S OFFICE

The above instrument ☒ lies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

COREY S WILSON

06/27/2018
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO

Time of Inspection: 20:17

Date of Inspection: 05/30/2018

Serial Number: 80-001287

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 32316080A3 Exp: 01/05/2019
0.000	049	0.077	0.196	0.081
0.000	049	0.077	0.198	0.081
0.000	049	0.078	0.197	0.080

Number of Simulators Used: 5

Remarks:

RECEIVED

MAY 31 2018

BAKER COUNTY
SHERIFF'S OFFICE

RECEIVED

JUN 05 2018

FDLE
Alcohol Testing Program

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

COREY S WILSON

05/30/2018
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 07:43

Date of Inspection: 04/29/2018

Serial Number: 80-001287
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32316080A3 Exp: 01/05/2019
0.000	0.049	0.078	0.199	0.082
0.000	0.049	0.078	0.200	0.082
0.000	0.050	0.079	0.201	0.082

Number of Simulators Used: 5

Remarks:

RECEIVED

APR 30 2018

BAKER COUNTY
SHERIFF'S OFFICE

RECEIVED

MAY 03 2018

FDLE
Alcohol Testing Program

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Corey S. Wilson

COREY S WILSON

Signature and Printed Name

04/29/2018
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

MAR 06 2018

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: BAKER COUNTY SO
Time of Inspection: 20:43

Date of Inspection: 02/27/2018

Serial Number: 80-001287
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32316080A3 Exp: 01/05/2019
0.000	0.049	0.078	0.196	0.082
0.000	0.049	0.079	0.198	0.082
0.000	0.049	0.078	0.199	0.082

Number of Simulators Used: 5

Remarks:


RECEIVED

MAR 02 2018

BAKER COUNTY
SHERIFF'S OFFICE

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name

COREY S WILSON

02/27/2018
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 14:50

Date of Inspection: 01/31/2018

Serial Number: 80-001287
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 32316080A3 Exp: 01/05/2019
0.000	0.049	0.079	0.199	0.082
0.000	0.049	0.079	0.199	0.082
0.000	0.050	0.079	0.199	0.082

Number of Simulators Used: 5

Remarks:

RECEIVED

FEB 01 2018

BAKER COUNTY
SHERIFF'S OFFICE

RECEIVED

FEB 12 2018

FDLE
Alcohol Testing Program

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name

COREY S WILSON

01/31/2018
Date