

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MARIANNA PD

Time of Inspection: 13:37

Date of Inspection: 07/30/2018

Serial Number: 80-001305

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) <i>201702B</i> Lot#: 95-100320 Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 906364 Exp: 02/07/2021
0.000	0.048	0.077	0.194	0.082
0.000	0.049	0.077	0.196	0.081
0.000	0.048	0.078	0.196	0.082

Number of Simulators Used: 5

Remarks: *0.20 Solution used product number instead of lot number
Lot # 201702B*

RECEIVED
 AUG 22 2018
FDLE
 Alcohol Testing Program

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Tyler Scarborough _____ TYLER J SCARBOROUGH
 Signature and Printed Name

07/30/2018
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MARIANNA PD

Time of Inspection: 13:32

Date of Inspection: 06/14/2018

Serial Number: 80-001305

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 906364 Exp: 02/07/2021
0.000	0.048	0.078	0.198	0.081
0.000	0.048	0.077	0.198	0.081
0.000	0.048	0.077	0.198	0.081

Number of Simulators Used: 5

Remarks:
NA

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



SUSAN M BARGE

Signature and Printed Name

06/14/2018
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MARIANNA PD

Time of Inspection: 13:42

Date of Inspection: 05/08/2018

Serial Number: 80-001305

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 201709A Exp: 09/12/2019
0.000	0.049	0.076	0.195	0.082
0.000	0.049	0.078	0.197	0.082
0.000	0.049	0.078	0.198	0.082

Number of Simulators Used:

Remarks:

Time-Date

SMB

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Susan M. Barber

Signature and Printed Name

SUSAN M. BARBER

05/08/2018
Date