

Florida Department of Law Enforcement
Alcohol Testing Program

RECEIVED
APR 22 2018
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: INDIAN RIVER CO. SO

Serial Number: 80-001328

Time of Inspection: 05:48


Date of Inspection: 03/23/2018

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: 684131 Exp: 01/23/2020
0.000	0.050	0.080	0.200	0.080
0.000	0.051	0.081	0.200	0.080
0.000	0.050	0.080	0.201	0.080

Number of Simulators Used: 4

Remarks: Printer failed to print inspection report prior to Com Transfer. It did print, but it was a blank page.  4-20-19

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MARK A KANTORSKI 

Signature and Printed Name

03/23/2018

Date