

1347
SF

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SARASOTA COUNTY SO
Time of Inspection: 07:39

Date of Inspection: 08/21/2018

Serial Number: 80-001347
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:10818080A4 Exp: 06/05/2020
0.000	0.048	0.078	0.195	0.070 / 0.071
0.000	0.048	0.077	0.195	0.070 / 0.071
0.000	0.048	0.078	0.195	0.070 / 0.071

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: CONTROL OUTSIDE TOL.

CHECKED ALL EQUIPMENT & SETUP. DID NOT OBSERVE ANY ISSUES. DISCONNECTED / RECONNECTED ALL AND RE-TESTED.

RECEIVED
AUG 21 2018
FDLE
Alcohol Testing Program

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DEP. M. FELTOVIC

DEP M FELTOVIC

Signature and Printed Name

08/21/2018
Date