

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PUTNAM COUNTY SO  
Time of Inspection: 13:26

Date of Inspection: 03/02/2018

Serial Number: 80-001648  
Software: 8100.27

*2*

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:29017080AJ Exp: 12/05/2019
0.000	0.046	0.074 / 0.076	0.191	0.079
0.000	0.047	0.075 / 0.077	0.193	0.079
0.000	0.047	0.076 / 0.077	0.193	0.079

Number of Simulators Used: 5

Remarks:  
: Control Outside Tolerance. Simulator temp was Low  
*MLR*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*[Signature]* 5187  
\_\_\_\_\_  
Signature and Printed Name

KEENE W. B. BOWERS

03/02/2018  
Date