

1653
SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ST PETERSBURG PD
Time of Inspection: 09:50

Date of Inspection: 11/20/2018

Serial Number: 80-001653
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 26316080A2 Exp: 11/05/2018 *
0.000	0.051	0.081	0.200	0.079
0.000	0.051	0.080	0.199	0.079
0.000	0.050	0.080	0.200	0.079

Number of Simulators Used: 5

Remarks:
IS NOT IN COMPLIANCE MW

* EXPIRED DRY GAS

RECEIVED
NOV 28 2018
FDLE
Alcohol Testing Program

The above instrument complies () does not comply () with Chapter 11D-8, FAC. ^{MW}

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name MICHAEL D WEISKOPF

11/20/2018
Date