Florida Department of Law Enforcement
Alcohol Testing Program

1/24/19

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 14:35
Date of Inspection: 12/19/2018
Serial Number: 80-001865
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22817080A5 Exp: 10/05/2019
0.000	0.045	0.077	* / 0.194	0.076
0.000	0.045	0.079	/ 0.193	0.077
0.000	0.046	0.079	/ 0.193	0.076

Number of Simulators Used: 5

Remarks:
* 20: Ambient Fail.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gordon L. Franz
GORDON L FRANZ
Signature and Printed Name

12/19/2018
Date

80-001865

RECEIVED
NOV 07 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 10:43 Date of Inspection: 11/06/2018 Serial Number: 80-001865
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	/ 0.048	0.078	0.191	
0.000	/ 0.047	0.077	0.189 *	
0.000	/ 0.047	0.077	0.190	

Number of Simulators Used: 5

Remarks:
05: Ambient Fail. 20: Control Outside Tolerance, Ambient FailUNABLE TO CLEAR SAMPLE CHAMBER PRIOR TO CONTROL TESTNon-compliance: .

* .20 SIMULATOR DRAWING AMBIENT AIR - POSSIBLE SEAL / HOSE

The above instrument complies ( ) does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gordon L. Franz
GORDON L FRANZ
Signature and Printed Name

11/06/2018
Date

FDLE/ATP Form 40 – March 2004

80-001865 NB

RECEIVED
OCT 17 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 15:11
Date of Inspection: 10/15/2018
Serial Number: 80-001865
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22817080A5 Exp: 10/05/2019
0.000	0.045	0.076	0.188 / 0.192	0.077
0.000	0.046	0.077	0.191 / 0.194	0.077
0.000	0.048	0.078	0.193 / 0.194	0.077

Number of Simulators Used: 5

Remarks:
20: Control Outside Tolerance. POSSIBLE Cold Tubing, .20 Alcohol REF. Solution, 2ND TEST within Limits - PASSED GLF

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gordon L. Franz
Signature and Printed Name GORDON L FRANZ

10/15/2018
Date

FDLE/ATP Form 40 – March 2004

80-001865

RECEIVED
SEP 24 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 11:06
Date of Inspection: 09/24/2018
Serial Number: 80-001865
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22817080A5 Exp: 10/05/2019
0.000	0.048	0.079	0.195	0.078
0.000	0.049	0.079	0.196	0.078
0.000	0.050	0.080	0.197	0.077

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gordon L Franz
GORDON L FRANZ
Signature and Printed Name

09/24/2018
Date

FDLE/ATP Form 40 – March 2004

80-001865 2018

RECEIVED
AUG 21 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

COPY

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 14:35 Date of Inspection: 08/20/2018
Serial Number: 80-001865
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22817080A5 Exp: 10/05/2019
0.000	0.049	0.077	0.196	0.077
0.000	0.049	0.078	0.197	0.077
0.000	0.049	0.078	0.197	0.077

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gordon L. Franz
GORDON L FRANZ
Signature and Printed Name

08/20/2018
Date

80-001865 DWB

RECEIVED
AUG 21 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

COPY

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 11:51
Date of Inspection: 08/20/2018
Serial Number: 80-001865
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22817080A5 Exp: 10/05/2019
0.000	0.046	0.075	0.193	0.077
0.000	0.047	0.076	0.195	0.077
0.000	0.048	0.077	0.196	0.077

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gordon L. Franz
GORDON L FRANZ
Signature and Printed Name

08/20/2018
Date

80-001865

RECEIVED
JUL 16 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 09:28 Date of Inspection: 07/16/2018
Serial Number: 80-001865
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22817080A5 Exp: 10/05/2019
0.000	0.048	0.078	0.196	0.077
0.000	0.049	0.079	0.198	0.077
0.000	0.049	0.079	0.199	0.077

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gordon L. Franz
Signature and Printed Name

GORDON L FRANZ

07/16/2018
Date

80-001865 Rom

RECEIVED
MAR 19 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 10:00 Date of Inspection: 03/19/2018

Serial Number: 80-001865
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22817080A5 Exp: 10/05/2019
0.000	0.047	0.076	0.190	0.078
0.000	0.048	0.077	0.193	0.078
0.000	0.048	0.077	0.195	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gordon L. Franz
GORDON L FRANZ
Signature and Printed Name

03/19/2018
Date

80-001865

RECEIVED
FEB 26 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 09:50
Date of Inspection: 02/26/2018
Serial Number: 80-001865
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22817080A5 Exp: 10/05/2019
0.000 / 0.000	0.049	0.076	0.194	0.077
INT / 0.000	0.048	0.077	0.194	0.077
/ 0.000	0.048	0.078	0.196	0.077

Number of Simulators Used: 5

Remarks:
00: Interferent Detect.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gordon L Franz
GORDON L FRANZ
Signature and Printed Name

02/26/2018
Date

80-001865 RWB

RECEIVED
JAN 23 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 10:15
Date of Inspection: 01/23/2018
Serial Number: 80-001865
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22817080A5 Exp: 10/05/2019
0.000	0.049	0.078	0.197	0.077
0.000	0.050	0.079	0.198	0.077
0.000	0.049	0.080	0.198	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gordon L. Franz
GORDON L FRANZ
Signature and Printed Name

01/23/2018
Date

FDLE/ATP Form 40 – March 2004