

5936
SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SARASOTA COUNTY SO

Serial Number: 80-005936

Time of Inspection: 09:23

Date of Inspection: 09/30/2018

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: 19817080A2 Exp: 09/05/2019
0.000	0.049	0.077	0.194	0.080
0.000	0.049	0.075	0.195	0.080
0.000	0.050	0.075	0.196	0.080

Number of Simulators Used: 5

Remarks:

INSPECTED AT N JAIL *SECOND INSPECTION PERFORMED DUE TO PRINTER ERROR AFTER FIRST INSPECTION. @1965*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Dep M. Feltovic 1965

DEP M FELTOVIC

Signature and Printed Name

09/30/2018

Date