

80-006567 ~~808~~

RECEIVED

JAN 04 2018

FDLE  
Alcohol Testing Program

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO  
Time of Inspection: 18:05

Date of Inspection: 01/03/2018

Serial Number: 80-006567  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:03417080A1 Exp: 03/05/2019
0.000	0.049	0.079	0.197	0.080
0.000	0.049	0.080	0.197	0.080
0.000	0.049	0.080	0.198	0.080

Number of Simulators Used: 5

Remarks:  
AI R. GRANNIS #257-256794 PRESENT ENTIRE INSPECTION. *2/4/18*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
Signature and Printed Name

ROGER G SKIPPER

01/03/2018  
Date

*REVISED  
1-3-18  
M/Sgt. J.L. Bell  
4/18*