

6/14/18
9

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP
Time of Inspection: 03:10

Date of Inspection: 05/25/2018

Serial Number: 80-006629
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201709A Exp: 03/09/2019	0.08g/210L Test (g/210L) Lot#:201811E Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201802B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:2017080AS Exp: 10/03/2018
/ 0.000	0.047	0.076	0.197	0.079
/ 0.000	0.048	0.077	0.199	0.079
/ 0.000	0.049	0.078	0.199	0.079

Number of Simulators Used: 4

Remarks:

dry Ambient Fail.

in possible alcohol bottle open new instrument

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JACOB A WASHIN

Signature and Printed Name

05/25/2018
Date