

6631 SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO
Time of Inspection: 09:02

Date of Inspection: 09/25/2018

Serial Number: 80-006631
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 962587 Exp: 06/04/2021
0.000	0.047	0.074 / 0.077	0.194	0.080
0.000	0.048	0.077 / 0.077	0.197	0.080
0.000	0.049	0.077 / 0.077	0.197	0.080

Number of Simulators Used: 5

Remarks:

Time-Date changed. 08: Control Outside Tolerance.

SIT-SPURLOCK PRESENT

Replaced "O" ring on connection, problem resolved.
(simulator)

RECEIVED
OCT 25 2018
FDLE
Alcohol Testing Program

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GLENN CLINE

Signature and Printed Name

09/25/2018
Date

6631
SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO
Time of Inspection: 09:02

Date of Inspection: 09/25/2018

Serial Number: 80-006631
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 962587 Exp: 06/04/2021
0.000	0.047	0.074 / 0.077	0.194	0.080
0.000	0.048	0.077 / 0.077	0.197	0.080
0.000	0.049	0.077 / 0.077	0.197	0.080

Number of Simulators Used: 5

Remarks:


Time-Date changed. 08: Control Outside Tolerance.

Sgt. Spurlack present 

RECEIVED
SEP 24 2018
FDLE
Alcohol Testing Program

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



GLENN CLINE

Signature and Printed Name

09/25/2018
Date