

6/14/18 (P)

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 03:24

Date of Inspection: 05/25/2018

Serial Number: 80-006635
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22817080A5 Exp: 10/05/2019
/ 0.000	0.044 / 0.047	0.076	0.195	0.078
/ 0.000	0.045 / 0.047	0.075	0.196	0.078
/ 0.000	0.046 / 0.047	0.075	0.196	0.078

Number of Simulators Used: 4

Remarks:

00: Ambient Fail, 05: Control Outside Tolerance.

00: Mouth Alcohol bottle separate via no measurement
05: Simulator not up to temp

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JACOB A VAUGHN

Signature and Printed Name

05/25/2018
Date