

6692 SP

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HOLMES BEACH PD  
Time of Inspection: 09:05

Date of Inspection: 07/23/2018

Serial Number: 80-006692  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 17316080A2 Exp: 08/05/2018
0.000	0.047	0.076	0.194	0.079
0.000	0.048	0.077	0.195	0.079
0.000	0.048	0.077	0.196	0.079

Number of Simulators Used: 5

Remarks:


WITH MSO DEPUTY CLINE

Fixed LOT # 201802B for .05/210 L TEST

RECEIVED  
AUG 20 2018  
FDLE  
Alcohol Testing Program

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 MICHAEL PILATO  
Signature and Printed Name

07/23/2018  
Date