

6694 SP

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO  
Time of Inspection: 09:12

Date of Inspection: 09/25/2018

Serial Number: 80-006694  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 962587 Exp: 06/04/2021
0.000	0.049	0.075	0.195	0.000 / 0.081
0.000	0.049	0.077	0.196	0.000 / 0.081
0.000	0.049	0.077	0.196	0.080 / 0.081

Number of Simulators Used: 5

Remarks:

Time-Date changed. 08: Control Outside Tolerance.

SLT SPOLOCK PRESENT ~~OK~~

08 DRY GAS DISCONNECTED

RECEIVED

OCT 25 2018

FDLE  
Alcohol Testing Program

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

GLENN CLINE

09/25/2018  
Date

6694  
SP

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO  
Time of Inspection: 09:12

Date of Inspection: 09/25/2018

Serial Number: 80-006694  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201709A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:962587 Exp: 06/04/2021
0.000	0.049	0.075	0.195	0.000 / 0.081
0.000	0.049	0.077	0.196	0.000 / 0.081
0.000	0.049	0.077	0.196	0.080 / 0.081

Number of Simulators Used: 5

Remarks:

Time-Date changed. 08: Control Outside Tolerance.


SLT SPURLOCK PRESENT 

08 dry gas disconnected

RECEIVED  
SEP 24 2018  
FDLE  
Alcohol Testing Program

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



GLENN CLINE

Signature and Printed Name

09/25/2018  
Date

6694 SP

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO  
Time of Inspection: 09:35

Date of Inspection: 07/24/2018

Serial Number: 80-006694  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AJ 256835 Exp: 09/19/2020
0.000	0.047	0.075	0.188 / 0.195	0.081
0.000	0.047	0.076	0.194 / 0.196	0.081
0.000	0.047	0.077	0.194 / 0.196	0.081

Number of Simulators Used: 5

Remarks:

20: Control Outside Tolerance. SGT SPURLOCK PRESENT

Sergeant in list Lot# on DOT GAS during inspection

**RECEIVED**  
AUG 16 2018  
FDLE  
Alcohol Testing Program

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GLENN CLINE  
Signature and Printed Name

07/24/2018  
Date