



INSTRUMENT PROCESSING SHEET

Agency Florida Department of Law Enforcement S/N 80-000224

Florida Department of Law Enforcement

Date In N/A DI Completion Date 04/27/2018 Ship P/U H/D CMI EE

Intake Performed By <u>DMB</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____ Final Release Date <div style="text-align: center; font-weight: bold; font-size: 1.2em;">FDLE</div> <div style="text-align: center; font-weight: bold; font-size: 1.2em;">APR 27 2018</div> <div style="text-align: center; font-weight: bold;">Alcohol Testing Program</div>	Quality Checks Performed By <u>DMB</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>218</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 102</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.496</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>G2835</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>G2840</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>SD1025</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG805702 02/26/2020</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	G2835	201707D 07/25/2019	0.080	G2840	201707E 07/25/2019	0.200	SD1025	201707C 07/24/2019	0.080 DGS	N/A	AG805702 02/26/2020	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.050	G2835	201707D 07/25/2019															
0.080	G2840	201707E 07/25/2019															
0.200	SD1025	201707C 07/24/2019															
0.080 DGS	N/A	AG805702 02/26/2020															
		Maintenance Performed By <u>DMB</u> <input type="checkbox"/> Battery Replacement <input checked="" type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>DMB</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.0</u> External Digital Therm. ID#: <u>300502</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>G2835</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>G2840</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1025</u>															

Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>DMB</u> Barometric Pressure ID# <u>28427</u> Gauge <u>1011</u> Instrument <u>1009</u> Mouth Alcohol Solution Lot # <u>2017-B</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G2880</td> </tr> <tr> <td>Interferent</td> <td>G2407</td> </tr> <tr> <td>0.050</td> <td>G2835</td> </tr> <tr> <td>0.080</td> <td>G2840</td> </tr> <tr> <td>0.200</td> <td>SD1025</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	G2880	Interferent	G2407	0.050	G2835	0.080	G2840	0.200	SD1025
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	G2880																																																												
Interferent	G2407																																																												
0.050	G2835																																																												
0.080	G2840																																																												
0.200	SD1025																																																												
Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____																																																													

Notes/Suggested Service: <u>Dry gas regulator replaced after determining DGS reading 0.000. DMB 4/26/18</u> _____ _____ _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="margin-top: 20px;"> <div style="display: flex; justify-content: space-between;"> <div style="text-align: center;"> Tech-Review / Date <u>4/27/18</u> </div> <div style="text-align: center;"> Admin Review / Date <u>4/27/18</u> </div> </div> </div>
---	--

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE
Time of Inspection: 11:46

Date of Inspection: 04/27/2018

Serial Number: 80-000224
Software: 8100.27

see

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805702 Exp: 02/26/2020
0.000	0.049	0.081	0.207 / 0.205	0.077
0.000	0.050	0.081	/ 0.206	0.078
0.000	0.051	0.082	/ 0.205	0.078
0.000	0.051	0.082	/ 0.205	0.078
0.000	0.051	0.082	/ 0.206	0.078
0.000	0.051	0.083	/ 0.205	0.078
0.000	0.051	0.082	/ 0.206	0.078
0.000	0.051	0.083	/ 0.205	0.077
0.000	0.051	0.082	/ 0.205	0.077
0.000	0.052	0.083	/ 0.205	0.077
Standard Deviations	0.0007	0.0007	/ 0.0004	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

20: Ambient Fail SIMULATOR TURNED OVER PRIOR TO CONNECTING .

(Handwritten mark)

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

(Handwritten Signature)

DANIELLE M BELL

Signature and Printed Name

04/27/2018
Date

*4/27/18
JD*

Stability Checks #80-000224 FDLE 4/26/18 ~~DOB~~ pg 1 of 2

DOB

FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000224
04/26/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:08
Control Test	0.049	14:09
Air Blank	0.000	14:10
Control Test	0.049	14:10
Air Blank	0.000	14:11
Control Test	0.049	14:12
Air Blank	0.000	14:12
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel. Std Dev(%)	0.0000	

FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000224
04/26/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:14
Control Test	0.080	14:15
Air Blank	0.000	14:15
Control Test	0.081	14:16
Air Blank	0.000	14:17
Control Test	0.081	14:17
Air Blank	0.000	14:18
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel. Std Dev(%)	0.7157	

FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000224
04/26/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:32
Control Test	0.200	14:32
Air Blank	0.000	14:33
Control Test	0.201	14:34
Air Blank	0.000	14:34
Control Test	0.201	14:35
Air Blank	0.000	14:36
Control Test Stats		
Average	0.2007	
Std Dev	0.0006	
Rel. Std Dev(%)	0.2877	

FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000224
04/26/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:28
Control Test	0.081	14:28
Air Blank	0.000	14:29
Control Test	0.079	14:29
Air Blank	0.000	14:30
Control Test	0.081	14:30
Air Blank	0.000	14:30
Control Test Stats		
Average	0.0803	
Std Dev	0.0012	
Rel. Std Dev(%)	1.4374	

4/27/18
DOB
Operator's Signature

DOB
Operator's Signature

DOB
Operator's Signature

DOB
Operator's Signature

Stability Checks #80-000224 FILE 3/26/18 RMB PG 2 of 2

RMB

FILE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000224
04/26/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:19
Control Test	0.000	14:19
Air Blank	0.000	14:20
Control Test	0.079	14:20
Air Blank	0.000	14:20
Control Test	0.000	14:21
Air Blank	0.000	14:21
Control Test Stats		
Average	0.0263	
Std Dev	0.0456	
Rel. Std Dev(%)	173.2051	

Resetted DA regulator plug. RMB

RMB

Operator's Signature

4/27/18
Z

RMB

FILE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000224
04/26/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:22
Control Test	0.000	14:23
Air Blank	0.000	14:23
Control Test	0.000	14:23
Air Blank	0.000	14:24
Control Test	0.000	14:24
Air Blank	0.000	14:25
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel. Std Dev(%)	0.0000	

Replaced regulator RMB

RMB

Operator's Signature



Florida Department of Law Enforcement
 Alcohol Testing Program
 2729 Fort Knox Blvd.
 Bldg. 2, Suite 1300
 Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000224, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000224</u>	UNCERTAINTY* ±	
Owning Agency:	<u>FDLE</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>04/27/2018</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>11:46</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
 Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
 *Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.
 Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 March 2018
 Issuing Authority: Alcohol Testing Program

04/27/2018 _____
 Date

 DANIELLE M BELL,
 Department Inspector

Service • Integrity • Respect • Quality

4/27/18

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE
Time of Inspection: 14:07

Date of Inspection: 04/26/2018

Serial Number: 80-000224
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:
BYPASSED AI TO OPERATE INSTRUMENT. NOT A COMPLIANCE CHECK.

[Handwritten mark]

*N/A Compliance
Not Determined
[Signature]*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Handwritten Signature]

DANIELLE M BELL

Signature and Printed Name

04/26/2018
Date

*4/27/18
[Signature]*