



INSTRUMENT PROCESSING SHEET

Agency FDLE

S/N 80-000229

Florida Department of Law Enforcement

Date In -----

DI Completion Date 6/14/18

Ship P/U H/D CMI EE

Intake Performed By <u>TSK</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>TSK</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>170</u> <input type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 102</u> 32 mm <u>156</u> (.139 - .169) 36 mm <u>171</u> (.156 - .190) 53 mm <u>242</u> (.228 - .278) 103 mm <u>507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1021</u></td> <td>201707D 7/25/2019</td> </tr> <tr> <td>0.080</td> <td><u>DR1275</u></td> <td>201707E 7/25/2019</td> </tr> <tr> <td>0.200</td> <td><u>SD1019</u></td> <td>201707C 7/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG805702</u> <u>2/26/20</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>SD1021</u>	201707D 7/25/2019	0.080	<u>DR1275</u>	201707E 7/25/2019	0.200	<u>SD1019</u>	201707C 7/24/2019	0.080 DGS	N/A	<u>AG805702</u> <u>2/26/20</u>	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	<u>SD1021</u>	201707D 7/25/2019																																																											
0.080	<u>DR1275</u>	201707E 7/25/2019																																																											
0.200	<u>SD1019</u>	201707C 7/24/2019																																																											
0.080 DGS	N/A	<u>AG805702</u> <u>2/26/20</u>																																																											
Final Release Date <p style="text-align: center;">FDLE</p> <p style="text-align: center;">JUN 14 2018</p> <p style="text-align: center;">Alcohol Testing Program</p>	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____	Temperature Checks Performed By <u>TSK</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.58</u> External Digital Therm. ID#: <u>300503</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1021</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>DR1275</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1019</u>																																																											
Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>TSK</u> Barometric Pressure ID# <u>28427</u> Gauge <u>1016</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2017-B</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G11621</u></td> </tr> <tr> <td>Interferent</td> <td><u>DR3855</u></td> </tr> <tr> <td>0.050</td> <td><u>SD1021</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1275</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1019</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	<u>G11621</u>	Interferent	<u>DR3855</u>	0.050	<u>SD1021</u>	0.080	<u>DR1275</u>	0.200	<u>SD1019</u>
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	<u>G11621</u>																																																												
Interferent	<u>DR3855</u>																																																												
0.050	<u>SD1021</u>																																																												
0.080	<u>DR1275</u>																																																												
0.200	<u>SD1019</u>																																																												
Notes/Suggested Service: _____ _____ _____ _____ _____	Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____																																																												
<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use		<div style="display: flex; justify-content: space-between;"> <div style="text-align: center;"> <u>6/14/18</u> Tech Review / Date </div> <div style="text-align: center;"> <u>6/14/18</u> Admin Review / Date </div> </div>																																																											

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE
Time of Inspection: 12:40

Date of Inspection: 06/14/2018

Serial Number: 80-000229
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805702 Exp: 02/26/2020
0.000	0.048	0.079	0.196	0.077
0.000	0.048	0.079	0.197	0.077
0.000	0.049	0.079	0.197	0.077
0.000	0.049	0.080	0.197	0.077
0.000	0.048	0.079	0.197	0.077
0.000	0.048	0.079	0.196	0.077
0.000	0.049	0.079	0.197	0.077
0.000	0.049	0.080	0.196	0.077
0.000	0.049	0.080	0.196	0.077
0.000	0.049	0.079	0.196	0.077

Standard Deviations	0.0005	0.0004	0.0005	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:



The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Brett Kirkland

BRETT H KIRKLAND

Signature and Printed Name

06/14/2018
Date

6/14/18
JS

Stability Checks
FDLE
80-000229
6/14/18 BK

FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000229
06/14/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:21
Control Test	0.047	10:22
Air Blank	0.000	10:23
Control Test	0.048	10:23
Air Blank	0.000	10:24
Control Test	0.047	10:25
Air Blank	0.000	10:25
Control Test Stats		
Average	0.0473	
Std Dev	0.0006	
Rel. Std Dev(%)	1.2198	

Brett Kirkland
Operator's Signature

6/14/18
JD

FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000229
06/14/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:27
Control Test	0.077	10:28
Air Blank	0.000	10:29
Control Test	0.078	10:29
Air Blank	0.000	10:30
Control Test	0.079	10:30
Air Blank	0.000	10:31
Control Test Stats		
Average	0.0780	
Std Dev	0.0010	
Rel. Std Dev(%)	1.2821	

Brett Kirkland
Operator's Signature

FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000229
06/14/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:32
Control Test	0.195	10:33
Air Blank	0.000	10:34
Control Test	0.195	10:34
Air Blank	0.000	10:35
Control Test	0.196	10:36
Air Blank	0.000	10:36
Control Test Stats		
Average	0.1953	
Std Dev	0.0006	
Rel. Std Dev(%)	0.2956	

Brett Kirkland
Operator's Signature

FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000229
06/14/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:38
Control Test	0.078	10:38
Air Blank	0.000	10:39
Control Test	0.078	10:39
Air Blank	0.000	10:40
Control Test	0.078	10:40
Air Blank	0.000	10:40
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel. Std Dev(%)	0.0000	

Brett Kirkland
Operator's Signature

DGS

(Signature)



Florida Department of Law Enforcement
 Alcohol Testing Program
 2729 Fort Knox Blvd.
 Bldg. 2, Suite 1300
 Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000229, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000229</u>	UNCERTAINTY* ±	
Owning Agency:	<u>FDLE</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/14/2018</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>12:40</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
 Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
 *Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.
 Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/14/2018

Date

Brett Kirkland

BRETT H KIRKLAND,
 Department Inspector

FDLE/ATP Form 69 March 2018
 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

6/14/18
JD

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE
Time of Inspection: 09:49

Date of Inspection: 06/14/2018

Serial Number: 80-000229
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:
BYPASSED TO OPERATE INSTRUMENT

not a compliance check
OK

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Brett Kirkland

BRETT H KIRKLAND

Signature and Printed Name

06/14/2018
Date

6/14/18
JB