



INSTRUMENT PROCESSING SHEET

Agency Hernando County Sheriff's Office S/N 80-000814

Florida Department of Law Enforcement

Date In 05/16/2018 DI Completion Date 05/17/2018 Ship P/U H/D CMI EE

Intake Performed By <u>TG</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>DMB</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>184</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 103</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																												
Final Release Date <div style="text-align: center; font-weight: bold; font-size: 1.2em;">FDLE</div> <div style="text-align: center; font-weight: bold; font-size: 1.2em;">MAY 17 2018</div> <div style="text-align: center; font-weight: bold;">Alcohol Testing Program</div>	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>G2835</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>G2840</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>SD1025</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG805702 02/26/2020</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	G2835	201707D 07/25/2019	0.080	G2840	201707E 07/25/2019	0.200	SD1025	201707C 07/24/2019	0.080 DGS	N/A	AG805702 02/26/2020	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>DMB</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.1</u> External Digital Therm. ID#: <u>300502</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>G2835</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>G2840</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1025</u>																																													
Simulator	Serial #	Lot #/Exp																																																												
0.050	G2835	201707D 07/25/2019																																																												
0.080	G2840	201707E 07/25/2019																																																												
0.200	SD1025	201707C 07/24/2019																																																												
0.080 DGS	N/A	AG805702 02/26/2020																																																												
Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>DMB</u> Barometric Pressure ID# <u>28427</u> Gauge <u>1010</u> Instrument <u>1011</u> Mouth Alcohol Solution Lot # <u>2017-B</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G2880</td> </tr> <tr> <td>Interferent</td> <td>G2407</td> </tr> <tr> <td>0.050</td> <td>G2835</td> </tr> <tr> <td>0.080</td> <td>G2840</td> </tr> <tr> <td>0.200</td> <td>SD1025</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	G2880	Interferent	G2407	0.050	G2835	0.080	G2840	0.200	SD1025	Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____
Simulator	Serial Number	Lot Number	Expiration																																																											
0.000		N/A	N/A																																																											
0.040																																																														
0.100																																																														
0.200																																																														
0.300																																																														
0.080 DGS	N/A																																																													
Simulator	Serial Number	Lot Number	Expiration																																																											
0.050																																																														
0.080																																																														
0.200																																																														
0.080 DGS	N/A																																																													
Simulator	Serial Number																																																													
0.000	G2880																																																													
Interferent	G2407																																																													
0.050	G2835																																																													
0.080	G2840																																																													
0.200	SD1025																																																													
Notes/Suggested Service: <u>Please change level 2 menu password to something unique. DMB 5/17/18</u> _____ _____ _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="display: flex; justify-content: space-between; margin-top: 10px;"> <div style="text-align: center;"> <u>5/17/18</u> Tech Review / Date </div> <div style="text-align: center;"> <u>5/17/18</u> Admin Review / Date </div> </div>																																																													

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 11:02

Date of Inspection: 05/17/2018

Serial Number: 80-000814
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805702 Exp: 02/26/2020
0.000	0.048	0.079	0.194	0.082
0.000	0.049	0.080	0.196	0.082
0.000	0.049	0.080	0.196	0.082
0.000	0.048	0.080	0.197	0.082
0.000	0.049	0.080	0.196	0.082
0.000	0.049	0.080	0.196	0.082
0.000	0.049	0.080	0.196	0.082
0.000	0.049	0.080	0.197	0.082
0.000	0.049	0.080	0.196	0.082
0.000	0.049	0.081	0.196	0.082

Standard Deviations	0.0004	0.0004	0.0008	0.0000
---------------------	--------	--------	--------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:



The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DANIELLE M BELL

Signature and Printed Name

05/17/2018
Date

5/17/18
DB

Stability Checks #80-000814 Hernando County S.O. 5/17/18 *RMS*

DG3

HERNANDO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000814
05/17/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:06
Control Test	0.049	09:08
Air Blank	0.000	09:09
Control Test	0.049	09:10
Air Blank	0.000	09:10
Control Test	0.049	09:11
Air Blank	0.000	09:11
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

RMS
Operator's Signature
5/17/18

HERNANDO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000814
05/17/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:12
Control Test	0.080	09:13
Air Blank	0.000	09:14
Control Test	0.080	09:14
Air Blank	0.000	09:15
Control Test	0.081	09:15
Air Blank	0.000	09:16
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

*Confirming cold tubing
- Keran RMS*

RMS
Operator's Signature

HERNANDO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000814
05/17/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:17
Control Test	0.194	09:18
Air Blank	0.000	09:18
Control Test	0.195	09:19
Air Blank	0.000	09:19
Control Test	0.196	09:20
Air Blank	0.000	09:21
Control Test Stats		
Average	0.1950	
Std Dev	0.0010	
Rel Std Dev(%)	0.5128	

*Confirming cold tubing
- Keran RMS*

RMS
Operator's Signature

HERNANDO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000814
05/17/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:25
Control Test	0.197	09:26
Air Blank	0.000	09:27
Control Test	0.197	09:27
Air Blank	0.000	09:28
Control Test	0.197	09:28
Air Blank	0.000	09:29
Control Test Stats		
Average	0.1970	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

RMS
Operator's Signature

HERNANDO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000814
05/17/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:22
Control Test	0.080	09:22
Air Blank	0.000	09:22
Control Test	0.082	09:23
Air Blank	0.000	09:23
Control Test	0.082	09:23
Air Blank	0.000	09:24
Control Test Stats		
Average	0.0813	
Std Dev	0.0012	
Rel Std Dev(%)	1.4197	

RMS
Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000814, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000814</u>	UNCERTAINTY* ±
Owning Agency:	<u>HERNANDO COUNTY S.O.</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>05/17/2018</u>	0.080 g/ 210 L 0.005
Calibration Time:	<u>11:02</u>	0.200 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

05/17/2018

Date

DANIELLE M BELL,
Department Inspector

FDLE/ATP Form 69 March 2018
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

5/17/18
JD



INSTRUMENT PROCESSING SHEET

Agency Hernando CountyS/N 80-000814

Florida Department of Law Enforcement

Date In 11/15/17DI Completion Date 01/08/2018 Ship P/U H/D CMI EE

Intake Performed By <u>DP</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>DP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>187</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 102</u> 32 mm <u>0.164</u> (.139 - .169) 36 mm <u>0.179</u> (.156 - .190) 53 mm <u>0.250</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28662</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
--	--	--

Final Release Date

FDLE

JAN 12 2018

Alcohol Testing Program

Simulator	Serial #	Lot #/Exp
0.050	SD3962	201707D 07/25/2019
0.080	SD3964	201707E 07/25/2019
0.200	DR3856	201707C 07/24/2019
0.080 DGS	N/A	AG626604 09/22/2018

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Temperature Checks Performed By DP

Lab Temp °C 21.5
 External Digital Therm. ID#: 300505
 34°C +/-2 Serial #: SD3962
 34°C +/-2 Serial #: SD3964
 34°C +/-2 Serial #: DR3856

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By DP

Barometric Pressure ID# 28662
 Gauge 1021 Instrument 1021
 Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.000	SD1019
Interferent	SD1021
0.050	SD3962
0.080	SD3964
0.200	DR3856

Attachments

Form 41 Flow Calibration
 Stability Checks Form 40
 Calibration Adjustment Other Calibration Certificate
 Post-Stability Checks

Notes/Suggested Service: Please remember to update the level 2 menu password to something unique.
Thanks. DMB 1/8/18

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

SP 1/8/18 J. Dahan 1/12/18
 Tech Review / Date Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 13:50

Date of Inspection: 01/08/2018

Serial Number: 80-000814
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626604 Exp: 09/22/2018
0.000	0.050	0.082	0.196	0.082
0.000	0.050	0.082	0.198	0.081
0.000	0.050	0.081	0.197	0.081
0.000	0.051	0.082	0.198	0.081
0.000	0.050	0.082	0.198	0.081
0.000	0.050	0.081	0.198	0.081
0.000	0.050	0.082	0.197	0.081
0.000	0.051	0.082	0.198	0.081
0.000	0.050	0.082	0.198	0.081
0.000	0.051	0.082	0.198	0.080

Standard Deviations	0.0004	0.0004	0.0006	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

SP

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Danielle M Bell

DANIELLE M BELL

Signature and Printed Name

01/08/2018
Date

*1/12/18
SP*

Stability Checks # 8D-000814 Hernando County S.O. '18/18

DES

HERNANDO COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 01/08/2018
 Software: 8100.27

HERNANDO COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 01/08/2018
 Software: 8100.27

HERNANDO COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 01/08/2018
 Software: 8100.27

HERNANDO COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 01/08/2018
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:39
Control Test	0.049	11:39
Air Blank	0.000	11:40
Control Test	0.049	11:41
Air Blank	0.000	11:41
Control Test	0.049	11:42
Air Blank	0.000	11:42
Control Test Stats		11:42
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Test	g/210L	Time
Air Blank	0.000	11:44
Control Test	0.080	11:44
Air Blank	0.000	11:45
Control Test	0.079	11:46
Air Blank	0.000	11:46
Control Test	0.081	11:47
Air Blank	0.000	11:47
Control Test Stats		11:47
Average	0.0800	
Std Dev	0.0010	
Rel Std Dev(%)	1.2500	

Test	g/210L	Time
Air Blank	0.000	11:48
Control Test	0.195	11:49
Air Blank	0.000	11:50
Control Test	0.196	11:50
Air Blank	0.000	11:51
Control Test	0.196	11:51
Air Blank	0.000	11:52
Control Test Stats		11:52
Average	0.1957	
Std Dev	0.0006	
Rel Std Dev(%)	0.2951	

Test	g/210L	Time
Air Blank	0.000	11:53
Control Test	0.083	11:53
Air Blank	0.000	11:54
Control Test	0.083	11:54
Air Blank	0.000	11:55
Control Test	0.083	11:55
Air Blank	0.000	11:55
Control Test Stats		11:55
Average	0.0830	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

11/2/18

Operator's Signature

Operator's Signature

Operator's Signature

Operator's Signature

SIP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 11:14

Date of Inspection: 01/08/2018

Serial Number: 80-000814
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:
BYPASSED AI TO OPERATE INSTRUMENT. NOT A COMPLIANCE CHECK.

SP

N/A Compliance
Not Determined

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Danielle M Bell

DANIELLE M BELL

Signature and Printed Name

01/08/2018
Date

1/11/18
SP



Florida Department of Law Enforcement
 Alcohol Testing Program
 2729 Fort Knox Blvd.
 Bldg. 2, Suite 1300
 Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000814, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000814</u>	UNCERTAINTY* ±	
Owning Agency:	<u>HERNANDO COUNTY S.O.</u>	0.05 g/ 210 L	0.004
Calibration Date:	<u>01/08/2018</u>	0.08 g/ 210 L	0.005
Calibration Time:	<u>13:50</u>	0.20 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by Calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the Target Alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS) ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/08/2018

Date


DANIELLE M BELL,
 Department Inspector

FDLE/ATP Form 69 January 2018
 Issuing Authority: Alcohol Testing Program

11/21/18
JP