



INSTRUMENT PROCESSING SHEET

Agency Miami-Dade Police Department

S/N 80-000883

Florida Department of Law Enforcement

Date In 2/12/2018

DI Completion Date 2/13/2018

Ship P/U H/D CMI EE

Intake Performed By <u>DELL</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>DELL</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>101</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 101</u> 32 mm <u>.167</u> (.139 - .169) 36 mm <u>.187</u> (.156 - .190) 53 mm <u>.250</u> (.228 - .278) 103 mm <u>.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28199</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG626605 09/22/2018</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD3967	201707D 07/25/2019	0.080	SD3968	201707E 07/25/2019	0.200	SD3969	201707C 07/24/2019	0.080 DGS	N/A	AG626605 09/22/2018	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.050	SD3967	201707D 07/25/2019															
0.080	SD3968	201707E 07/25/2019															
0.200	SD3969	201707C 07/24/2019															
0.080 DGS	N/A	AG626605 09/22/2018															
Final Release Date FDLE FEB 21 2018 Alcohol Testing Program		Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>DELL</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.50c</u> External Digital Therm. ID#: <u>300918</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD3967</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD3968</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD3969</u>															

Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																																													
0.000		N/A	N/A																																													
0.040																																																
0.100																																																
0.200																																																
0.300																																																
0.080 DGS	N/A																																															
Simulator	Serial Number	Lot Number	Expiration																																													
0.050																																																
0.080																																																
0.200																																																
0.080 DGS	N/A																																															

Department Inspection Performed By <u>DELL</u> Barometric Pressure ID# <u>68639</u> Gauge <u>1026</u> Instrument <u>1025</u> Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2017-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD3965</td> </tr> <tr> <td>Interferent</td> <td>SD3966</td> </tr> <tr> <td>0.050</td> <td>SD3967</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	SD3965	Interferent	SD3966	0.050	SD3967	0.080	SD3968	0.200	SD3969
Simulator	Serial Number											
0.000	SD3965											
Interferent	SD3966											
0.050	SD3967											
0.080	SD3968											
0.200	SD3969											

Notes/Suggested Service: E-mailed **APPROVED**

After the inspection, while dialing the server to upload the data the instrument froze and had to be restarted. Once restarted instrument uploaded the data on the first try to the COBRA server.

Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____
<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use
<u>_____</u> 2/21/18 <u>_____</u> 2/21/18 Tech Review / Date Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI-DADE PD
Time of Inspection: 09:43

Date of Inspection: 02/13/2018

Serial Number: 80-000883
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626605 Exp: 09/22/2018
0.000	0.051	0.081	0.200	0.080
0.000	0.051	0.082	0.201	0.081
0.000	0.051	0.081	0.201	0.080
0.000	0.051	0.082	0.201	0.081
0.000	0.052	0.081	0.200	0.081
0.000	0.051	0.081	0.201	0.081
0.000	0.051	0.081	0.200	0.080
0.000	0.051	0.082	0.200	0.081
0.000	0.051	0.082	0.200	0.080
0.000	0.052	0.082	0.200	0.081
Standard Deviations	0.0004	0.0005	0.0005	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Reyes-Rivera Signature and Printed Name

DAVID E REYES-RIVERA

02/13/2018
Date

*2/13/18
JLZ*

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-000883	Miami-Dade Police Department	02/13/2018	DELL

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
SN: SD3967 Temp: 34.05c	SN: SD3968 Temp: 34.00c	SN: SD3969 Temp: 34.06c	Lot AG626605
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

Test	g/210L	Time	Test	g/210L	Time	Test	g/210L	Time	Test	g/210L	Time
Air Blank	0.000	07:13	Air Blank	0.000	07:18	Air Blank	0.000	07:24	Air Blank	0.000	07:29
Control Test	0.050	07:14	Control Test	0.081	07:19	Control Test	0.197	07:24	Control Test	0.081	07:29
Air Blank	0.000	07:15	Air Blank	0.000	07:20	Air Blank	0.000	07:25	Air Blank	0.000	07:30
Control Test	0.050	07:15	Control Test	0.080	07:20	Control Test	0.199	07:26	Control Test	0.080	07:30
Air Blank	0.000	07:16	Air Blank	0.000	07:21	Air Blank	0.000	07:26	Air Blank	0.000	07:31
Control Test	0.050	07:17	Control Test	0.081	07:22	Control Test	0.198	07:27	Control Test	0.081	07:31
Air Blank	0.000	07:17	Air Blank	0.000	07:22	Air Blank	0.000	07:27	Air Blank	0.000	07:31
Control Test Stats			Control Test Stats			Control Test Stats			Control Test Stats		
Average	0.0500		Average	0.0807		Average	0.1980		Average	0.0807	
Std Dev	0.0000		Std Dev	0.0006		Std Dev	0.0010		Std Dev	0.0006	
Rel. Std Dev(%)	0.0000		Rel. Std Dev(%)	0.7157		Rel. Std Dev(%)	0.5051		Rel. Std Dev(%)	0.7157	

DELL
Operator's Signature

DELL
Operator's Signature

DELL
Operator's Signature

DELL
Operator's Signature

DELL

2/13/18
DELL



Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000883, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000883</u>	UNCERTAINTY* ±
Owning Agency:	<u>MIAMI-DADE PD</u>	0.05 g/ 210 L
Calibration Date:	<u>02/13/2018</u>	0.08 g/ 210 L
Calibration Time:	<u>09:43</u>	0.20 g/ 210 L
		0.080 g/ 210 L Dry Gas Control
		0.005

All results are reported in g/210 L.

Bias is limited by Calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the Target Alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS) ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/13/2018 Date
David E Reyes-Rivera DAVID E REYES-RIVERA,
 Department Inspector

FDLE/ATP Form 69 January 2018
 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

2/21/18
[Signature]