



INSTRUMENT PROCESSING SHEET

Agency North Miami Beach Police Department S/N 80-000886

Florida Department of Law Enforcement

Date In 10/29/2018 DI Completion Date 10/29/2018 Ship P/U H/D CMI EE

Intake Performed By <u>DEE</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>DEE</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>165</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 101</u> 32 mm <u>.148</u> (.139 - .169) 36 mm <u>.160</u> (.156 - .190) 53 mm <u>.226</u> (.228 - .278) 103 mm <u>.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28199</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG805701 02/26/2020</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD3967	201707D 07/25/2019	0.080	SD3968	201707E 07/25/2019	0.200	SD3969	201707C 07/24/2019	0.080 DGS	N/A	AG805701 02/26/2020	Flow Calibration Performed By <u>DEE</u> Flow Column # <u>ATP104</u> <input checked="" type="checkbox"/> 5L/min - 17mm <input checked="" type="checkbox"/> 15L/min - 53mm <input checked="" type="checkbox"/> 30L/min - 103mm <input checked="" type="checkbox"/> R-Value <u>165</u> <input checked="" type="checkbox"/> Post Calibration Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>148</u> (.139 - .169) 36 mm <u>167</u> (.156 - .190) 53 mm <u>234</u> (.228 - .278) 103 mm <u>500</u> (.447 - .547)																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	SD3967	201707D 07/25/2019																																																											
0.080	SD3968	201707E 07/25/2019																																																											
0.200	SD3969	201707C 07/24/2019																																																											
0.080 DGS	N/A	AG805701 02/26/2020																																																											
Final Release Date <div style="text-align: center; font-weight: bold; font-size: 1.2em;">FDLE</div> <div style="text-align: center; font-weight: bold; font-size: 1.2em;">NOV 05 2018</div> <div style="text-align: center;">Alcohol Testing Program</div>	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____	Temperature Checks Performed By <u>DEE</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.56C</u> External Digital Therm. ID#: <u>300949</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD3967</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD3968</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD3969</u>																																																											
Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>DEE</u> Barometric Pressure ID# <u>68639</u> Gauge <u>1018</u> Instrument <u>1016</u> Mouth Alcohol Solution Lot # <u>2017-B</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD3965</td> </tr> <tr> <td>Interferent</td> <td>SD3966</td> </tr> <tr> <td>0.050</td> <td>SD3967</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	SD3965	Interferent	SD3966	0.050	SD3967	0.080	SD3968	0.200	SD3969
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	SD3965																																																												
Interferent	SD3966																																																												
0.050	SD3967																																																												
0.080	SD3968																																																												
0.200	SD3969																																																												
Notes/Suggested Service: <u>E-mailed</u> <input checked="" type="checkbox"/> APPROVED	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use	<table style="width:100%; border-collapse: collapse;"> <tr> <td style="width:50%; text-align: center;"> <u>P. Dean 11/5/18</u> Tech Review / Date </td> <td style="width:50%; text-align: center;"> <u>J. P. ... 11/5/18</u> Admin Review / Date </td> </tr> </table>	<u>P. Dean 11/5/18</u> Tech Review / Date	<u>J. P. ... 11/5/18</u> Admin Review / Date																																																									
<u>P. Dean 11/5/18</u> Tech Review / Date	<u>J. P. ... 11/5/18</u> Admin Review / Date																																																												

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: NORTH MIAMI BEACH PD
Time of Inspection: 14:15

Date of Inspection: 10/29/2018

Serial Number: 80-000886
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.049	0.081	0.198	0.079
0.000	0.050	0.083	0.201	0.080
0.000	0.050	0.082	0.202	0.080
0.000	0.051	0.082	0.201	0.080
0.000	0.051	0.082	0.201	0.080
0.000	0.050	0.082	0.201	0.080
0.000	0.051	0.082	0.200	0.080
0.000	0.051	0.082	0.200	0.079
0.000	0.051	0.082	0.200	0.080
0.000	0.050	0.082	0.200	0.080

Standard Deviations	0.0006	0.0004	0.0010	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

PDM

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes Rivera

DAVID E REYES-RIVERA

Signature and Printed Name

10/29/2018
Date

*11/5/18
JO*

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-000886	North Miami Beach Police Department	10/29/2018	<i>DELL</i>

0.05g/210L	0.08g/210L	0.20g/210L	0.08g/210L	0.20g/210L																																																																																																																																						
SN: SD3967 Temp: 34.08c 0.047 to 0.053 <input checked="" type="checkbox"/>	SN: SD3968 Temp: 34.08c 0.077 to 0.083 <input checked="" type="checkbox"/>	SN: SD3969 Temp: 34.09c 0.194 to 0.206 <input checked="" type="checkbox"/>	SN: SD3968 Temp: 34.08c 0.077 to 0.083 <input checked="" type="checkbox"/>	SN: SD3969 Temp: 34.09c 0.194 to 0.206 <input checked="" type="checkbox"/>																																																																																																																																						
NORTH MIAMI BEACH PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000886 10/29/2018 Software: 8100.27	NORTH MIAMI BEACH PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000886 10/29/2018 Software: 8100.27	NORTH MIAMI BEACH PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000886 10/29/2018 Software: 8100.27	NORTH MIAMI BEACH PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000886 10/29/2018 Software: 8100.27	NORTH MIAMI BEACH PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000886 10/29/2018 Software: 8100.27																																																																																																																																						
<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:29</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:30</td></tr> <tr><td>Control Test</td><td>0.050</td><td>12:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:32</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:33</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:33</td></tr> <tr><td>Average</td><td>0.0493</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:29	Control Test	0.049	12:30	Air Blank	0.000	12:30	Control Test	0.050	12:31	Air Blank	0.000	12:32	Control Test	0.049	12:32	Air Blank	0.000	12:33	Control Test	0.049	12:33	Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:34</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:35</td></tr> <tr><td>Control Test</td><td>0.081</td><td>12:36</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:36</td></tr> <tr><td>Control Test</td><td>0.082</td><td>12:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:38</td></tr> <tr><td>Control Test</td><td>0.0810</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2346</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:34	Control Test	0.080	12:35	Air Blank	0.000	12:35	Control Test	0.081	12:36	Air Blank	0.000	12:36	Control Test	0.082	12:37	Air Blank	0.000	12:38	Control Test	0.0810		Std Dev	0.0010		Rel Std Dev(%)	1.2346		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:39</td></tr> <tr><td>Control Test</td><td>0.197</td><td>12:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:40</td></tr> <tr><td>Control Test</td><td>0.199</td><td>12:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:42</td></tr> <tr><td>Control Test</td><td>0.200</td><td>12:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:43</td></tr> <tr><td>Control Test</td><td>0.1987</td><td></td></tr> <tr><td>Std Dev</td><td>0.0015</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7689</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:39	Control Test	0.197	12:40	Air Blank	0.000	12:40	Control Test	0.199	12:41	Air Blank	0.000	12:42	Control Test	0.200	12:42	Air Blank	0.000	12:43	Control Test	0.1987		Std Dev	0.0015		Rel Std Dev(%)	0.7689		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:44</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:45</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:46</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:46</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:47</td></tr> <tr><td>Control Test</td><td>0.0797</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:44	Control Test	0.079	12:44	Air Blank	0.000	12:45	Control Test	0.080	12:45	Air Blank	0.000	12:46	Control Test	0.080	12:46	Air Blank	0.000	12:47	Control Test	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247	
Test	g/210L	Time																																																																																																																																								
Air Blank	0.000	12:29																																																																																																																																								
Control Test	0.049	12:30																																																																																																																																								
Air Blank	0.000	12:30																																																																																																																																								
Control Test	0.050	12:31																																																																																																																																								
Air Blank	0.000	12:32																																																																																																																																								
Control Test	0.049	12:32																																																																																																																																								
Air Blank	0.000	12:33																																																																																																																																								
Control Test	0.049	12:33																																																																																																																																								
Average	0.0493																																																																																																																																									
Std Dev	0.0006																																																																																																																																									
Rel Std Dev(%)	1.1703																																																																																																																																									
Test	g/210L	Time																																																																																																																																								
Air Blank	0.000	12:34																																																																																																																																								
Control Test	0.080	12:35																																																																																																																																								
Air Blank	0.000	12:35																																																																																																																																								
Control Test	0.081	12:36																																																																																																																																								
Air Blank	0.000	12:36																																																																																																																																								
Control Test	0.082	12:37																																																																																																																																								
Air Blank	0.000	12:38																																																																																																																																								
Control Test	0.0810																																																																																																																																									
Std Dev	0.0010																																																																																																																																									
Rel Std Dev(%)	1.2346																																																																																																																																									
Test	g/210L	Time																																																																																																																																								
Air Blank	0.000	12:39																																																																																																																																								
Control Test	0.197	12:40																																																																																																																																								
Air Blank	0.000	12:40																																																																																																																																								
Control Test	0.199	12:41																																																																																																																																								
Air Blank	0.000	12:42																																																																																																																																								
Control Test	0.200	12:42																																																																																																																																								
Air Blank	0.000	12:43																																																																																																																																								
Control Test	0.1987																																																																																																																																									
Std Dev	0.0015																																																																																																																																									
Rel Std Dev(%)	0.7689																																																																																																																																									
Test	g/210L	Time																																																																																																																																								
Air Blank	0.000	12:44																																																																																																																																								
Control Test	0.079	12:44																																																																																																																																								
Air Blank	0.000	12:45																																																																																																																																								
Control Test	0.080	12:45																																																																																																																																								
Air Blank	0.000	12:46																																																																																																																																								
Control Test	0.080	12:46																																																																																																																																								
Air Blank	0.000	12:47																																																																																																																																								
Control Test	0.0797																																																																																																																																									
Std Dev	0.0006																																																																																																																																									
Rel Std Dev(%)	0.7247																																																																																																																																									
<i>DELL</i> Operator's Signature	<i>DELL</i> Operator's Signature	<i>DELL</i> Operator's Signature	<i>DELL</i> Operator's Signature	<i>DELL</i> Operator's Signature																																																																																																																																						

ADAM
 11/5/18
DELL



Florida Department of Law Enforcement
 Alcohol Testing Program
 4700 Terminal Drive, Suite 1
 Ft. Myers, FL 33907

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000886, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000886</u>	UNCERTAINTY* ±
Owning Agency:	<u>NORTH MIAMI BEACH PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>10/29/2018</u>	0.080 g/ 210 L 0.005
Calibration Time:	<u>14:15</u>	0.200 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.
 Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
 *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.
 Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

10/29/2018 Date

DAVID E REYES-RIVERA,
 Department Inspector

FDLE/ATP Form 69 July 2018
 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

11/5/18
ADAM

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Flow calibration	80-000886	North Miami Beach Police Department	10/29/2018	<i>[Signature]</i>

NORTH MIAMI BEACH PD
 Intoxilyzer - Alconoh Analyzer
 Model 8000 SN 80-000886
 10/29/2018
 Software: 8100.27

Flow Rate Calibration*****
 1: Rate (Liters/min) = 5
 SORT(Diff)) = 6.655
 2: Rate (Liters/min) = 15
 SORT(Diff)) = 11.223
 3: Rate (Liters/min) = 30
 SORT(Diff)) = 21.271

Dependent Data Scale Factor = 10000 L/min
 Independent Data Scale Factor = 256
 Rounded Slope = 671
 Rounded Intercept = -576234
 Correlation = 0.99459

Adam

11/5/18
[Signature]