



INSTRUMENT PROCESSING SHEET

Agency St. Johns County SO

S/N 80-001134

Florida Department of Law Enforcement

Date In 03/20/2018 DI Completion Date 4/25/18

Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>JB</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>See note shipped with instrument.</u>	<b>Quality Checks</b> Performed By <u>JB</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>179</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>A7P102</u> 32 mm <u>216/265</u> (.139 - .169) 36 mm <u>250/300</u> (.156 - .190) 53 mm <u>312/421</u> (.228 - .278) 103 mm <u>544/921</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks	<b>Flow Calibration</b> Performed By <u>JB</u> Flow Column # <u>A7P103</u> <input checked="" type="checkbox"/> 5L/min - 17mm <input checked="" type="checkbox"/> 15L/min - 53mm <input checked="" type="checkbox"/> 30L/min - 103mm <input checked="" type="checkbox"/> R-Value <u>180</u> <input checked="" type="checkbox"/> Post Calibration Verification (L/s) Flow Column # <u>A7P102</u> 32 mm <u>148</u> (.139 - .169) 36 mm <u>171</u> (.156 - .190) 53 mm <u>238</u> (.228 - .278) 103 mm <u>494</u> (.447 - .547)															
<b>Final Release Date</b> <p style="text-align: center;"><b>FDLE</b>  APR 25 2018  Alcohol Testing Program</p>	<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>G2835</u></td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td><u>G2840</u></td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG805702</u> 2/26/20</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>G2835</u>	201707D 07/25/2019	0.080	<u>G2840</u>	201707E 07/25/2019	0.200	<u>SD1025</u>	201707C 07/24/2019	0.080 DGS	N/A	<u>AG805702</u> 2/26/20	<b>Maintenance</b> Performed By <u>JB</u> <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input checked="" type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Temperature Checks</b> Performed By <u>JB</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.8</u> External Digital Therm. ID#: <u>200502</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>G2835</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>G2840</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1025</u>
Simulator	Serial #	Lot #/Exp															
0.050	<u>G2835</u>	201707D 07/25/2019															
0.080	<u>G2840</u>	201707E 07/25/2019															
0.200	<u>SD1025</u>	201707C 07/24/2019															
0.080 DGS	N/A	<u>AG805702</u> 2/26/20															

<b>Calibration Adjustment</b> Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<b>Department Inspection</b> Performed By <u>JB</u> Barometric Pressure ID# <u>28427</u> Gauge <u>1011</u> Instrument <u>1011</u> Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G2880</u></td> </tr> <tr> <td>Interferent</td> <td><u>G2407</u></td> </tr> <tr> <td>0.050</td> <td><u>G2835</u></td> </tr> <tr> <td>0.080</td> <td><u>G2840</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	<u>G2880</u>	Interferent	<u>G2407</u>	0.050	<u>G2835</u>	0.080	<u>G2840</u>	0.200	<u>SD1025</u>
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	<u>G2880</u>																																																												
Interferent	<u>G2407</u>																																																												
0.050	<u>G2835</u>																																																												
0.080	<u>G2840</u>																																																												
0.200	<u>SD1025</u>																																																												
<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																																												

Notes/Suggested Service:  
Breath tube replaced  
Did a flow cal upon replacement of new breath tube

Tech Review / Date 4/25/18 Admin Review / Date 4/25/18

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ST JOHNS COUNTY SO  
Time of Inspection: 08:26

Date of Inspection: 04/25/2018

Serial Number: 80-001134  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805702 Exp: 02/26/2020
0.000	0.048	0.080	0.199	0.078
0.000	0.048	0.080	0.199	0.078
0.000	0.048	0.080	0.199	0.078
0.000	0.048	0.080	0.199	0.078
0.000	0.048	0.080	0.199	0.077
0.000	0.049	0.080	0.199	0.077
0.000	0.048	0.080	0.199	0.077
0.000	0.048	0.080	0.199	0.077
0.000	0.049	0.080	0.198	0.077
0.000	0.048	0.080	0.199	0.077

Standard Deviations	0.0004	0.0000	0.0003	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

RDS

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAKE L SHANAHAN

Signature and Printed Name

04/25/2018  
Date

4/25/18  
JD

80-001134  
 Flow Calibration  
 Stability Checks  
 4/25/18

INTOXILYZER 8000  
 Instrument Initialization  
 06:17 04/25/2018

ST JOHNS COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001134  
 04/25/2018  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:50
Control Test	0.079	06:50
Air Blank	0.000	06:50
Control Test	0.079	06:51
Air Blank	0.000	06:51
Control Test	0.079	06:52
Air Blank	0.000	06:52
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

WOS

ST JOHNS COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001134  
 04/25/2018  
 Software: 8100.27

Flow Rate Calibration\*\*\*\*\*  
 1: Rate (Liters/min) = 5  
 SQRTOff) = 6.533  
 2: Rate (Liters/min) = 15  
 SQRTOff) = 11.355  
 3: Rate (Liters/min) = 30  
 SQRTOff) = 20.734  
 Dependent Data Scale Factor = 100000 L/min  
 Independent Data Scale Factor = 256  
 Rounded Slope = 683  
 Rounded Intercept = -990011  
 Correlation = 0.99731

ST JOHNS COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001134  
 04/25/2018  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:53
Control Test	0.048	06:54
Air Blank	0.000	06:54
Control Test	0.048	06:55
Air Blank	0.000	06:55
Control Test	0.048	06:56
Air Blank	0.000	06:57
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

ST JOHNS COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001134  
 04/25/2018  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:57
Control Test	0.079	06:58
Air Blank	0.000	06:59
Control Test	0.079	06:59
Air Blank	0.000	07:00
Control Test	0.079	07:01
Air Blank	0.000	07:01
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

Operator's Signature

Operator's Signature

ST JOHNS COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001134  
 04/25/2018  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:02
Control Test	0.197	07:03
Air Blank	0.000	07:03
Control Test	0.196	07:04
Air Blank	0.000	07:04
Control Test	0.197	07:05
Air Blank	0.000	07:06
Control Test Stats		
Average	0.1967	
Std Dev	0.0006	
Rel Std Dev(%)	0.2936	

Operator's Signature

4/25/18  
 SD

SD



Florida Department of Law Enforcement  
 Alcohol Testing Program  
 2729 Fort Knox Blvd.  
 Bldg. 2, Suite 1300  
 Tallahassee, FL 32308

## Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001134, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001134</u>	UNCERTAINTY* ±	
Owning Agency:	<u>ST JOHNS COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>04/25/2018</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>08:26</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
 \*Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 March 2018  
 Issuing Authority: Alcohol Testing Program

04/25/2018

Date

  
**JAKE L SHANAHAN,**  
 Department Inspector

*Service • Integrity • Respect • Quality*

*4/25/18*  
*JD*

Please check the LCD screen it is starting to get dim.

Thank you.

80-001134

Rec'd

3/20/18

JR

4/25/18  
JR