



INSTRUMENT PROCESSING SHEET

Agency Daytona Beach P.D.

S/N 80-001135

Florida Department of Law Enforcement

Date In 04/23/2018

DI Completion Date 4/26/18

Ship P/U H/D CMI EE

Intake Performed By <u>JLS</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____	Quality Checks Performed By <u>PGM</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>136</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>.164</u> (.139 - .169) 36 mm <u>.187</u> (.156 - .190) 53 mm <u>.250</u> (.228 - .278) 103 mm <u>.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
--	--	--

Final Release Date

FDLE

APR 30 2018

Alcohol Testing Program

Simulator	Serial #	Lot #/Exp
0.050	G2835	201707D 7/25/19
0.080	G2840	201707E 7/25/19
0.200	S01025	201707C 7/24/19
0.080 DGS	N/A	AG805702 2/26/20

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Temperature Checks Performed By PGM

Lab Temp °C 21.8
 External Digital Therm. ID#: S00502
 34°C +/- .2 Serial #: G2835
 34°C +/- .2 Serial #: G2840
 34°C +/- .2 Serial #: S01025

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By PGM

Barometric Pressure ID# 28427
 Gauge 1010 Instrument 1011
 Mouth Alcohol Solution Lot # 2017-B
 Acetone Stock Solution Lot # 2018-A

Simulator	Serial Number
0.000	G2880
Interferent	G2407
0.050	G2835
0.080	G2840
0.200	S01025

Attachments

Form 41 Post-Stability Checks
 Stability Checks Flow Calibration
 Calibration Certificate Form 40
 Calibration Adjustment Other _____

Notes/Suggested Service: _____

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

PGM 4/30/18 JJ 4/30/18
 Tech Review / Date Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: DAYTONA BEACH P.D.
Time of Inspection: 12:15

Date of Inspection: 04/26/2018

Serial Number: 80-001135
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805702 Exp: 02/26/2020
0.000	0.048	0.078	0.197	0.080
0.000	0.048	0.079	0.198	0.079
0.000	0.048	0.080	0.197	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.048	0.080	0.198	0.080
0.000	0.048	0.079	0.197	0.079
0.000	0.048	0.080	0.198	0.080
0.000	0.049	0.079	0.198	0.080
0.000	0.048	0.080	0.198	0.079
0.000	0.048	0.080	0.198	0.079
Standard Deviations	0.0004	0.0006	0.0004	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

✓

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick J Murphy

PATRICK J MURPHY

Signature and Printed Name

04/26/2018
Date

4/30/18
22

DAYTONA BEACH P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001135
04/26/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:58
Control Test	0.048	09:59
Air Blank	0.000	10:00
Control Test	0.048	10:00
Air Blank	0.000	10:01
Control Test	0.048	10:01
Air Blank	0.000	10:02
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy
Operator's Signature

DAYTONA BEACH P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001135
04/26/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:09
Control Test	0.079	10:10
Air Blank	0.000	10:11
Control Test	0.079	10:11
Air Blank	0.000	10:12
Control Test	0.080	10:13
Air Blank	0.000	10:13
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

P Murphy
Operator's Signature

DAYTONA BEACH P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001135
04/26/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:04
Control Test	0.197	10:04
Air Blank	0.000	10:05
Control Test	0.197	10:06
Air Blank	0.000	10:06
Control Test	0.197	10:07
Air Blank	0.000	10:07
Control Test Stats		
Average	0.1970	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy
Operator's Signature

DAYTONA BEACH P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001135
04/26/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:18
Control Test	0.079	10:18
Air Blank	0.000	10:18
Control Test	0.080	10:19
Air Blank	0.000	10:19
Control Test	0.081	10:20
Air Blank	0.000	10:20
Control Test Stats		
Average	0.0800	
Std Dev	0.0010	
Rel Std Dev(%)	1.2500	

DGS

P Murphy
Operator's Signature

4/30/18
[Signature]



Florida Department of Law Enforcement
 Alcohol Testing Program
 2729 Fort Knox Blvd.
 Bldg. 2, Suite 1300
 Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001135, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001135</u>	UNCERTAINTY* ±
Owning Agency:	<u>DAYTONA BEACH P.D.</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>04/26/2018</u>	0.080 g/ 210 L 0.005
Calibration Time:	<u>12:15</u>	0.200 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.
 Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
 *Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.
 Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 March 2018
 Issuing Authority: Alcohol Testing Program

04/26/2018

Date

Patrick J Murphy

PATRICK J MURPHY,
 Department Inspector

Service • Integrity • Respect • Quality

4/30/18

Return Material Authorization form

SHIP TO: CMI, Inc.

Enforcement Electronics

Shipment to repair facility authorized by: BRETT KIRKLAND on 10/13/17

Items Returned: Instrument Supplies other Please Describe: _____

Instrument Model: 8000 Serial Number 80-001135

Bill To Address:

FDLE ATP
2331 Phillips Rd
Tallahassee FL 32308

Ship To Address:

FDLE ATP
2331 Phillips Rd
Tallahassee FL 32308

Reason for Return:

CONNECTOR ATTACHING TUBE TO EXHAUST PORT (SIMULATOR) BROKEN.
PLEASE REPLACE CONNECTOR AND RETURN. NO FURTHER WORK NEEDED OR
DESIRED

Please choose one of the following options:

- 1. I _____, authorize all repairs.
- 2. I _____, authorize repairs up to \$ _____.
- 3. I require an estimate **BEFORE** any repairs will be authorized and/or conducted,
Please contact, Name: BRETT KIRKLAND Phone Number: 850-617-1275
BRETTKIRKLAND@fdle.state.fl.us

Notes:



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Daytona Beach PD S/N 20-001135

Date In 10/2/17 Date Out _____ Ship P/U H/D CMI EE

Intake <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>ok</u> Case <u>ok</u> Handle <u>ok</u> Dry Gas Holder <u>ok</u> Feet <u>ok</u> Keyboard/Plug <u>ok</u> Back/Plugs <u>ok</u> Screws tight <u>ok</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u> Notes: _____ _____ _____	Quality Checks Performed By <u>DS</u> <input type="checkbox"/> Lab Temp °C _____ <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input type="checkbox"/> Instrument Set Up Verified <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Flow Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) <input type="checkbox"/> Barometric Pressure Check Gauge ID # _____ <input type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05			0.08			0.20			0.08 DGS	N/A		Flow Calibration Performed By _____ <input type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) Maintenance Performed By <u>DSB</u> <input checked="" type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Quality Checks Cont. Performed By _____ Simulator Temperatures °C External Digital Therm. ID#: _____ <input type="checkbox"/> 34°C +/- .2 Serial #: _____ <input type="checkbox"/> 34°C +/- .2 Serial #: _____ <input type="checkbox"/> 34°C +/- .2 Serial #: _____
Simulator	Serial #	Lot #/Exp															
0.05																	
0.08																	
0.20																	
0.08 DGS	N/A																

Calibration Adjustment Performed By _____ <input type="checkbox"/> Calibration Adjustment N/A <input type="checkbox"/> Calibration Adjustment Complete Barometric Pressure Gauge ID # _____																							
Simulator	Serial Number	Lot Number	Expiration																				
0.000		N/A	N/A																				
0.040																							
0.100																							
0.200																							
0.300																							
0.080 DGS	N/A																						
<input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>				Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																				
0.05																							
0.08																							
0.20																							
0.08 DGS	N/A																						

Department Inspection Performed By _____ <input type="checkbox"/> Barometric Pressure _____ Gauge ID# _____ Instrument Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____	
Simulator	Serial Number
0.00	
Interferent	
0.05	
0.08	
0.20	

Attachments <input type="checkbox"/> Form 41 <input type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____	
---	--

Notes/Suggested Service:
Connector broke rendering inspection impossible, sending to CMI for repair

<input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use
--

Quality Control Review _____ Date 10/17/17

