



INSTRUMENT PROCESSING SHEET

Agency Lady Lake Police DepartmentS/N 80-001177Florida Department of
Law EnforcementDate In 06/27/2018DI Completion Date 07/03/2018 Ship P/U H/D CMI EE

Intake Performed By <u>SQC</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>DMB</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>217</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 105</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.250</u> (.228 - .278) 103 mm <u>0.531</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28662</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																											
Final Release Date FDLE JUL 03 2018 Alcohol Testing Program	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1018</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>SD3962</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>G2078</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG805701 02/26/2020</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD1018	201707D 07/25/2019	0.080	SD3962	201707E 07/25/2019	0.200	G2078	201707C 07/24/2019	0.080 DGS	N/A	AG805701 02/26/2020	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>DMB</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.1</u> External Digital Therm. ID#: <u>300505</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1018</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3962</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>G2078</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	SD1018	201707D 07/25/2019																																																											
0.080	SD3962	201707E 07/25/2019																																																											
0.200	G2078	201707C 07/24/2019																																																											
0.080 DGS	N/A	AG805701 02/26/2020																																																											
Calibration Adjustment Performed By <u>DMB</u> Barometric Pressure Gauge <u>28427</u> ID # <u>1017</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G8144</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>G2403</td> <td>17410</td> <td>12/06/2019</td> </tr> <tr> <td>0.100</td> <td>G2879</td> <td>18070</td> <td>02/26/2020</td> </tr> <tr> <td>0.200</td> <td>G3709</td> <td>17340</td> <td>10/09/2019</td> </tr> <tr> <td>0.300</td> <td>G8149</td> <td>18110</td> <td>04/02/2020</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>17817080A2</td> <td>08/05/2019</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1018</td> <td>201707D</td> <td>07/25/2019</td> </tr> <tr> <td>0.080</td> <td>SD3962</td> <td>201707E</td> <td>07/25/2019</td> </tr> <tr> <td>0.200</td> <td>G2078</td> <td>201707C</td> <td>07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG805701</td> <td>02/26/2020</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	G8144	N/A	N/A	0.040	G2403	17410	12/06/2019	0.100	G2879	18070	02/26/2020	0.200	G3709	17340	10/09/2019	0.300	G8149	18110	04/02/2020	0.080 DGS	N/A	17817080A2	08/05/2019	Simulator	Serial Number	Lot Number	Expiration	0.050	SD1018	201707D	07/25/2019	0.080	SD3962	201707E	07/25/2019	0.200	G2078	201707C	07/24/2019	0.080 DGS	N/A	AG805701	02/26/2020	Department Inspection Performed By <u>DMB</u> Barometric Pressure ID# <u>28662</u> Gauge <u>1017</u> Instrument <u>1017</u> Mouth Alcohol Solution Lot # <u>2017-B</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G4444</td> </tr> <tr> <td>Interferent</td> <td>G6621</td> </tr> <tr> <td>0.050</td> <td>SD1018</td> </tr> <tr> <td>0.080</td> <td>SD3962</td> </tr> <tr> <td>0.200</td> <td>G2078</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	G4444	Interferent	G6621	0.050	SD1018	0.080	SD3962	0.200	G2078
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	G8144	N/A	N/A																																																										
0.040	G2403	17410	12/06/2019																																																										
0.100	G2879	18070	02/26/2020																																																										
0.200	G3709	17340	10/09/2019																																																										
0.300	G8149	18110	04/02/2020																																																										
0.080 DGS	N/A	17817080A2	08/05/2019																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050	SD1018	201707D	07/25/2019																																																										
0.080	SD3962	201707E	07/25/2019																																																										
0.200	G2078	201707C	07/24/2019																																																										
0.080 DGS	N/A	AG805701	02/26/2020																																																										
Simulator	Serial Number																																																												
0.000	G4444																																																												
Interferent	G6621																																																												
0.050	SD1018																																																												
0.080	SD3962																																																												
0.200	G2078																																																												
Notes/Suggested Service: <u>Performed optical bench calibration adjustment to bring values closer to nominal. DMB 7/3/18 Please change menu level 2 password to something unique. DMB 7/3/18</u> _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <u>S Platt 7/3/18</u> <u>J Jahan 7/3/18</u> Tech Review / Date Admin Review / Date																																																												

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LADY LAKE P.D.
Time of Inspection: 12:29

Date of Inspection: 07/03/2018

Serial Number: 80-001177
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.049	0.081	0.199	0.080
0.000	0.050	0.082	0.204	0.080
0.000	0.050	0.082	0.205	0.079
0.000	0.051	0.082	0.204	0.079
0.000	0.050	0.082	0.204	0.079
0.000	0.051	0.082	0.203	0.079
0.000	0.050	0.082	0.202	0.080
0.000	0.051	0.082	0.202	0.080
0.000	0.050	0.082	0.202	0.080
0.000	0.050	0.082	0.203	0.079

Standard Deviations	0.0006	0.0003	0.0016	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0007 Number of Simulators Used: 5

Remarks:

SP

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Danielle M Bell

DANIELLE M BELL

Signature and Printed Name

07/03/2018
Date

7/3/18
JD

Stability Checks # 80-001177 Lady Lake P.D. 7/31/18 ~~08/18~~

DES

LADY LAKE P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001177
07/03/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:18
Control Test	0.047	08:18
Air Blank	0.000	08:19
Control Test	0.047	08:20
Air Blank	0.000	08:20
Control Test	0.047	08:21
Air Blank	0.000	08:21
Control Test Stats		
Average	0.0470	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

LADY LAKE P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001177
07/03/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:23
Control Test	0.078	08:23
Air Blank	0.000	08:24
Control Test	0.078	08:25
Air Blank	0.000	08:25
Control Test	0.079	08:26
Air Blank	0.000	08:26
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

LADY LAKE P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001177
07/03/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:28
Control Test	0.196	08:29
Air Blank	0.000	08:29
Control Test	0.195	08:30
Air Blank	0.000	08:30
Control Test	0.196	08:31
Air Blank	0.000	08:32
Control Test Stats		
Average	0.1957	
Std Dev	0.0006	
Rel Std Dev(%)	0.2951	

LADY LAKE P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001177
07/03/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:33
Control Test	0.077	08:33
Air Blank	0.000	08:34
Control Test	0.076	08:34
Air Blank	0.000	08:35
Control Test	0.076	08:35
Air Blank	0.000	08:36
Control Test Stats		
Average	0.0763	
Std Dev	0.0006	
Rel Std Dev(%)	0.7564	

RS

RS
Operator's Signature
7/31/18

RS
Operator's Signature

RS
Operator's Signature

RS
Operator's Signature



Florida Department of Law Enforcement
 Alcohol Testing Program
 2729 Fort Knox Blvd.
 Bldg. 2, Suite 1300
 Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001177, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001177</u>	UNCERTAINTY* ±
Owning Agency:	<u>LADY LAKE P.D.</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>07/03/2018</u>	0.080 g/ 210 L 0.005
Calibration Time:	<u>12:29</u>	0.200 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.
 Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
 *Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/03/2018

Date



DANIELLE M BELL,
 Department Inspector

FDLE/ATP Form 69 March 2018
 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

7/3/18


Optical Bench Calibration Adjustment Data #80-001177 Lady Lake P.D. 7/3/18 RWS

LADY LAKE P.D.
Intoxilyzer - Alcotest Analyzer
Model 8000
07/03/2018
09:57:47

Auto Calibration
Max Power Res Value = 34
Auto Range Res Value = 18

Sol. Value = 0.000 g/210L ***
Fit. Value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12784, Sum Io = 13188
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.0960 (-0.0030)
Sample #2 = 0.1210 (0.0110)
Sample #3 = 0.0910 (0.0570)
Sample #4 = 0.1210 (0.0530)
Avg % Abs = 0.1110 (0.0403)
STD DEV = 0.0173 (0.0255)
REL STD DEV = 15.604 (63.179)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1210 (-0.0030)
Sample #2 = 0.1050 (0.0030)
Sample #3 = 0.1000 (0.0210)
Sample #4 = 0.1030 (0.0050)
Avg % Abs = 0.1027 (0.0097)
STD DEV = 0.0025 (0.0099)
REL STD DEV = 2.451 (102.060)

Sol. Value = 0.040 g/210L ***
Fit. Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12772, Sum Io = 13185
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.8800 (-0.0190)
Sample #2 = 0.8660 (0.0100)
Sample #3 = 0.8720 (0.0440)
Sample #4 = 0.8780 (0.0220)
Avg % Abs = 0.8720 (0.0253)
STD DEV = 0.0060 (0.0172)
REL STD DEV = 0.688 (68.066)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5800 (-0.0210)
Sample #2 = 1.5740 (0.0080)
Sample #3 = 1.5840 (0.0230)
Sample #4 = 1.5640 (0.0130)
Avg % Abs = 1.5740 (0.0147)
STD DEV = 0.0100 (0.0076)
REL STD DEV = 0.635 (52.075)

Sol. Value = 0.100 g/210L ***
Fit. Value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12764, Sum Io = 13180
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.9550 (-0.0130)
Sample #2 = 1.9940 (-0.0050)
Sample #3 = 1.9860 (0.0110)
Sample #4 = 1.9570 (0.0250)
Avg % Abs = 1.9790 (0.0103)
STD DEV = 0.0195 (0.0150)
REL STD DEV = 0.984 (145.269)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6740 (-0.0170)
Sample #2 = 3.6800 (0.0170)
Sample #3 = 3.6240 (0.0370)
Sample #4 = 3.6350 (0.0460)
Avg % Abs = 3.6463 (0.0333)
STD DEV = 0.0297 (0.0148)
REL STD DEV = 0.814 (44.531)

Sol. Value = 0.200 g/210L ***
Fit. Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12758, Sum Io = 13176
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6360 (-0.0100)
Sample #2 = 3.6310 (0.0020)
Sample #3 = 3.6450 (0.0310)
Sample #4 = 3.6150 (0.0450)
Avg % Abs = 3.6303 (0.0260)
STD DEV = 0.0150 (0.0219)
REL STD DEV = 0.413 (84.353)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.8380 (-0.0230)
Sample #2 = 6.8090 (0.0360)
Sample #3 = 6.8100 (0.0580)
Sample #4 = 6.8290 (0.0660)
Avg % Abs = 6.8160 (0.0533)
STD DEV = 0.0113 (0.0155)
REL STD DEV = 0.165 (29.128)

Sol. Value = 0.300 g/210L ***
Fit. Value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12756, Sum Io = 13177
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 5.3160 (0.0010)
Sample #2 = 5.3310 (0.0310)
Sample #3 = 5.3600 (0.0610)
Sample #4 = 5.3840 (0.0640)
Avg % Abs = 5.3583 (0.0520)
STD DEV = 0.0265 (0.0182)
REL STD DEV = 0.495 (35.093)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 9.9150 (0.0110)
Sample #2 = 9.9230 (0.0960)
Sample #3 = 9.9830 (0.1170)
Sample #4 = 10.0320 (0.1140)
Avg % Abs = 9.9793 (0.1090)
STD DEV = 0.0546 (0.0114)
REL STD DEV = 0.547 (10.420)

Sol. Value = 0.000 mg/l or 0.000 g/210L
% Abs = 0.111
Std Dev = 0.02 Rel Std Dev = 15.60
Sol. Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.872
Std Dev = 0.01 Rel Std Dev = 0.69
Sol. Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.979
Std Dev = 0.02 Rel Std Dev = 0.98
Sol. Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.630
Std Dev = 0.02 Rel Std Dev = 0.41
Sol. Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.358
Std Dev = 0.03 Rel Std Dev = 0.50
Zero Order Coef = -333.10
First Order Coef = 2565.45
Second Order Coef = 31.83
Standard Deviation = 88.499664

<<<<< CHANNEL 1 >>>>>
Sol. Val = 0.000 mg/l or 0.000 g/210L
% Abs = 0.111
Std Dev = 0.02 Rel Std Dev = 15.60
Sol. Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.872
Std Dev = 0.01 Rel Std Dev = 0.69
Sol. Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.979
Std Dev = 0.02 Rel Std Dev = 0.98
Sol. Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.630
Std Dev = 0.02 Rel Std Dev = 0.41
Sol. Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.358
Std Dev = 0.03 Rel Std Dev = 0.50
Zero Order Coef = -333.10
First Order Coef = 2565.45
Second Order Coef = 31.83
Standard Deviation = 88.499664

<<<<< CHANNEL 2 >>>>>
Sol. Val = 0.000 mg/l or 0.000 g/210L
% Abs = 0.110
Std Dev = 0.00 Rel Std Dev = 2.45
Sol. Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.574
Std Dev = 0.01 Rel Std Dev = 0.64
Sol. Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.646
Std Dev = 0.03 Rel Std Dev = 0.81
Sol. Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.816
Std Dev = 0.01 Rel Std Dev = 0.17
Sol. Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.979
Std Dev = 0.05 Rel Std Dev = 0.55
Zero Order Coef = -176.00
First Order Coef = 1323.02
Second Order Coef = 12.93
Standard Deviation = 58.266525

<<<<< CHANNEL 1 >>>>>
Sol. Value = 0.080 g/210L ***
Fit. Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
<<<<< CHANNEL 2 >>>>>
Sample #1 = 3037.00
Sample #2 = 2982.00
Sample #3 = 2997.00
Sample #4 = 2982.00
Average Result = 2987.0000
STD DEV = 8.6603
REL STD DEV = 0.290

***** CHANNEL 2
Sample #1 = 3362.00
Sample #2 = 3337.00
Sample #3 = 3375.00
Sample #4 = 3359.00
Average Result = 3357.0000
STD DEV = 19.0788
REL STD DEV = 0.568

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1017
3 um H2O Adjust (mg/l*10,000) = 822
9 um H2O Adjust (mg/l*10,000) = 482
**** AUTO CAL PASS

<<<<< CHANNEL 1 >>>>>
Sol. Value = 0.000 mg/l or 0.000 g/210L
% Abs = 0.111
Std Dev = 0.02 Rel Std Dev = 15.60
Sol. Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.872
Std Dev = 0.01 Rel Std Dev = 0.69
Sol. Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.979
Std Dev = 0.02 Rel Std Dev = 0.98
Sol. Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.630
Std Dev = 0.02 Rel Std Dev = 0.41
Sol. Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.358
Std Dev = 0.03 Rel Std Dev = 0.50
Zero Order Coef = -333.10
First Order Coef = 2565.45
Second Order Coef = 31.83
Standard Deviation = 88.499664

SR

7/3/18
J2

Post Calibration Adjustment

Stability Checks #80-001177 Lady Lake P.D. 7/3/18 *RWB*

DGS

LADY LAKE P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001177
07/03/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:40
Control Test	0.049	10:40
Air Blank	0.000	10:41
Control Test	0.050	10:42
Air Blank	0.000	10:42
Control Test	0.049	10:43
Air Blank	0.000	10:43
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

LADY LAKE P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001177
07/03/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:44
Control Test	0.079	10:45
Air Blank	0.000	10:45
Control Test	0.080	10:46
Air Blank	0.000	10:47
Control Test	0.081	10:47
Air Blank	0.000	10:48
Control Test Stats		
Average	0.0800	
Std Dev	0.0010	
Rel Std Dev(%)	1.2500	

LADY LAKE P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001177
07/03/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:49
Control Test	0.196	10:50
Air Blank	0.000	10:50
Control Test	0.200	10:51
Air Blank	0.000	10:51
Control Test	0.200	10:52
Air Blank	0.000	10:53
Control Test Stats		
Average	0.1987	
Std Dev	0.0023	
Rel Std Dev(%)	1.1625	

LADY LAKE P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001177
07/03/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:54
Control Test	0.079	10:55
Air Blank	0.000	10:55
Control Test	0.079	10:55
Air Blank	0.000	10:56
Control Test	0.080	10:56
Air Blank	0.000	10:57
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

SP

RWB

Operator's Signature

7/3/18
SP

RWB

Operator's Signature

RWB

Operator's Signature

RWB

Operator's Signature



INSTRUMENT PROCESSING SHEET

Agency Lady Lake P.D.S/N 80-001177

Florida Department of Law Enforcement

Date In 12/18/2017 DI Completion Date 01/03/2018 Ship P/U H/D CMI EE

Intake Performed By <u>SP</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>DMB</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>220</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 102</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>30793</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
---	---	--

Final Release Date FDLE JAN 03 2018 Alcohol Testing Program
--

Simulator	Serial #	Lot #/Exp
0.050	SD3962	201707D 07/25/2019
0.080	SD3964	201707E 07/25/2019
0.200	DR3856	201707C 07/24/2019
0.080 DGS	N/A	AG626604 09/22/2018

Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____	Temperature Checks Performed By <u>DMB</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.5</u> External Digital Therm. ID#: <u>300505</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3962</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3964</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>DR3856</u>
---	---

Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																																													
0.000		N/A	N/A																																													
0.040																																																
0.100																																																
0.200																																																
0.300																																																
0.080 DGS	N/A																																															
Simulator	Serial Number	Lot Number	Expiration																																													
0.050																																																
0.080																																																
0.200																																																
0.080 DGS	N/A																																															

Department Inspection Performed By <u>DMB</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1017</u> Instrument <u>1017</u> Mouth Alcohol Solution Lot # <u>2016C</u> Acetone Stock Solution Lot # <u>2017A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD1019</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.050</td> <td>SD3962</td> </tr> <tr> <td>0.080</td> <td>SD3964</td> </tr> <tr> <td>0.200</td> <td>DR3856</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	SD1019	Interferent	SD1021	0.050	SD3962	0.080	SD3964	0.200	DR3856
Simulator	Serial Number											
0.000	SD1019											
Interferent	SD1021											
0.050	SD3962											
0.080	SD3964											
0.200	DR3856											

Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Checks	<input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Other <u>Calibration Certificate</u>
--	---

Notes/Suggested Service: Please remember to update the password for the level 2 menu upon receipt of instrument. DMB 1/3/18

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use
<u>DMB 1/3/18</u> Tech Review / Date <u>J. Dehan 1/3/18</u> Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LADY LAKE P.D.

Time of Inspection: 14:43

Date of Inspection: 01/03/2018

Serial Number: 80-001177

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626604 Exp: 09/22/2018
0.000	0.050	0.082	0.201	0.078
0.000	0.051	0.082	0.202	0.078
0.000	0.051	0.082	0.203	0.078
0.000	0.051	0.082	0.203	0.078
0.000	0.051	0.083	0.203	0.078
0.000	0.051	0.082	0.203	0.078
0.000	0.051	0.082	0.202	0.078
0.000	0.051	0.083	0.203	0.078
0.000	0.051	0.083	0.203	0.078
0.000	0.051	0.082	0.202	0.078

Standard Deviations	0.0003	0.0004	0.0007	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

RAM

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Danielle M Bell

DANIELLE M BELL

Signature and Printed Name

01/03/2018
Date

*1/3/18
JD*

Stability Checks #80-001177 Lady Lake P.D. 1/3/18

DS

LADY LAKE P.O.
Intoxilyzer - Alcohol Analyzer
Model 8000
01/03/2018
Software: 8100.27

LADY LAKE P.O.
Intoxilyzer - Alcohol Analyzer
Model 8000
01/03/2018
Software: 8100.27

LADY LAKE P.O.
Intoxilyzer - Alcohol Analyzer
Model 8000
01/03/2018
Software: 8100.27

LADY LAKE P.O.
Intoxilyzer - Alcohol Analyzer
Model 8000
01/03/2018
Software: 8100.27

LADY LAKE P.O.
Intoxilyzer - Alcohol Analyzer
Model 8000
01/03/2018
Software: 8100.27

LADY LAKE P.O.
Intoxilyzer - Alcohol Analyzer
Model 8000
01/03/2018
Software: 8100.27

LADY LAKE P.O.
Intoxilyzer - Alcohol Analyzer
Model 8000
01/03/2018
Software: 8100.27

SN 80-001177

Test	g/210L	Time
Air Blank	0.000	12:22
Control Test	0.050	12:23
Air Blank	0.000	12:24
Control Test	0.050	12:24
Air Blank	0.000	12:25
Control Test	0.050	12:26
Air Blank	0.000	12:26
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Test	g/210L	Time
Air Blank	0.000	12:27
Control Test	0.081	12:28
Air Blank	0.000	12:28
Control Test	0.080	12:29
Air Blank	0.000	12:30
Control Test	0.081	12:30
Air Blank	0.000	12:31
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

Test	g/210L	Time
Air Blank	0.000	12:32
Control Test	0.200	12:33
Air Blank	0.000	12:33
Control Test	0.199	12:34
Air Blank	0.000	12:35
Control Test	0.199	12:35
Air Blank	0.000	12:36
Control Test Stats		
Average	0.1993	
Std Dev	0.0006	
Rel Std Dev(%)	0.2896	

Test	g/210L	Time
Air Blank	0.000	12:37
Control Test	0.078	12:37
Air Blank	0.000	12:38
Control Test	0.078	12:38
Air Blank	0.000	12:39
Control Test	0.079	12:39
Air Blank	0.000	12:39
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

gum

DS

Operator's Signature

DS

Operator's Signature

DS

Operator's Signature

DS

Operator's Signature

1/3/18
90



Florida Department of Law Enforcement
 Alcohol Testing Program
 2729 Fort Knox Blvd.
 Bldg. 2, Suite 1300
 Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001177, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001177</u>	UNCERTAINTY* ±	
Owning Agency:	<u>LADY LAKE P.D.</u>	0.05 g/ 210 L	0.004
Calibration Date:	<u>01/03/2018</u>	0.08 g/ 210 L	0.005
Calibration Time:	<u>14:43</u>	0.20 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by Calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the Target Alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS) ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/03/2018

Date

DANIELLE M BELL,
 Department Inspector

FDLE/ATP Form 69 January 2018
 Issuing Authority: Alcohol Testing Program

1/3/18