



INSTRUMENT PROCESSING SHEET

Agency Walton County SO

S/N 80-001311

Florida Department of Law Enforcement

Date In 6/19/18

DI Completion Date 6/29/18

Ship P/U H/D CMI EE

Intake Performed By <u>SQC</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>Center port is missing a screw</u>		Quality Checks Performed By <u>PJGM</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>131</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>.140</u> (.139 - .169) 36 mm <u>.160</u> (.156 - .190) 53 mm <u>.242</u> (.228 - .278) 103 mm <u>.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks		Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																							
Final Release Date <p style="text-align: center;">FDLE</p> <p style="text-align: center;">JUL 02 2018</p> <p style="text-align: center;">Alcohol Testing Program</p>		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1021</td> <td>261707D 7/25/19</td> </tr> <tr> <td>0.080</td> <td>DR1275</td> <td>201707E 7/25/19</td> </tr> <tr> <td>0.200</td> <td>SD1019</td> <td>201707C 7/24/19</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG805702 2/26/20</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	SD1021	261707D 7/25/19	0.080	DR1275	201707E 7/25/19	0.200	SD1019	201707C 7/24/19	0.080 DGS	N/A	AG805702 2/26/20	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>PJGM</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.5</u> External Digital Therm. ID#: <u>300503</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1021</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>DR1275</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1019</u>																								
Simulator	Serial #	Lot #/Exp																																									
0.050	SD1021	261707D 7/25/19																																									
0.080	DR1275	201707E 7/25/19																																									
0.200	SD1019	201707C 7/24/19																																									
0.080 DGS	N/A	AG805702 2/26/20																																									
Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____		Department Inspection Performed By <u>PJGM</u> Barometric Pressure ID# <u>28427</u> Gauge <u>1015</u> Instrument <u>1012</u> Mouth Alcohol Solution Lot # <u>2017-B</u> Acetone Stock Solution Lot # <u>2018-A</u>																																									
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr> <tr><td>0.040</td><td></td><td></td><td></td></tr> <tr><td>0.100</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.300</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr><td>0.000</td><td>G11621</td></tr> <tr><td>Interferent</td><td>DR3855</td></tr> <tr><td>0.050</td><td>SD1021</td></tr> <tr><td>0.080</td><td>DR1275</td></tr> <tr><td>0.200</td><td>SD1019</td></tr> </tbody> </table>		Simulator	Serial Number	0.000	G11621	Interferent	DR3855	0.050	SD1021	0.080	DR1275	0.200	SD1019
Simulator	Serial Number	Lot Number	Expiration																																								
0.000		N/A	N/A																																								
0.040																																											
0.100																																											
0.200																																											
0.300																																											
0.080 DGS	N/A																																										
Simulator	Serial Number																																										
0.000	G11621																																										
Interferent	DR3855																																										
0.050	SD1021																																										
0.080	DR1275																																										
0.200	SD1019																																										
<input type="checkbox"/> Post Calibration Adjustment Stability Checks		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.050</td><td></td><td></td><td></td></tr> <tr><td>0.080</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A																						
Simulator	Serial Number	Lot Number	Expiration																																								
0.050																																											
0.080																																											
0.200																																											
0.080 DGS	N/A																																										
Notes/Suggested Service: _____ _____ _____ _____ _____		<table border="1"> <tr> <td><input checked="" type="checkbox"/> Form 41</td> <td><input type="checkbox"/> Post-Stability Checks</td> </tr> <tr> <td><input checked="" type="checkbox"/> Stability Checks</td> <td><input type="checkbox"/> Flow Calibration</td> </tr> <tr> <td><input checked="" type="checkbox"/> Calibration Certificate</td> <td><input type="checkbox"/> Form 40</td> </tr> <tr> <td><input type="checkbox"/> Calibration Adjustment</td> <td><input type="checkbox"/> Other _____</td> </tr> </table>		<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Post-Stability Checks	<input checked="" type="checkbox"/> Stability Checks	<input type="checkbox"/> Flow Calibration	<input checked="" type="checkbox"/> Calibration Certificate	<input type="checkbox"/> Form 40	<input type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Other _____																																
<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Post-Stability Checks																																										
<input checked="" type="checkbox"/> Stability Checks	<input type="checkbox"/> Flow Calibration																																										
<input checked="" type="checkbox"/> Calibration Certificate	<input type="checkbox"/> Form 40																																										
<input type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Other _____																																										
		<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																									
		<u>S. Platt 7/2/18</u> <u>J. Dahan 7/2/18</u> Tech Review / Date Admin Review / Date																																									

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO
Time of Inspection: 10:36

Date of Inspection: 06/29/2018

Serial Number: 80-001311
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805702 Exp: 02/26/2020
0.000	0.049	0.080	0.197	0.079
0.000	0.049	0.081	0.200	0.079
0.000	0.049	0.081	0.200	0.079
0.000	0.049	0.081	0.200	0.079
0.000	0.049	0.081	0.200	0.079
0.000	0.049	0.081	0.200	0.079
0.000	0.049	0.081	0.200	0.079
0.000	0.050	0.081	0.200	0.079
0.000	0.049	0.081	0.200	0.079
0.000	0.049	0.081	0.200	0.078
Standard Deviations	0.0003	0.0003	0.0009	0.0003

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

SP

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick J Murphy

PATRICK J MURPHY

Signature and Printed Name

06/29/2018
Date

7/2/18
JD

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001311
06/29/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:31
Control Test	0.049	07:32
Air Blank	0.000	07:32
Control Test	0.049	07:33
Air Blank	0.000	07:34
Control Test	0.049	07:34
Air Blank	0.000	07:35
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy
Operator's Signature

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001311
06/29/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:21
Control Test	0.079	07:22
Air Blank	0.000	07:22
Control Test	0.079	07:23
Air Blank	0.000	07:23
Control Test	0.080	07:24
Air Blank	0.000	07:25
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

P Murphy
Operator's Signature

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001311
06/29/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:37
Control Test	0.199	07:38
Air Blank	0.000	07:38
Control Test	0.200	07:39
Air Blank	0.000	07:40
Control Test	0.199	07:40
Air Blank	0.000	07:41
Control Test Stats		
Average	0.1993	
Std Dev	0.0006	
Rel Std Dev(%)	0.2896	

P Murphy
Operator's Signature

SP

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001311
06/29/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:43
Control Test	0.079	07:43
Air Blank	0.000	07:44
Control Test	0.079	07:44
Air Blank	0.000	07:45
Control Test	0.078	07:45
Air Blank	0.000	07:45
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

DGS

P Murphy
Operator's Signature

7/2/18
JL



Florida Department of Law Enforcement
 Alcohol Testing Program
 2729 Fort Knox Blvd.
 Bldg. 2, Suite 1300
 Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001311, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001311</u>	UNCERTAINTY* ±
Owning Agency:	<u>WALTON COUNTY SO</u>	0.050 g/ 210 L
Calibration Date:	<u>06/29/2018</u>	0.080 g/ 210 L
Calibration Time:	<u>10:36</u>	0.200 g/ 210 L
		0.080 g/ 210 L Dry Gas Control
		0.005

All results are reported in g/ 210 L.
 Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
 *Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/29/2018 _____
 Date Patrick J Murphy
PATRICK J MURPHY
 Department Inspector

FDLE/ATP Form 69 March 2018
 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Handwritten initials/signature