



INSTRUMENT PROCESSING SHEET

Agency Miccosukee Police Department S/N 80-001657

Florida Department of Law Enforcement Date In 1/16/2018 DI Completion Date 1/17/2018 Ship P/U H/D CMI EE

Intake Performed By <u>DELL</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____		Quality Checks Performed By <u>DELL</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>182</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 101</u> 32 mm <u>.160</u> (.139 - .169) 36 mm <u>.175</u> (.156 - .190) 53 mm <u>.246</u> (.228 - .278) 103 mm <u>.511</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks		Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																
Final Release Date FDLE JAN 30 2018 Alcohol Testing Program		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>SD3963</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG626604 09/22/2018</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	SD3967	201707D 07/25/2019	0.080	SD3963	201707E 07/25/2019	0.200	SD3969	201707C 07/24/2019	0.080 DGS	N/A	AG626604 09/22/2018	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>DELL</u> <input checked="" type="checkbox"/> Lab Temp °C <u>20.88c</u> External Digital Therm. ID#: <u>300918</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD3967</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD3963</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD3969</u>	
Simulator	Serial #	Lot #/Exp																		
0.050	SD3967	201707D 07/25/2019																		
0.080	SD3963	201707E 07/25/2019																		
0.200	SD3969	201707C 07/24/2019																		
0.080 DGS	N/A	AG626604 09/22/2018																		

Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____				Department Inspection Performed By <u>DELL</u> Barometric Pressure ID# <u>68639</u> Gauge <u>1022</u> Instrument <u>1021</u> Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2017-A</u>																																											
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>				Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD3965</td> </tr> <tr> <td>Interferent</td> <td>SD3966</td> </tr> <tr> <td>0.050</td> <td>SD3967</td> </tr> <tr> <td>0.080</td> <td>SD3963</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> </tr> </tbody> </table>				Simulator	Serial Number	0.000	SD3965	Interferent	SD3966	0.050	SD3967	0.080	SD3963	0.200	SD3969
Simulator	Serial Number	Lot Number	Expiration																																												
0.000		N/A	N/A																																												
0.040																																															
0.100																																															
0.200																																															
0.300																																															
0.080 DGS	N/A																																														
Simulator	Serial Number																																														
0.000	SD3965																																														
Interferent	SD3966																																														
0.050	SD3967																																														
0.080	SD3963																																														
0.200	SD3969																																														
<input type="checkbox"/> Post Calibration Adjustment Stability Checks				Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____																																											

Notes/Suggested Service: E-mailed **APPROVED**

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Pgm 1/29/18 J. Dehn 1/30/18
 Tech Review / Date Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MICCOSUKEE PD
Time of Inspection: 09:00

Date of Inspection: 01/17/2018

Serial Number: 80-001657
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG526604 Exp: 09/22/2018
0.000	0.050	0.081	0.203	0.080
0.000	0.050	0.082	0.203	0.081
0.000	0.050	0.082	0.203	0.081
0.000	0.050	0.082	0.203	0.081
0.000	0.050	0.082	0.203	0.081
0.000	0.051	0.083	0.202	0.081
0.000	0.050	0.082	0.203	0.081
0.000	0.051	0.082	0.203	0.081
0.000	0.051	0.083	0.203	0.081
0.000	0.051	0.083	0.203	0.081
0.000	0.051	0.083	0.203	0.082

Standard Deviations	0.0005	0.0006	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Reyes Rivera

DAVID E REYES-RIVERA

Signature and Printed Name

01/17/2018
Date

1/30/18
JD

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-001657	Miccosukee Police Department	01/17/2018	<i>DKL</i>

0.05g/210L	0.08g/210L	0.20g/210L	0.08g/210L
SN: SD3967 Temp: 34.06c	SN: SD3963 Temp: 34.06c	SN: SD3969 Temp: 34.08c	Lot AG626604
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

Test	g/210L	Time	Test	g/210L	Time	Test	g/210L	Time	Test	g/210L	Time
<p>MICCOSUKEE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001657 01/17/2018 Software: 8100.27</p>											
Air Blank	0.000	06:27	Air Blank	0.000	06:32	Air Blank	0.000	06:36	Air Blank	0.000	06:43
Control Test	0.050	06:28	Control Test	0.083	06:33	Control Test	0.202	06:38	Control Test	0.080	06:44
Air Blank	0.000	06:28	Air Blank	0.000	06:33	Air Blank	0.000	06:39	Air Blank	0.000	06:44
Control Test	0.051	06:29	Control Test	0.083	06:34	Control Test	0.201	06:39	Control Test	0.081	06:44
Air Blank	0.000	06:29	Air Blank	0.000	06:35	Air Blank	0.000	06:40	Air Blank	0.000	06:45
Control Test	0.050	06:30	Control Test	0.082	06:35	Control Test	0.202	06:41	Control Test	0.081	06:45
Air Blank	0.000	06:31	Air Blank	0.000	06:36	Air Blank	0.000	06:41	Air Blank	0.000	06:46
<p>Control Test Stats</p> <p>Average 0.0503 Average 0.2017 Average 0.0807</p> <p>Std Dev 0.0006 Std Dev 0.0006 Std Dev 0.0006</p> <p>Rel. Std Dev(%) 1.1471 Rel. Std Dev(%) 0.5984 Rel. Std Dev(%) 0.7157</p>											
<p><i>DKL</i> Operator's Signature</p>			<p><i>DKL</i> Operator's Signature</p>			<p><i>DKL</i> Operator's Signature</p>			<p><i>DKL</i> Operator's Signature</p>		

DKL
8/10/18

WPA



Florida Department of Law Enforcement
 Alcohol Testing Program
 4700 Terminal Drive, Suite 1
 Ft. Myers, FL 33907

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001657, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001657</u>	UNCERTAINTY* ±
Owning Agency:	<u>MICCOSUKEE PD</u>	0.05 g/ 210 L
Calibration Date:	<u>01/17/2018</u>	0.08 g/ 210 L
Calibration Time:	<u>09:00</u>	0.20 g/ 210 L
		0.080 g/ 210 L Dry Gas Control
		0.005

All results are reported in g/210 L.

Bias is limited by Calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the Taretget Alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS) ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/17/2018 Date *David Reyes-Rivera*
DAVID E REYES-RIVERA,
 Department Inspector

FDLE/ATP Form 69 January 2018
 Issuing Authority: Alcohol Testing Program

Pagan

1/30/18
SD