



INSTRUMENT PROCESSING SHEET

Agency Miami Gardens Police Department S/N 80-002987

Florida Department of Law Enforcement Date In 10/29/2018 DI Completion Date 10/29/2018 Ship P/U H/D CMI EE

Intake Performed By DELL Quality Checks Performed By DELL Flow Calibration Performed By... Final Release Date FDLE NOV 05 2018 Alcohol Testing Program

Calibration Adjustment Performed By Department Inspection Performed By DELL Barometric Pressure Gauge ID# 68639 Gauge 1019 Instrument 1020

Notes/Suggested Service: E-mailed APPROVED Instrument Complies with Chapter 11D-8, FAC

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI GARDENS PD  
Time of Inspection: 09:51

Date of Inspection: 10/29/2018

Serial Number: 80-002987  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.050	0.081	0.197	0.077
0.000	0.050	0.082	0.198	0.077
0.000	0.051	0.082	0.198	0.078
0.000	0.051	0.082	0.199	0.077
0.000	0.051	0.082	0.198	0.077
0.000	0.051	0.082	0.198	0.078
0.000	0.051	0.083	0.200	0.077
0.000	0.052	0.082	0.199	0.078
0.000	0.051	0.082	0.200	0.078
0.000	0.051	0.082	0.200	0.078

Standard Deviations	0.0005	0.0004	0.0010	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

*PDM*

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*David E Reyes Rivera*

DAVID E REYES-RIVERA

Signature and Printed Name

10/29/2018  
Date

*11/5/18  
JO*

4678

<b>TYPE OF TEST</b>	<b>SERIAL NUMBER</b>	<b>AGENCY</b>	<b>DATE</b>	<b>PERFORMED BY</b>
Stabilities	80-002987	Miami Gardens Police Department	10/29/2018	<i>See</i>

<b>0.05g/210L</b>	<b>0.08g/210L</b>	<b>0.20g/210L</b>	<b>DGS 0.08g/210L</b>
SN: SD3967 Temp: 34.08c	SN: SD3968 Temp: 34.08c	SN: SD3969 Temp: 34.09c	Lot AG805701
<b>0.047 to 0.053</b> <input checked="" type="checkbox"/>	<b>0.077 to 0.083</b> <input checked="" type="checkbox"/>	<b>0.194 to 0.206</b> <input checked="" type="checkbox"/>	<b>0.077 to 0.083</b> <input checked="" type="checkbox"/>

MIAMI GARDENS PD Intoxilyzer - Alconol Analyzer Model: 8000 10/29/2018 Software: 8100.27	MIAMI GARDENS PD Intoxilyzer - Alconol Analyzer Model: 8000 10/29/2018 Software: 8100.27	MIAMI GARDENS PD Intoxilyzer - Alconol Analyzer Model: 8000 10/29/2018 Software: 8100.27	MIAMI GARDENS PD Intoxilyzer - Alconol Analyzer Model: 8000 10/25/2018 Software: 8100.27
Test	Test	Test	Test
g/210L	g/210L	g/210L	g/210L
Time	Time	Time	Time
Air Blank 0.000 07:51	Air Blank 0.000 07:56	Air Blank 0.000 08:01	Air Blank 0.000 08:14
Control Test 0.049 07:52	Control Test 0.090 07:57	Control Test 0.197 08:01	Control Test 0.078 08:14
Air Blank 0.000 07:52	Air Blank 0.000 07:57	Air Blank 0.000 08:02	Air Blank 0.000 08:15
Control Test 0.050 07:53	Control Test 0.081 07:58	Control Test 0.198 08:03	Control Test 0.077 08:15
Air Blank 0.000 07:53	Air Blank 0.000 07:58	Air Blank 0.000 08:03	Air Blank 0.000 08:16
Control Test 0.050 07:54	Control Test 0.081 07:59	Control Test 0.196 08:04	Control Test 0.078 08:16
Air Blank 0.000 07:55	Air Blank 0.000 08:00	Air Blank 0.000 08:05	Air Blank 0.000 08:16
Control Test Stats	Control Test Stats	Control Test Stats	Control Test Stats
Average 0.0497	Average 0.0807	Average 0.1970	Average 0.0777
Std Dev 0.0005	Std Dev 0.0006	Std Dev 0.0010	Std Dev 0.0006
Rel. Std Dev(%) 1.1625	Rel. Std Dev(%) 0.7157	Rel. Std Dev(%) 0.5076	Rel. Std Dev(%) 0.7434

*See*

Operator's Signature

*See*

Operator's Signature

*See*

Operator's Signature

*See*

Operator's Signature

11/5/18  
*JD*



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-002987, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-002987</u>	UNCERTAINTY* ±	
Owning Agency:	<u>MIAMI GARDENS PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>10/29/2018</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>09:51</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

10/29/2018 \_\_\_\_\_ Date David Reyes-Rivera  
**DAVID E REYES-RIVERA,**  
 Department Inspector

FDLE/ATP Form 69 July 2018  
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

11/5/18  
DR

DR

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI GARDENS PD  
Time of Inspection: 07:43

Date of Inspection: 10/29/2018

Serial Number: 80-002987  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_


**Remarks:**

COMPLIANCE NOT DETERMINED AI NOT CONDUCTED

*Qom*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


DAVID E REYES-RIVERA

Signature and Printed Name

10/29/2018  
Date

*11/5/18*  
*JO*