



# INSTRUMENT PROCESSING SHEET

Agency Florida Highway PatrolS/N 80-003411Florida Department of  
Law EnforcementDate In 2/2/2018DI Completion Date 02/22/2018 Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>TG</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>DMB</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>210</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 105</u> 32 mm <u>0.160</u> (.139 - .169) 36 mm <u>0.179</u> (.156 - .190) 53 mm <u>0.250</u> (.228 - .278) 103 mm <u>0.523</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks	<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
---	---	--

**Final Release Date**  
**FDLE**  
**FEB 23 2018**  
**Alcohol Testing Program**

Simulator	Serial #	Lot #/Exp
0.050	G2835	201707D 07/25/2019
0.080	SD1013	201707E 07/25/2019
0.200	SD1025	201707C 07/24/2019
0.080 DGS	N/A	AG708807 03/29/2019

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Temperature Checks** Performed By DMB  
 Lab Temp °C 22.3  
 External Digital Therm. ID#: 300503  
 34°C +/- .2 Serial #: G2835  
 34°C +/- .2 Serial #: SD1013  
 34°C +/- .2 Serial #: SD1025

**Calibration Adjustment** Performed By DMB  
 Barometric Pressure Gauge 1030 ID # 28662

Simulator	Serial Number	Lot Number	Expiration
0.000	G2879	N/A	N/A
0.040	SD1024	16320	10/21/2018
0.100	G2834	17280	09/11/2019
0.200	SD1011	17090	02/24/2019
0.300	DR1275	17140	05/15/2019
0.080 DGS	N/A	17817080A2	08/05/2019

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050	G2835	201707D	07/25/2019
0.080	SD1013	201707E	07/25/2019
0.200	SD1025	201707C	07/24/2019
0.080 DGS	N/A	AG715202	06/01/2019

**Department Inspection** Performed By DMB  
 Barometric Pressure ID# 28427  
 Gauge 1028 Instrument 1029  
 Mouth Alcohol Solution Lot # 2016-C  
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.000	G2880
Interferent	G8144
0.050	G2835
0.080	SD1013
0.200	SD1025

**Attachments**

Form 41  Post-Stability Checks  
 Stability Checks  Flow Calibration  
 Calibration Certificate  Form 40  
 Calibration Adjustment  Other \_\_\_\_\_

Notes/Suggested Service: Performed optical bench calibration adjustment to bring values closer to nominal. DMB 2/22/18  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

DMB 2/23/18 J. Rahon 2/23/18  
 Tech Review / Date Admin Review / Date

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL  
Time of Inspection: 13:32

Date of Inspection: 02/22/2018

Serial Number: 80-003411  
Software: 8100.27

*EC*

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG715202 Exp: 06/01/2019
0.000	0.049	0.080	0.199	0.079
0.000	0.049	0.081	0.200	0.079
0.000	0.050	0.082	0.200	0.079
0.000	0.050	0.081	0.200	0.079
0.000	0.049	0.082	0.200	0.079
0.000	0.049	0.081	0.200	0.079
0.000	0.050	0.082	0.200	0.079
0.000	0.050	0.081	0.201	0.079
0.000	0.049	0.081	0.200	0.079
0.000	0.050	0.081	0.200	0.080

Standard Deviations	0.0005	0.0006	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

*EC*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Danielle M Bell*

DANIELLE M BELL

Signature and Printed Name

02/22/2018  
Date

*2/23/18  
JD*

Stability Checks #80-003411 FL Highway Patrol 2/7/18 ~~2018~~

*ES*

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003411  
02/07/2018  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:53
Control Test	0.050	09:53
Air Blank	0.000	09:54
Control Test	0.050	09:54
Air Blank	0.000	09:55
Control Test	0.050	09:55
Air Blank	0.000	09:56
Control Test	0.000	09:56
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003411  
02/07/2018  
Software: 8100.27

Test	g/210C	Time
Air Blank	0.000	09:57
Control Test	0.081	09:58
Air Blank	0.000	09:59
Control Test	0.081	09:59
Air Blank	0.000	10:00
Control Test	0.082	10:01
Air Blank	0.000	10:01
Control Test	0.000	10:01
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003411  
02/07/2018  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:03
Control Test	0.199	10:04
Air Blank	0.000	10:04
Control Test	0.199	10:05
Air Blank	0.000	10:06
Control Test	0.200	10:06
Air Blank	0.000	10:07
Control Test	0.000	10:07
Control Test Stats		
Average	0.1993	
Std Dev	0.0006	
Rel Std Dev(%)	0.2896	

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003411  
02/07/2018  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:08
Control Test	0.078	10:09
Air Blank	0.000	10:09
Control Test	0.078	10:09
Air Blank	0.000	10:10
Control Test	0.078	10:10
Air Blank	0.000	10:11
Control Test	0.000	10:11
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*ES*

*ES*

*ES*

*ES*

Operator's Signature

Operator's Signature

Operator's Signature

2/23/18  
JA



Florida Department of Law Enforcement  
 Alcohol Testing Program  
 2729 Fort Knox Blvd.  
 Bldg. 2, Suite 1300  
 Tallahassee, FL 32308

# Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003411, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-003411</u>	UNCERTAINTY* ±
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.05 g/ 210 L
Calibration Date:	<u>02/22/2018</u>	0.08 g/ 210 L
Calibration Time:	<u>13:32</u>	0.20 g/ 210 L
		0.080 g/ 210 L Dry Gas Control
		0.005

All results are reported in g/ 210 L.  
 Bias is limited by Calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the Taretget Alcohol concentration.  
 \*Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS) ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/22/2018

Date

**DANIELLE M BELL,**  
 Department Inspector

FDLE/ATP Form 69 January 2018  
 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

*2/23/18*  
*[Signature]*

Optical Bench Calibration Adjustment #80-003411 FL Highway Patrol 2/22/18 ~~8085~~

FL HIGHWAY PATROL  
 Intoxilyzer - Alcoloh Analyzer  
 Model 8000  
 02/22/2018 09:49:18  
 SN 80-003411  
 Auto Calibration  
 Max Power Res Value = 20  
 Auto Range Res Value = 8

Internal printer  
 Paper jammed.  
 See external  
 printer print-  
 out. ~~8085~~

\*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 1.4640 (0.0040)  
 Sample #2 = 1.4900 (-0.0250)  
 Sample #3 = 1.4630 (0.0000)  
 Sample #4 = 1.5160 (-0.0100)  
 Avg % Abs = 1.4897 (-0.0117)  
 STD DEV = 0.0265 (0.0126)  
 REL STD DEV = 1.779 (107.855)

\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
 Sol Value = 0.100 g/210L \*\*\*  
 Fit Value = 0.4762 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12600, Sum Io = 13412  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 1.8340 (-0.0120)  
 Sample #2 = 1.8070 (-0.0060)  
 Sample #3 = 1.8450 (-0.0200)  
 Sample #4 = 1.8010 (0.0180)  
 Avg % Abs = 1.8177 (-0.0060)  
 STD DEV = 0.0239 (0.0140)  
 REL STD DEV = 1.313 (233.333)

\*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
 Sol Value = 0.040 g/210L \*\*\*  
 Fit Value = 0.1905 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12604, Sum Io = 13412  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 0.1030 (-0.0140)  
 Sample #2 = 0.0850 (0.0010)  
 Sample #3 = 0.0780 (0.0040)  
 Sample #4 = 0.1020 (-0.0130)  
 Avg % Abs = 0.0883 (-0.0027)  
 STD DEV = 0.0123 (0.0091)  
 REL STD DEV = 13.972 (340.267)

\*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
 Sol Value = 0.080 g/210L \*\*\*  
 Fit Value = 0.3810 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 3384.00  
 Sample #2 = 3382.00  
 Sample #3 = 3387.00  
 Sample #4 = 3371.00  
 Average Result = 3380.0000  
 STD DEV = 8.1854  
 REL STD DEV = 0.242

\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
 Sol Value = 0.300 g/210L \*\*\*  
 Fit Value = 1.4286 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12595, Sum Io = 13410  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 5.2080 (-0.0160)  
 Sample #2 = 5.1840 (0.0000)  
 Sample #3 = 5.1840 (0.0000)  
 Sample #4 = 5.1470 (0.0210)  
 Avg % Abs = 5.1717 (0.0077)  
 STD DEV = 0.0214 (0.0121)  
 REL STD DEV = 0.413 (173.205)

\*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
 Sol Value = 0.200 g/210L \*\*\*  
 Fit Value = 0.9524 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12599, Sum Io = 13413  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 3.5740 (-0.0210)  
 Sample #2 = 3.5420 (0.0010)  
 Sample #3 = 3.5470 (-0.0010)  
 Sample #4 = 3.5100 (0.0110)  
 Avg % Abs = 3.5330 (0.0037)  
 STD DEV = 0.0201 (0.0064)  
 REL STD DEV = 0.568 (175.339)

\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
 Sol Value = 0.200 g/210L \*\*\*  
 Fit Value = 0.9524 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12599, Sum Io = 13413  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 3.5570 (-0.0020)  
 Sample #2 = 3.5270 (0.0200)  
 Sample #3 = 3.5690 (0.0070)  
 Sample #4 = 3.5190 (0.0360)  
 Avg % Abs = 3.5383 (0.0210)  
 STD DEV = 0.0269 (0.0145)  
 REL STD DEV = 0.759 (69.171)

\*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
 Sol Value = 0.300 g/210L \*\*\*  
 Fit Value = 1.4286 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12595, Sum Io = 13410  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 5.2080 (-0.0160)  
 Sample #2 = 5.1840 (0.0000)  
 Sample #3 = 5.1840 (0.0000)  
 Sample #4 = 5.1470 (0.0210)  
 Avg % Abs = 5.1717 (0.0077)  
 STD DEV = 0.0214 (0.0121)  
 REL STD DEV = 0.413 (173.205)

\*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
 Sol Value = 0.080 g/210L \*\*\*  
 Fit Value = 0.3810 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 3384.00  
 Sample #2 = 3382.00  
 Sample #3 = 3387.00  
 Sample #4 = 3371.00  
 Average Result = 3380.0000  
 STD DEV = 8.1854  
 REL STD DEV = 0.242

\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
 Sol Value = 0.300 g/210L \*\*\*  
 Fit Value = 1.4286 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12595, Sum Io = 13410  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 5.2080 (-0.0160)  
 Sample #2 = 5.1840 (0.0000)  
 Sample #3 = 5.1840 (0.0000)  
 Sample #4 = 5.1470 (0.0210)  
 Avg % Abs = 5.1717 (0.0077)  
 STD DEV = 0.0214 (0.0121)  
 REL STD DEV = 0.413 (173.205)

\*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
 Sol Value = 0.200 g/210L \*\*\*  
 Fit Value = 0.9524 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12599, Sum Io = 13413  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 3.5740 (-0.0210)  
 Sample #2 = 3.5420 (0.0010)  
 Sample #3 = 3.5470 (-0.0010)  
 Sample #4 = 3.5100 (0.0110)  
 Avg % Abs = 3.5330 (0.0037)  
 STD DEV = 0.0201 (0.0064)  
 REL STD DEV = 0.568 (175.339)

\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
 Sol Value = 0.200 g/210L \*\*\*  
 Fit Value = 0.9524 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12599, Sum Io = 13413  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 3.5570 (-0.0020)  
 Sample #2 = 3.5270 (0.0200)  
 Sample #3 = 3.5690 (0.0070)  
 Sample #4 = 3.5190 (0.0360)  
 Avg % Abs = 3.5383 (0.0210)  
 STD DEV = 0.0269 (0.0145)  
 REL STD DEV = 0.759 (69.171)

\*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
 Sol Value = 0.300 g/210L \*\*\*  
 Fit Value = 1.4286 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12595, Sum Io = 13410  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 5.2080 (-0.0160)  
 Sample #2 = 5.1840 (0.0000)  
 Sample #3 = 5.1840 (0.0000)  
 Sample #4 = 5.1470 (0.0210)  
 Avg % Abs = 5.1717 (0.0077)  
 STD DEV = 0.0214 (0.0121)  
 REL STD DEV = 0.413 (173.205)

Solution Stats Quadratic Fit Chan 2  
 Act Fit Residual  
 g/210L g/210L g/210L  
 0.000 0.000 -0.0003  
 0.040 0.040 0.0002  
 0.100 0.099 0.0006  
 0.200 0.201 -0.0007  
 0.300 0.300 0.0003

Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1030  
 3 um H2O Adjust (mg/l\*10,000) = 429  
 9 um H2O Adjust (mg/l\*10,000) = 369  
 \*\*\*\*\* AUTO CAL PASS

\*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
 Sol Value = 0.040 g/210L \*\*\*  
 Fit Value = 0.1905 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12604, Sum Io = 13412  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 0.1030 (-0.0140)  
 Sample #2 = 0.0850 (0.0010)  
 Sample #3 = 0.0780 (0.0040)  
 Sample #4 = 0.1020 (-0.0130)  
 Avg % Abs = 0.0883 (-0.0027)  
 STD DEV = 0.0123 (0.0091)  
 REL STD DEV = 13.972 (340.267)

\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
 Sol Value = 0.200 g/210L \*\*\*  
 Fit Value = 0.9524 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12599, Sum Io = 13413  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 3.5570 (-0.0020)  
 Sample #2 = 3.5270 (0.0200)  
 Sample #3 = 3.5690 (0.0070)  
 Sample #4 = 3.5190 (0.0360)  
 Avg % Abs = 3.5383 (0.0210)  
 STD DEV = 0.0269 (0.0145)  
 REL STD DEV = 0.759 (69.171)

\*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
 Sol Value = 0.300 g/210L \*\*\*  
 Fit Value = 1.4286 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12595, Sum Io = 13410  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 5.2080 (-0.0160)  
 Sample #2 = 5.1840 (0.0000)  
 Sample #3 = 5.1840 (0.0000)  
 Sample #4 = 5.1470 (0.0210)  
 Avg % Abs = 5.1717 (0.0077)  
 STD DEV = 0.0214 (0.0121)  
 REL STD DEV = 0.413 (173.205)

FL HIGHWAY PATROL  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-003411  
 02/22/2018 09:49:18

Optical Bench Cal Adjust  
 Data  
 2/22/18 RMB

Auto Calibration

	<<<<<	3um	>>>>>	<<<<<	9um	>>>>>
-----						
Solution =	0.000 g/210L	or	0.0000 mg/l,	Samples =	4,	Discarded = 1
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)	
Sample #1	0.0600	(-0.0050)		0.1030	(-0.0140)	
Sample #2	0.0480	(0.0230)		0.0850	(0.0010)	
Sample #3	0.0430	(0.0280)		0.0780	(0.0040)	
Sample #4	0.0330	(0.0400)		0.1020	(-0.0130)	
Avg % Abs	0.0413	(0.0303)		0.0883	(-0.0027)	
STD DEV	0.0076	(0.0087)		0.0123	(0.0091)	
REL STD DEV	18.478	(28.803)		13.972	(340.267)	

-----						
Solution =	0.040 g/210L	or	0.1905 mg/l,	Samples =	4,	Discarded = 1
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)	
Sample #1	0.7390	(-0.0100)		1.4640	(0.0040)	
Sample #2	0.7850	(-0.0240)		1.4900	(-0.0250)	
Sample #3	0.7450	(0.0000)		1.4630	(0.0000)	
Sample #4	0.7710	(0.0000)		1.5160	(-0.0100)	
Avg % Abs	0.7670	(-0.0080)		1.4897	(-0.0117)	
STD DEV	0.0203	(0.0139)		0.0265	(0.0126)	
REL STD DEV	2.646	(173.205)		1.779	(107.855)	

-----						
Solution =	0.100 g/210L	or	0.4762 mg/l,	Samples =	4,	Discarded = 1
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)	
Sample #1	1.8340	(-0.0120)		3.5740	(-0.0210)	
Sample #2	1.8070	(-0.0060)		3.5420	(0.0010)	
Sample #3	1.8450	(-0.0200)		3.5470	(-0.0010)	
Sample #4	1.8010	(0.0080)		3.5100	(0.0110)	
Avg % Abs	1.8177	(-0.0060)		3.5330	(0.0037)	
STD DEV	0.0239	(0.0140)		0.0201	(0.0064)	
REL STD DEV	1.313	(233.333)		0.568	(175.339)	

-----						
Solution =	0.200 g/210L	or	0.9524 mg/l,	Samples =	4,	Discarded = 1
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)	
Sample #1	3.5570	(-0.0020)		6.8560	(-0.0100)	
Sample #2	3.5270	(0.0200)		6.8220	(0.0390)	
Sample #3	3.5690	(0.0070)		6.8510	(0.0120)	
Sample #4	3.5190	(0.0360)		6.8220	(0.0370)	
Avg % Abs	3.5383	(0.0210)		6.8317	(0.0293)	
STD DEV	0.0269	(0.0145)		0.0167	(0.0150)	
REL STD DEV	0.759	(69.171)		0.245	(51.288)	

-----						
Solution =	0.300 g/210L	or	1.4286 mg/l,	Samples =	4,	Discarded = 1
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)	
Sample #1	5.2080	(-0.0160)		9.9260	(-0.0240)	
Sample #2	5.1840	(0.0000)		9.8810	(0.0160)	
Sample #3	5.1840	(0.0000)		9.8750	(0.0130)	
Sample #4	5.1470	(0.0210)		9.8570	(0.0380)	
Avg % Abs	5.1717	(0.0070)		9.8710	(0.0223)	
STD DEV	0.0214	(0.0121)		0.0125	(0.0137)	
REL STD DEV	0.413	(173.205)		0.127	(61.121)	

100

2/23/18  
 JZ

FL HIGHWAY PATROL  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-003411  
 02/22/2018 09:49:18

Auto Calibration

pg 2 of 2

```

<<<<<      3um      >>>>>
-----
Zero Order Coef   -106.54
First Order Coef  2605.83
Second Order Coef  34.03
  
```

```

<<<<<      9um      >>>>>
-----
Zero Order Coef   -101.21
First Order Coef  1319.58
Second Order Coef  13.83
  
```

```

-----
Act      Fit      Residual
(g/210L) (g/210L) (g/210L)
0.000    0.000    -0.0000
0.040    0.040    -0.0002
0.100    0.100    0.0004
0.200    0.200    -0.0003
0.300    0.300    0.0001
-----
  
```

```

-----
Act      Fit      Residual
(g/210L) (g/210L) (g/210L)
0.000    0.000    -0.0003
0.040    0.040    0.0002
0.100    0.099    0.0006
0.200    0.201    -0.0007
0.300    0.300    0.0003
-----
  
```

```

<<<<<      3um      >>>>>
-----
  
```

```

<<<<<      9um      >>>>>
-----
  
```

Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1  
 Sample

Sample #1	3384.00	3450.00
Sample #2	3382.00	3449.00
Sample #3	3387.00	3430.00
Sample #4	3371.00	3441.00
Avg	3380.0000	3440.0000
STD DEV	8.1854	9.5394
REL STD DEV	0.242	0.277
H2O adjust (mg/l*10k)	429	369

Barometric Pressure = 1030

\*\*\*\*\*CALIBRATION SUCCESSFUL\*\*\*\*\*

*le*

*2/23/18  
sa*

# Post - Calibration Adjust

Stability Checks #80-003411 FL Highway Patrol 2/22/18 ~~2018~~

*DGS*

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003411  
02/22/2018  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:57
Control Test	0.050	10:58
Air Blank	0.000	10:58
Control Test	0.049	10:59
Air Blank	0.000	11:00
Control Test	0.050	11:00
Air Blank	0.000	11:01
Control Test Stats	0.0497	
Average	0.0006	
Std Dev	1.1625	
Rel Std Dev(%)		

*DGS*  
Operator's Signature

2/23/18  
*DGS*

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003411  
02/22/2018  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:07
Control Test	0.200	11:08
Air Blank	0.000	11:08
Control Test	0.200	11:09
Air Blank	0.000	11:09
Control Test	0.200	11:10
Air Blank	0.000	11:11
Control Test Stats	0.2000	
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*DGS*  
Operator's Signature

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003411  
02/22/2018  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:16
Control Test	0.080	11:16
Air Blank	0.000	11:17
Control Test	0.079	11:17
Air Blank	0.000	11:18
Control Test	0.079	11:18
Air Blank	0.000	11:19
Control Test Stats	0.0793	
Average	0.0006	
Std Dev	0.7277	
Rel Std Dev(%)		

*DGS*  
Operator's Signature