



INSTRUMENT PROCESSING SHEET

Agency Patrick AFBS/N 80-005146

Florida Department of Law Enforcement

Date In 11/27/18DI Completion Date 12/7/18 Ship P/U H/D CMI EE

Intake Performed By <u>SQC</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>SP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>143</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>.148</u> (.139 - .169) 36 mm <u>.160</u> (.156 - .190) 53 mm <u>.238</u> (.228 - .278) 103 mm <u>.523</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28421</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																											
Final Release Date <div style="text-align: center; font-weight: bold; font-size: 1.2em;">FDLE</div> <div style="text-align: center; font-weight: bold; font-size: 1.2em;">DEC 10 2018</div> <div style="text-align: center; font-weight: bold;">Alcohol Testing Program</div>	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1012</td> <td>201707D 7-25-19</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> <td>201707E 7-25-19</td> </tr> <tr> <td>0.200</td> <td>DR3856</td> <td>201707C 7-24-19</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG805701 2-26-20</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD1012	201707D 7-25-19	0.080	DR1279	201707E 7-25-19	0.200	DR3856	201707C 7-24-19	0.080 DGS	N/A	AG805701 2-26-20	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>SP</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.5</u> External Digital Therm. ID#: <u>300504</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD1012</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>DR1279</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>DR3856</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	SD1012	201707D 7-25-19																																																											
0.080	DR1279	201707E 7-25-19																																																											
0.200	DR3856	201707C 7-24-19																																																											
0.080 DGS	N/A	AG805701 2-26-20																																																											
Calibration Adjustment Performed By <u>SP</u> Barometric Pressure Gauge <u>1028</u> ID # <u>30793</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G2834</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>SD1022</td> <td>17410</td> <td>12-6-19</td> </tr> <tr> <td>0.100</td> <td>SD3964</td> <td>18070</td> <td>2-26-20</td> </tr> <tr> <td>0.200</td> <td>SD1025</td> <td>17340</td> <td>10-9-19</td> </tr> <tr> <td>0.300</td> <td>SD1024</td> <td>18110</td> <td>4-2-20</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>17817080A2</td> <td>8-5-19</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1012</td> <td>201707D</td> <td>7-25-19</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> <td>201707E</td> <td>7-25-19</td> </tr> <tr> <td>0.200</td> <td>DR3856</td> <td>201707C</td> <td>7-24-19</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG805701</td> <td>2-26-20</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	G2834	N/A	N/A	0.040	SD1022	17410	12-6-19	0.100	SD3964	18070	2-26-20	0.200	SD1025	17340	10-9-19	0.300	SD1024	18110	4-2-20	0.080 DGS	N/A	17817080A2	8-5-19	Simulator	Serial Number	Lot Number	Expiration	0.050	SD1012	201707D	7-25-19	0.080	DR1279	201707E	7-25-19	0.200	DR3856	201707C	7-24-19	0.080 DGS	N/A	AG805701	2-26-20	Department Inspection Performed By <u>SP</u> Barometric Pressure ID# <u>28421</u> Gauge <u>1028</u> Instrument <u>1029</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G2408</td> </tr> <tr> <td>Interferent</td> <td>G2882</td> </tr> <tr> <td>0.050</td> <td>SD1012</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> </tr> <tr> <td>0.200</td> <td>DR3856</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	G2408	Interferent	G2882	0.050	SD1012	0.080	DR1279	0.200	DR3856
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	G2834	N/A	N/A																																																										
0.040	SD1022	17410	12-6-19																																																										
0.100	SD3964	18070	2-26-20																																																										
0.200	SD1025	17340	10-9-19																																																										
0.300	SD1024	18110	4-2-20																																																										
0.080 DGS	N/A	17817080A2	8-5-19																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050	SD1012	201707D	7-25-19																																																										
0.080	DR1279	201707E	7-25-19																																																										
0.200	DR3856	201707C	7-24-19																																																										
0.080 DGS	N/A	AG805701	2-26-20																																																										
Simulator	Serial Number																																																												
0.000	G2408																																																												
Interferent	G2882																																																												
0.050	SD1012																																																												
0.080	DR1279																																																												
0.200	DR3856																																																												
Notes/Suggested Service: <u>Performed cal adjust due to barometric pressure difference > 10 pts. SP</u> _____ _____ _____	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____ <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="display: flex; justify-content: space-between; margin-top: 10px;"> <div style="text-align: center;"> <u>AGM 12/10/18</u> Tech Review / Date </div> <div style="text-align: center;"> <u>J. Jordan 12/10/18</u> Admin Review / Date </div> </div>																																																												

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PATRICK AFB
Time of Inspection: 11:28

Date of Inspection: 12/07/2018

Serial Number: 80-005146
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.049	0.079	0.197	0.078
0.000	0.049	0.080	0.198	0.078
0.000	0.050	0.080	0.199	0.078
0.000	0.050	0.080	0.200	0.078
0.000	0.050	0.080	0.200	0.077
0.000	0.050	0.080	0.199	0.078
0.000	0.050	0.080	0.199	0.078
0.000	0.049	0.080	0.199	0.078
0.000	0.049	0.080	0.199	0.077
0.000	0.050	0.080	0.200	0.078

Standard Deviations	0.0005	0.0003	0.0009	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT
Signature and Printed Name

12/07/2018
Date

*12/10/18
JD*

80-005146
STABILITY CHECKS

PATRICK AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005146
12/04/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:44
Control Test	0.083	09:45
Air Blank	0.000	09:45
Control Test	0.082	09:46
Air Blank	0.000	09:46
Control Test	0.082	09:46
Air Blank	0.000	09:47
Control Test Stats		
Average	0.0823	
Std Dev	0.0006	
Rel Std Dev(%)	0.7012	

DGS

SP

Operator's Signature

PATRICK AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005146
12/04/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:49
Control Test	0.199	09:49
Air Blank	0.000	09:50
Control Test	0.197	09:51
Air Blank	0.000	09:51
Control Test	0.197	09:52
Air Blank	0.000	09:52
Control Test Stats		
Average	0.1977	
Std Dev	0.0012	
Rel Std Dev(%)	0.5842	

SP

Operator's Signature

PATRICK AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005146
12/04/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:39
Control Test	0.080	09:40
Air Blank	0.000	09:41
Control Test	0.079	09:41
Air Blank	0.000	09:42
Control Test	0.080	09:43
Air Blank	0.000	09:43
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

SP

Operator's Signature

PATRICK AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005146
12/04/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:35
Control Test	0.049	09:36
Air Blank	0.000	09:36
Control Test	0.050	09:37
Air Blank	0.000	09:37
Control Test	0.050	09:38
Air Blank	0.000	09:38
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

SP

Operator's Signature

SP
SP
SP



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005146, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005146</u>	UNCERTAINTY* ±	
Owning Agency:	<u>PATRICK AFB</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>12/07/2018</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>11:28</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Shayla Platt

12/07/2018 Date
SHAYLA D PLATT,
Department Inspector

FDLE/ATP Form 69 July 2018
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Adam
8/10/18
EC

PATRICK AFB

Intoxilyzer - Alcohol Analyzer

SN 80-005146

08:43:14

Auto Calibration

Max Power Res Value = 33

Auto Range Res Value = 15

Sol Value = 0.000 g/210L ***

Fit value = 0.0000 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12727, Sum Io = 12530

<<<<< CHANNEL 1 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.1430 (-0.0180)

Sample #2 = 0.0830 (0.0820)

Sample #3 = 0.1350 (0.1160)

Sample #4 = 0.0600 (0.1880)

Avg % Abs = 0.0927 (0.1287)

STD DEV = 0.0384 (0.0541)

REL STD DEV = 41.464 (42.065)

<<<<< CHANNEL 2 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.1910 (0.0000)

Sample #2 = 0.1280 (0.0410)

Sample #3 = 0.1580 (0.0510)

Sample #4 = 0.1370 (0.0860)

Avg % Abs = 0.1410 (0.0593)

STD DEV = 0.0154 (0.0236)

REL STD DEV = 10.918 (39.824)

Sol Value = 0.040 g/210L ***

Fit value = 0.1905 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12693, Sum Io = 12514

<<<<< CHANNEL 1 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.7940 (-0.0100)

Sample #2 = 0.8140 (-0.0030)

Sample #3 = 0.7770 (0.0520)

Sample #4 = 0.8210 (0.0610)

Avg % Abs = 0.8040 (0.0367)

STD DEV = 0.0236 (0.0346)

REL STD DEV = 2.941 (94.489)

<<<<< CHANNEL 2 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 1.5010 (0.0030)

Sample #2 = 1.5260 (-0.0030)

Sample #3 = 1.4830 (0.0270)

Sample #4 = 1.5270 (0.0230)

Avg % Abs = 1.5120 (0.0157)

STD DEV = 0.0251 (0.0163)

REL STD DEV = 1.661 (103.973)

Sol Value = 0.100 g/210L ***

Fit value = 0.4762 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12683, Sum Io = 12508

<<<<< CHANNEL 1 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 1.9030 (-0.0110)

Sample #2 = 1.9270 (0.0040)

Sample #3 = 1.8740 (0.0380)

Sample #4 = 1.9130 (0.0340)

Avg % Abs = 1.9047 (0.0253)

STD DEV = 0.0275 (0.0186)

REL STD DEV = 1.442 (73.355)

<<<<< CHANNEL 2 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 3.4890 (-0.0040)

Sample #2 = 3.5030 (0.0000)

Sample #3 = 3.4960 (0.0050)

Sample #4 = 3.4890 (0.0120)

Avg % Abs = 3.4960 (0.0057)

STD DEV = 0.0070 (0.0060)

REL STD DEV = 0.200 (106.371)

Sol Value = 0.200 g/210L ***

Fit value = 0.9524 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12675, Sum Io = 12502

<<<<< CHANNEL 1 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 3.5670 (-0.0070)

Sample #2 = 3.5600 (0.0370)

Sample #3 = 3.5760 (0.0450)

Sample #4 = 3.6280 (0.0260)

Avg % Abs = 3.5880 (0.0360)

STD DEV = 0.0356 (0.0095)

REL STD DEV = 0.991 (26.498)

<<<<< CHANNEL 2 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 6.5900 (-0.0160)

Sample #2 = 6.5670 (0.0310)

Sample #3 = 6.5860 (0.0260)

Sample #4 = 6.5920 (0.0300)

Avg % Abs = 6.5823 (0.0290)

STD DEV = 0.0134 (0.0026)

REL STD DEV = 0.204 (9.123)

Sol Value = 0.300 g/210L ***

Fit value = 1.4286 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12670, Sum Io = 12500

<<<<< CHANNEL 1 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 5.3020 (-0.0100)

Sample #2 = 5.3320 (0.0110)

Sample #3 = 5.2930 (0.0340)

Sample #4 = 5.3170 (0.0420)

Avg % Abs = 5.3140 (0.0290)

STD DEV = 0.0197 (0.0161)

REL STD DEV = 0.370 (55.495)

<<<<< CHANNEL 2 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 9.6360 (-0.0120)

Sample #2 = 9.6180 (0.0460)

Sample #3 = 9.6210 (0.0510)

Sample #4 = 9.5660 (0.0840)

Avg % Abs = 9.6017 (0.0603)

STD DEV = 0.0309 (0.0206)

REL STD DEV = 0.322 (34.223)

***** AUTO CAL DATA *****

<<<<< CHANNEL 1 >>>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.093

Std Dev = 0.04 Rel Std Dev = 41.46

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 0.804

Std Dev = 0.02 Rel Std Dev = 2.94

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 1.905

Std Dev = 0.03 Rel Std Dev = 1.44

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 3.588

Std Dev = 0.04 Rel Std Dev = 0.99

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 5.314

Std Dev = 0.02 Rel Std Dev = 0.37

Zero Order Coef = -251.93

First Order Coef = 2636.11

Second Order Coef = 19.48

Standard Deviation = 53.790192

<<<<< CHANNEL 2 >>>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.141

Std Dev = 0.02 Rel Std Dev = 10.92

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 1.512

Std Dev = 0.03 Rel Std Dev = 1.66

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 3.496

Std Dev = 0.01 Rel Std Dev = 0.20

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 6.582

Std Dev = 0.01 Rel Std Dev = 0.20

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 9.602

Std Dev = 0.03 Rel Std Dev = 0.32

Zero Order Coef = -213.59

First Order Coef = 1388.27

Second Order Coef = 12.85

Standard Deviation = 30.305580

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.000 0.0002
0.040 0.039 0.0005
0.100 0.102 -0.0016
0.200 0.199 0.0014
0.300 0.300 -0.0004

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.000 0.0004
0.040 0.040 -0.0002
0.100 0.101 -0.0007
0.200 0.199 0.0009
0.300 0.300 -0.0003

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1
Sample #1 = 3345.00
Sample #2 = 3378.00
Sample #3 = 3359.00
Sample #4 = 3375.00
Average Result = 3370.6667
STD DEV = 10.2144
REL STD DEV = 0.303

***** CHANNEL 2
Sample #1 = 3363.00
Sample #2 = 3377.00
Sample #3 = 3367.00
Sample #4 = 3390.00
Average Result = 3378.0000
STD DEV = 11.5326
REL STD DEV = 0.341

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1028
3 um H2O Adjust (mg/l*10,000) = 439
9 um H2O Adjust (mg/l*10,000) = 431
**** AUTO CAL PASS

CAL ADJUSTMENT
#80-005146

ADP
2/10/16
JA

POST CAL ADJUST STABILITY CHECKS # 80-005146

PATRICK AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005146
12/07/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:39
Control Test	0.048	09:40
Air Blank	0.000	09:40
Control Test	0.049	09:41
Air Blank	0.000	09:41
Control Test	0.049	09:42
Air Blank	0.000	09:43
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

SP

Operator's Signature

PATRICK AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005146
12/07/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:44
Control Test	0.079	09:44
Air Blank	0.000	09:45
Control Test	0.080	09:46
Air Blank	0.000	09:46
Control Test	0.080	09:47
Air Blank	0.000	09:47
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

SP

Operator's Signature

PATRICK AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005146
12/07/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:34
Control Test	0.198	09:35
Air Blank	0.000	09:35
Control Test	0.198	09:36
Air Blank	0.000	09:36
Control Test	0.198	09:37
Air Blank	0.000	09:38
Control Test Stats		
Average	0.1980	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

PATRICK AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005146
12/07/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:51
Control Test	0.079	09:51
Air Blank	0.000	09:52
Control Test	0.079	09:52
Air Blank	0.000	09:53
Control Test	0.080	09:53
Air Blank	0.000	09:54
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DAS

SP

Operator's Signature

WDM
12/10/18
SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PATRICK AFB
Time of Inspection: 09:17

Date of Inspection: 12/04/2018

Serial Number: 80-005146
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:

Time-Date changed. BYPASSED AI TO OPERATE INSTRUMENT

Handwritten initials/signature

N/A COMPLIANCE NOT DETERMINED

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT
Signature and Printed Name

12/04/2018
Date

12/10/18 SD