



INSTRUMENT PROCESSING SHEET

Agency Pinellas County Sheriff's OfficeS/N 80-005290Florida Department of
Law EnforcementDate In 02/23/2018DI Completion Date 03/14/2018 Ship P/U H/D CMI EE

Intake Performed By <u>PJM</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____		Quality Checks Performed By <u>DMB</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>126</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 103</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.496</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28662 1001/1020</u> <input checked="" type="checkbox"/> Stability Checks		Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																							
Final Release Date <p style="text-align: center;">FDLE</p> <p style="text-align: center;">MAR 14 2018</p> <p style="text-align: center;">Alcohol Testing Program</p>		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>G11739</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>SD3964</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>DR3856</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG715202 06/01/2019</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	G11739	201707D 07/25/2019	0.080	SD3964	201707E 07/25/2019	0.200	DR3856	201707C 07/24/2019	0.080 DGS	N/A	AG715202 06/01/2019	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>DMB</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.6</u> External Digital Therm. ID#: <u>300503</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>G11739</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3964</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>DR3856</u>																								
Simulator	Serial #	Lot #/Exp																																									
0.050	G11739	201707D 07/25/2019																																									
0.080	SD3964	201707E 07/25/2019																																									
0.200	DR3856	201707C 07/24/2019																																									
0.080 DGS	N/A	AG715202 06/01/2019																																									
Calibration Adjustment Performed By <u>DMB</u> Barometric Pressure Gauge <u>1020</u> ID # <u>28427</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G2879</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>SD1024</td> <td>16320</td> <td>10/21/2018</td> </tr> <tr> <td>0.100</td> <td>G2834</td> <td>17280</td> <td>09/11/2019</td> </tr> <tr> <td>0.200</td> <td>SD1011</td> <td>17090</td> <td>02/24/2019</td> </tr> <tr> <td>0.300</td> <td>DR1275</td> <td>17140</td> <td>05/15/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>22817080A5</td> <td>10/05/2019</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks		Simulator	Serial Number	Lot Number	Expiration	0.000	G2879	N/A	N/A	0.040	SD1024	16320	10/21/2018	0.100	G2834	17280	09/11/2019	0.200	SD1011	17090	02/24/2019	0.300	DR1275	17140	05/15/2019	0.080 DGS	N/A	22817080A5	10/05/2019	Department Inspection Performed By <u>DMB</u> Barometric Pressure ID# <u>28662</u> Gauge <u>1021</u> Instrument <u>1020</u> Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2017-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD1019</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.050</td> <td>G11739</td> </tr> <tr> <td>0.080</td> <td>SD3964</td> </tr> <tr> <td>0.200</td> <td>DR3856</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	SD1019	Interferent	SD1021	0.050	G11739	0.080	SD3964	0.200	DR3856
Simulator	Serial Number	Lot Number	Expiration																																								
0.000	G2879	N/A	N/A																																								
0.040	SD1024	16320	10/21/2018																																								
0.100	G2834	17280	09/11/2019																																								
0.200	SD1011	17090	02/24/2019																																								
0.300	DR1275	17140	05/15/2019																																								
0.080 DGS	N/A	22817080A5	10/05/2019																																								
Simulator	Serial Number																																										
0.000	SD1019																																										
Interferent	SD1021																																										
0.050	G11739																																										
0.080	SD3964																																										
0.200	DR3856																																										
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>G11739</td> <td>201707D</td> <td>07/25/2019</td> </tr> <tr> <td>0.080</td> <td>SD3964</td> <td>201707E</td> <td>07/25/2019</td> </tr> <tr> <td>0.200</td> <td>DR2856</td> <td>201707C</td> <td>07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG715202</td> <td>06/01/2019</td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.050	G11739	201707D	07/25/2019	0.080	SD3964	201707E	07/25/2019	0.200	DR2856	201707C	07/24/2019	0.080 DGS	N/A	AG715202	06/01/2019	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____																					
Simulator	Serial Number	Lot Number	Expiration																																								
0.050	G11739	201707D	07/25/2019																																								
0.080	SD3964	201707E	07/25/2019																																								
0.200	DR2856	201707C	07/24/2019																																								
0.080 DGS	N/A	AG715202	06/01/2019																																								
Notes/Suggested Service: <u>Optical Bench Calibration</u> <u>Adjustment performed to bring barometric pressure</u> <u>values closer to nominal. 3/14/18 DMB</u> _____ _____ _____		<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <u>DMB 3/14/18</u> <u>J. Dehan 3/14/18</u> Tech Review / Date Admin Review / Date																																									

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS COUNTY SO
Time of Inspection: 14:28

Date of Inspection: 03/14/2018

Serial Number: 80-005290
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG715202 Exp: 06/01/2019
0.000	0.049	0.080	0.199	0.080
0.000	0.050	0.081	0.199	0.080
0.000	0.050	0.081	0.200	0.080
0.000	0.050	0.082	0.200	0.080
0.000	0.050	0.081	0.200	0.080
0.000	0.049	0.081	0.200	0.080
0.000	0.050	0.082	0.199	0.080
0.000	0.049	0.082	0.201	0.079
0.000	0.049	0.082	0.200	0.080
0.000	0.050	0.082	0.201	0.080

Standard Deviations	0.0005	0.0006	0.0007	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Room

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Danielle M Bell

DANIELLE M BELL

Signature and Printed Name

03/14/2018
Date

*3/14/18
JD*

Stability Checks #80-005290 Pinellas County SO. 3/13/18 ~~2018~~

DOE

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 80-005290
03/13/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:40
Control Test	0.050	11:40
Air Blank	0.000	11:41
Control Test	0.049	11:42
Air Blank	0.000	11:42
Control Test	0.049	11:43
Air Blank	0.000	11:43
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel. Std Dev(%)	1.1703	

[Signature]

Operator's Signature

3/14/18
[Signature]

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 80-005290
03/13/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:45
Control Test	0.060	11:46
Air Blank	0.000	11:46
Control Test	0.091	11:47
Air Blank	0.000	11:48
Control Test	0.077	11:48
Air Blank	0.000	11:49
Control Test Stats		
Average	0.0793	
Std Dev	0.0021	
Rel. Std Dev(%)	2.6239	

[Signature]

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 80-005290
03/13/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:51
Control Test	0.196	11:51
Air Blank	0.000	11:52
Control Test	0.199	11:53
Air Blank	0.000	11:53
Control Test	0.199	11:54
Air Blank	0.000	11:54
Control Test Stats		
Average	0.1967	
Std Dev	0.0006	
Rel. Std Dev(%)	0.2906	

[Signature]

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 80-005290
03/13/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:55
Control Test	0.083	11:56
Air Blank	0.000	11:56
Control Test	0.081	11:56
Air Blank	0.000	11:57
Control Test	0.082	11:57
Air Blank	0.000	11:58
Control Test Stats		
Average	0.0820	
Std Dev	0.0010	
Rel. Std Dev(%)	1.2195	

[Signature]

Operator's Signature

[Signature]



Florida Department of Law Enforcement
 Alcohol Testing Program
 2729 Fort Knox Blvd.
 Bldg. 2, Suite 1300
 Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005290, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005290</u>	UNCERTAINTY* ±
Owning Agency:	<u>PINELLAS COUNTY SO</u>	0.050 g/ 210 L
Calibration Date:	<u>03/14/2018</u>	0.080 g/ 210 L
Calibration Time:	<u>14:28</u>	0.200 g/ 210 L
		0.080 g/ 210 L Dry Gas Control
		0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration. *Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/14/2018

Date

DANIELLE M BELL,
 Department Inspector

FDLE/ATP Form 69 March 2018
 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

3/14/18
ee

ee

Optical bench calibration adjustment data #80-005290 Pinellas County S.O. 3/14/18 ~~PHS~~

PINELLAS COUNTY SO
 Intoxilyzer - Hicorol Analyzer
 Model 8000
 03/14/2018
 09:53:20

Auto Calibration
 Max Power Res Value = 46
 Auto Range Res Value = 29

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.5660 (0.0050)
 Sample #2 = 1.5550 (0.0190)
 Sample #3 = 1.5460 (0.0170)
 Sample #4 = 1.5090 (0.0070)
 Avg % Abs = 1.5633 (0.0143)
 STD DEV = 0.0227 (0.0064)
 REL STD DEV = 1.451 (44.854)

Sol Value = 0.040 g/210L ***
 Fit Value = 0.1905 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12767, Sum Io = 13491
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.7840 (-0.0170)
 Sample #2 = 0.7800 (-0.0050)
 Sample #3 = 0.8120 (-0.0040)
 Sample #4 = 0.7440 (-0.0390)
 Avg % Abs = 0.7787 (0.0100)
 STD DEV = 0.0340 (0.0251)
 REL STD DEV = 4.369 (251.197)

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.5540 (0.0050)
 Sample #2 = 1.5740 (-0.0070)
 Sample #3 = 1.5860 (-0.0030)
 Sample #4 = 1.5510 (0.0160)
 Avg % Abs = 1.5703 (0.0020)
 STD DEV = 0.0178 (0.0123)
 REL STD DEV = 1.133 (614.410)

Sol Value = 0.100 g/210L ***
 Fit Value = 0.4782 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12754, Sum Io = 13399
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.8830 (-0.0030)
 Sample #2 = 1.8770 (0.0200)
 Sample #3 = 1.8370 (0.0350)
 Sample #4 = 1.8620 (0.0100)
 Avg % Abs = 1.8587 (0.0217)
 STD DEV = 0.0282 (0.0126)
 REL STD DEV = 1.387 (58.876)

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.6310 (-0.0020)
 Sample #2 = 3.6480 (0.0160)
 Sample #3 = 3.6160 (0.0110)
 Sample #4 = 3.6320 (0.0040)
 Avg % Abs = 3.6320 (0.0103)
 STD DEV = 0.0160 (0.0060)
 REL STD DEV = 0.441 (58.333)

Sol Value = 0.200 g/210L ***
 Fit Value = 0.9524 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12759, Sum Io = 13398
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.5960 (-0.0070)
 Sample #2 = 3.6340 (-0.0140)
 Sample #3 = 3.5950 (0.0100)
 Sample #4 = 3.6480 (-0.0120)
 Avg % Abs = 3.6257 (-0.0053)
 STD DEV = 0.0275 (0.0133)
 REL STD DEV = 0.758 (249.687)

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 6.9260 (0.0000)
 Sample #2 = 6.9820 (-0.0020)
 Sample #3 = 6.9310 (-0.0190)
 Sample #4 = 6.9730 (-0.0050)
 Avg % Abs = 6.9620 (0.0040)
 STD DEV = 0.0272 (0.0131)
 REL STD DEV = 0.391 (326.517)

Sol Value = 0.300 g/210L ***
 Fit Value = 1.4286 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12757, Sum Io = 13396
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 5.2910 (-0.0160)
 Sample #2 = 5.3450 (-0.0230)
 Sample #3 = 5.3250 (0.0070)
 Sample #4 = 5.2850 (0.0080)
 Avg % Abs = 5.3183 (-0.0027)
 STD DEV = 0.0306 (0.0176)
 REL STD DEV = 0.574 (660.611)

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 10.0300 (0.0020)
 Sample #2 = 10.0410 (0.0030)
 Sample #3 = 10.0290 (0.0130)
 Sample #4 = 10.0190 (0.0130)
 Avg % Abs = 10.0297 (0.0097)
 STD DEV = 0.0110 (0.0058)
 REL STD DEV = 0.110 (59.726)

***** AUTO CAL DATA *****
 <<<<< CHANNEL 1 >>>>>
 Sol Val = 0.0099 mg/l or 0.000 g/210L
 % Abs = 0.051
 Std Dev = 0.03 Rel Std Dev = 59.35
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 0.779
 Std Dev = 0.03 Rel Std Dev = 4.37
 Sol Val = 0.4782 mg/l or 0.100 g/210L
 % Abs = 1.859
 Std Dev = 0.02 Rel Std Dev = 1.09
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 3.626
 Std Dev = 0.03 Rel Std Dev = 0.76
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 5.318
 Std Dev = 0.03 Rel Std Dev = 0.57
 Zero Order Coef = -121.42
 First Order Coef = 2573.40
 Second Order Coef = 25.26
 Standard Deviation = 12.589899

<<<<< CHANNEL 2 >>>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.177
 Std Dev = 0.01 Rel Std Dev = 5.98
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 1.570
 Std Dev = 0.02 Rel Std Dev = 1.13
 Sol Val = 0.4782 mg/l or 0.100 g/210L
 % Abs = 3.632
 Std Dev = 0.02 Rel Std Dev = 0.44
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 6.962
 Std Dev = 0.03 Rel Std Dev = 0.39
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 10.030
 Std Dev = 0.01 Rel Std Dev = 0.11
 Zero Order Coef = -210.55
 First Order Coef = 1314.31
 Second Order Coef = 12.91
 Standard Deviation = 30.318558

 ***** CHANNEL 2 *****
 Sample #1 = 3298.00
 Sample #2 = 3294.00
 Sample #3 = 3274.00
 Sample #4 = 3275.00
 Average Result = 3281.0000
 STD DEV = 11.2694
 REL STD DEV = 0.343

 ***** CHANNEL 1 *****
 Sample #1 = 3298.00
 Sample #2 = 3294.00
 Sample #3 = 3274.00
 Sample #4 = 3275.00
 Average Result = 3281.0000
 STD DEV = 11.2694
 REL STD DEV = 0.343

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1020
 3 um H2O Adjust (mg/l*10.000) = 433
 9 um H2O Adjust (mg/l*10.000) = 528
 ***** AUTO CAL PASS *****

 Solution Stats Quadratic Fit Chan 1
 Act Fit Residual
 g/210L g/210L g/210L
 0.000 0.000 -0.0032
 0.040 0.040 0.0001
 0.100 0.100 0.0003
 0.200 0.200 -0.0004
 0.300 0.300 0.0001

 Solution Stats Quadratic Fit Chan 2
 Act Fit Residual
 g/210L g/210L g/210L
 0.000 0.000 -0.0005
 0.040 0.040 0.0004
 0.100 0.099 0.0006
 0.200 0.201 -0.0009
 0.300 0.300 0.0003

Sol Value = 0.080 g/210L ***
 Fit Value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1
 ***** CHANNEL 1 *****
 Sample #1 = 3330.00
 Sample #2 = 3371.00
 Sample #3 = 3394.00
 Sample #4 = 3363.00
 Average Result = 3376.0000
 STD DEV = 16.0935
 REL STD DEV = 0.477

***** CHANNEL 2 *****
 Sample #1 = 3298.00
 Sample #2 = 3294.00
 Sample #3 = 3274.00
 Sample #4 = 3275.00
 Average Result = 3281.0000
 STD DEV = 11.2694
 REL STD DEV = 0.343

 ***** CHANNEL 1 *****
 Sample #1 = 3298.00
 Sample #2 = 3294.00
 Sample #3 = 3274.00
 Sample #4 = 3275.00
 Average Result = 3281.0000
 STD DEV = 11.2694
 REL STD DEV = 0.343

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1020
 3 um H2O Adjust (mg/l*10.000) = 433
 9 um H2O Adjust (mg/l*10.000) = 528
 ***** AUTO CAL PASS *****

3/14/18

Post Calibration Adjustment
 Stability Checks #80-005290 Pinellas County S.O. 3/14/18 ~~QAS~~

QAS

PINELLAS COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 03/14/2018
 Software: 8100.27

PINELLAS COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 03/14/2018
 Software: 8100.27

PINELLAS COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 03/14/2018
 Software: 8100.27

PINELLAS COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 03/14/2018
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:25
Control Test	0.050	11:26
Air Blank	0.000	11:26
Control Test	0.050	11:27
Air Blank	0.000	11:27
Control Test	0.050	11:28
Air Blank	0.000	11:28
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel. Std Dev(%)	0.0000	

Test	g/210L	Time
Air Blank	0.000	11:29
Control Test	0.081	11:30
Air Blank	0.000	11:31
Control Test	0.081	11:31
Air Blank	0.000	11:32
Control Test	0.081	11:33
Air Blank	0.000	11:33
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel. Std Dev(%)	0.0000	

Test	g/210L	Time
Air Blank	0.000	11:35
Control Test	0.199	11:35
Air Blank	0.000	11:36
Control Test	0.199	11:36
Air Blank	0.000	11:37
Control Test	0.200	11:38
Air Blank	0.000	11:38
Control Test Stats		
Average	0.1993	
Std Dev	0.0006	
Rel. Std Dev(%)	0.2956	

Test	g/210L	Time
Air Blank	0.000	11:40
Control Test	0.081	11:41
Air Blank	0.000	11:41
Control Test	0.080	11:42
Air Blank	0.000	11:42
Control Test	0.080	11:42
Air Blank	0.000	11:43
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel. Std Dev(%)	0.7187	

QAS
 Operator's Signature
 3/14/18
ee

QAS
 Operator's Signature

QAS
 Operator's Signature

QAS
 Operator's Signature

QAS