



# INSTRUMENT PROCESSING SHEET

Agency Broward County Sheriff's OfficeS/N 80-007006

Florida Department of Law Enforcement

Date In 05/15/2018 DI Completion Date 05/16/2018 Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>DELR</u> <input checked="" type="checkbox"/> Annual <input checked="" type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>DELR</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>201</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 101</u> 32 mm <u>.160</u> (.139 - .169) 36 mm <u>.183</u> (.156 - .190) 53 mm <u>.250</u> (.228 - .278) 103 mm <u>.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28199</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG805701 02/26/2020</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD3967	201707D 07/25/2019	0.080	SD3968	201707E 07/25/2019	0.200	SD3969	201707C 07/24/2019	0.080 DGS	N/A	AG805701 02/26/2020	<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) <b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Temperature Checks</b> Performed By <u>DELR</u> <input checked="" type="checkbox"/> Lab Temp °C <u>23.09C</u> External Digital Therm. ID#: <u>300949</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3967</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3968</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3969</u>																																													
Simulator	Serial #	Lot #/Exp																																																												
0.050	SD3967	201707D 07/25/2019																																																												
0.080	SD3968	201707E 07/25/2019																																																												
0.200	SD3969	201707C 07/24/2019																																																												
0.080 DGS	N/A	AG805701 02/26/2020																																																												
<b>Final Release Date</b> <p style="text-align: center; font-size: 1.2em;">FDLE</p> <p style="text-align: center; font-size: 1.2em;">MAY 24 2018</p> <p style="text-align: center;">Alcohol Testing Program</p>																																																														
<b>Calibration Adjustment</b> Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<b>Department Inspection</b> Performed By <u>DELR</u> Barometric Pressure ID# <u>28663</u> Gauge <u>1015</u> Instrument <u>1016</u> Mouth Alcohol Solution Lot # <u>2017-B</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD3965</td> </tr> <tr> <td>Interferent</td> <td>SD3966</td> </tr> <tr> <td>0.050</td> <td>SD3967</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> </tr> </tbody> </table> <b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input checked="" type="checkbox"/> Other <u>Form 47</u>		Simulator	Serial Number	0.000	SD3965	Interferent	SD3966	0.050	SD3967	0.080	SD3968	0.200	SD3969
Simulator	Serial Number	Lot Number	Expiration																																																											
0.000		N/A	N/A																																																											
0.040																																																														
0.100																																																														
0.200																																																														
0.300																																																														
0.080 DGS	N/A																																																													
Simulator	Serial Number	Lot Number	Expiration																																																											
0.050																																																														
0.080																																																														
0.200																																																														
0.080 DGS	N/A																																																													
Simulator	Serial Number																																																													
0.000	SD3965																																																													
Interferent	SD3966																																																													
0.050	SD3967																																																													
0.080	SD3968																																																													
0.200	SD3969																																																													
Notes/Suggested Service: <u>E-mailed</u> <input checked="" type="checkbox"/> <span style="color: red; font-weight: bold; font-size: 1.2em;">APPROVED</span> _____ _____ _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="display: flex; justify-content: space-between; margin-top: 10px;"> <div style="text-align: center;"> <u>9:00am 5/24/18</u>              Tech Review / Date           </div> <div style="text-align: center;"> <u>[Signature]</u>              Admin Review / Date <u>5/24/18</u> </div> </div>																																																													

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO  
Time of Inspection: 11:31

Date of Inspection: 05/16/2018

Serial Number: 80-007006  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.047	0.079	0.198	0.079
0.000	0.048	0.080	0.200	0.078
0.000	0.048	0.080	0.199	0.078
0.000	0.049	0.080	0.199	0.078
0.000	0.049	0.079	0.200	0.078
0.000	0.049	0.080	0.199	0.079
0.000	0.049	0.080	0.198	0.078
0.000	0.048	0.079	0.198	0.079
0.000	0.049	0.080	0.198	0.078
0.000	0.049	0.080	0.198	0.078

Standard Deviations	0.0007	0.0004	0.0008	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

*DM*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*David E Reyes Rivera*

DAVID E REYES-RIVERA

Signature and Printed Name

05/16/2018  
Date

*5/24/18  
JO*

<b>TYPE OF TEST</b>	<b>SERIAL NUMBER</b>	<b>AGENCY</b>	<b>DATE</b>	<b>PERFORMED BY</b>
Stabilities	80-007006	Broward County Sheriff's Office	05/16/2018	<i>Dee</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
<b>SN: SD3967 Temp: 34.10c</b> <b>0.047 to 0.053</b> <input checked="" type="checkbox"/>	<b>SN: SD3968 Temp: 34.09c</b> <b>0.077 to 0.083</b> <input checked="" type="checkbox"/>	<b>SN: SD3969 Temp: 34.10c</b> <b>0.194 to 0.206</b> <input checked="" type="checkbox"/>	<b>Lot AG805701</b> <b>0.077 to 0.083</b> <input checked="" type="checkbox"/>																																																																																																																																																
<p>BROWARD COUNTY SO Intoxilyzer - Alcoloh Analyzer Model 8000 SN 80-107006 05/16/2018 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:16</td></tr> <tr><td>Control Test</td><td>0.047</td><td>06:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:17</td></tr> <tr><td>Control Test</td><td>0.048</td><td>06:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:19</td></tr> <tr><td>Control Test</td><td>0.048</td><td>06:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:20</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0477</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2112</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:16	Control Test	0.047	06:17	Air Blank	0.000	06:17	Control Test	0.048	06:18	Air Blank	0.000	06:19	Control Test	0.048	06:19	Air Blank	0.000	06:20	Control Test Stats			Average	0.0477		Std Dev	0.0006		Rel Std Dev(%)	1.2112		<p>BROWARD COUNTY SO Intoxilyzer - Alcoloh Analyzer Model 8000 SN 80-107006 05/16/2018 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:24</td></tr> <tr><td>Control Test</td><td>0.078</td><td>06:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:24</td></tr> <tr><td>Control Test</td><td>0.079</td><td>06:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:24</td></tr> <tr><td>Control Test</td><td>0.080</td><td>06:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:25</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2658</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:24	Control Test	0.078	06:24	Air Blank	0.000	06:24	Control Test	0.079	06:25	Air Blank	0.000	06:24	Control Test	0.080	06:24	Air Blank	0.000	06:25	Control Test Stats			Average	0.0790		Std Dev	0.0010		Rel Std Dev(%)	1.2658		<p>BROWARD COUNTY SO Intoxilyzer - Alcoloh Analyzer Model 8000 SN 80-107006 05/16/2018 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:27</td></tr> <tr><td>Control Test</td><td>0.196</td><td>06:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:28</td></tr> <tr><td>Control Test</td><td>0.196</td><td>06:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:29</td></tr> <tr><td>Control Test</td><td>0.197</td><td>06:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:30</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1963</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2941</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:27	Control Test	0.196	06:27	Air Blank	0.000	06:28	Control Test	0.196	06:29	Air Blank	0.000	06:29	Control Test	0.197	06:30	Air Blank	0.000	06:30	Control Test Stats			Average	0.1963		Std Dev	0.0006		Rel Std Dev(%)	0.2941		<p>BROWARD COUNTY SO Intoxilyzer - Alcoloh Analyzer Model 8000 SN 80-107006 05/16/2018 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:33</td></tr> <tr><td>Control Test</td><td>0.078</td><td>06:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:34</td></tr> <tr><td>Control Test</td><td>0.079</td><td>06:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:35</td></tr> <tr><td>Control Test</td><td>0.080</td><td>06:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:36</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2658</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:33	Control Test	0.078	06:34	Air Blank	0.000	06:34	Control Test	0.079	06:34	Air Blank	0.000	06:35	Control Test	0.080	06:35	Air Blank	0.000	06:36	Control Test Stats			Average	0.0790		Std Dev	0.0010		Rel Std Dev(%)	1.2658	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:16																																																																																																																																																	
Control Test	0.047	06:17																																																																																																																																																	
Air Blank	0.000	06:17																																																																																																																																																	
Control Test	0.048	06:18																																																																																																																																																	
Air Blank	0.000	06:19																																																																																																																																																	
Control Test	0.048	06:19																																																																																																																																																	
Air Blank	0.000	06:20																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0477																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.2112																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:24																																																																																																																																																	
Control Test	0.078	06:24																																																																																																																																																	
Air Blank	0.000	06:24																																																																																																																																																	
Control Test	0.079	06:25																																																																																																																																																	
Air Blank	0.000	06:24																																																																																																																																																	
Control Test	0.080	06:24																																																																																																																																																	
Air Blank	0.000	06:25																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	1.2658																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:27																																																																																																																																																	
Control Test	0.196	06:27																																																																																																																																																	
Air Blank	0.000	06:28																																																																																																																																																	
Control Test	0.196	06:29																																																																																																																																																	
Air Blank	0.000	06:29																																																																																																																																																	
Control Test	0.197	06:30																																																																																																																																																	
Air Blank	0.000	06:30																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1963																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2941																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:33																																																																																																																																																	
Control Test	0.078	06:34																																																																																																																																																	
Air Blank	0.000	06:34																																																																																																																																																	
Control Test	0.079	06:34																																																																																																																																																	
Air Blank	0.000	06:35																																																																																																																																																	
Control Test	0.080	06:35																																																																																																																																																	
Air Blank	0.000	06:36																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	1.2658																																																																																																																																																		
<p><i>Stacy/18</i> <i>JD</i></p> <p><i>Dee</i> Operator's Signature</p>	<p><i>Dee</i> Operator's Signature</p>	<p><i>Dee</i> Operator's Signature</p>	<p><i>Dee</i> Operator's Signature</p>																																																																																																																																																

*WJPD*



Florida Department of Law Enforcement  
 Alcohol Testing Program  
 4700 Terminal Drive, Suite 1  
 Ft. Myers, FL 33907

# Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007006, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007006</u>	UNCERTAINTY* ±	
Owning Agency:	<u>BROWARD COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>05/16/2018</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>11:31</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.  
 Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
 \*Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

05/16/2018

Date

*David E Reyes-Rivera*

DAVID E REYES-RIVERA,  
 Department Inspector

FDLE/ATP Form 69 March 2018  
 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

*5/24/18*  
*JD*

*QPM*

# Florida Department of Law Enforcement Alcohol Testing Program

---

## REGISTRATION OF EVIDENTIARY BREATH TEST INSTRUMENT

**MANUFACTURER:** CMI, Inc.  
**MODEL:** Intoxilyzer 8000  
**SERIAL NUMBER:** 80-007006  
**OWNER:** Broward County Sheriff's Office  
**DATE OF REGISTRATION:** May 29, 2018

The above instrument is hereby approved for evidentiary breath alcohol testing in the State of Florida pursuant to Chapter 11D-8, Florida Administrative Code. This instrument and related records are subject to inspection at any time by the Florida Department of Law Enforcement.



---

Authorized Representative  
Alcohol Testing Program  
Florida Department of Law Enforcement



Florida Department of Law Enforcement

# REQUEST FOR REGISTRATION

MAKE AND MODEL OF INSTRUMENT: Intoxilyzer 8000

SERIAL NUMBER: 80-007006

OWNING AGENCY: Broward County Sheriff's Office

DATE OF DEPARTMENT INSPECTION: 5/16/2018

AGENCY INSPECTOR: Yvette Washington

ADDRESS: 200 NW 27TH Avenue

CITY, STATE, ZIP: Fort Lauderdale, Florida 33311

TELEPHONE NUMBER: (954) 321-4849

FAX NUMBER: (954) 321-4684

EMAIL ADDRESS (if available): yvette\_washington@sheriff.org

**For Program Office Use Only:**

- Registration Issued *BK*
- Instrument Added to Evidentiary Instrument Spreadsheet *BK*
- Instrument Added to Evidentiary Instrument Database *SP*
- Instrument Added to Monthly Statistics Database *SP*
- Contact Information Added to Contact Spreadsheet *SP*

*5/24/18*  
*J12*